अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ)



All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, GE Road, Raipur-492 099 (CG) <u>www.aiimsraipur.edu.in</u>

Admin/ Store/Purchase/Non-Consumable/Anaesthesiology /2016-AIIMS.RPR Date: 03.08.2016

Inviting Quotations for Purchase of Non-Consumable items for department of Anaesthesiology, AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of Non-Consumable Items for department of Anaesthesiology, AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer** up to 08/08/2016 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

Item Description	Qty Required	Purpose	Unit Rate
 Peripheral Nerve Stimulator (PNS) Specification Amplitude settings: Digital and Analog dial display of the stimulting voltage. Display Size should be W*H in mm: 50*20 Display Type: Graphic/Standard/ Custom LCD. Current : 0-5 mAmp for cannuls mode and 0-6- mA for percutaneous nerve mapping mode. Max Voltage Output : 60-90 V Max Output load (Impedance) Nominal/Max: 12/17kOhm 	Qty Required One	Purpose Teaching and training purpose.	Unit Rate
 7. Impulse Duration : 0.05,0.1,0.2,0.3,0.5, 1.0 ms. 8. Stimulus Frequency : 1Hz, 2Hz. 9. Should have a 			
	 Peripheral Nerve Stimulator (PNS) Specification Amplitude settings: Digital and Analog dial display of the stimulting voltage. Display Size should be W*H in mm: 50*20 Display Type: Graphic/Standard/ Custom LCD. Current : 0-5 mAmp for cannuls mode and 0-6- mA for percutaneous nerve mapping mode. Max Voltage Output : 60-90 V Max Output load (Impedance) Nominal/Max: 12/17kOhm (5mAmp) Impulse Duration : 0.05,0.1,0.2,0.3,0.5, 1.0 ms. Stimulus Frequency : 1Hz, 2Hz. 	RequiredPeripheralNerveOneStimulator (PNS)OneSpecification1. Amplitude settings: Digital and Analog dial display of the stimulting voltage.2. Display Size should be W*H in mm: 50*203. Display Type: Graphic/Standard/ Custom LCD.3. DisplayType: Graphic/Standard/ Custom LCD.4. Current : 0-5 mAmp for cannuls mode and 0-6- mA for percutaneous nerve mapping mode.5. Max Voltage Output : 60-90 V6. Max Output load (Impedance) Nominal/Max: 12/17kOhm (5mAmp)7. Impulse Duration : 0.05,0.1,0.2,0.3,0.5, 1.0 ms.8. Stimulus Frequency : 1Hz, 2Hz.9. Should have a	RequiredPeripheralNerveStimulator (PNS)OneSpecificationTeaching and training purpose.1. Amplitude settings: Digital and Analog dial display of the stimulting voltage.Display Size should be W*H in mm: 50*202. Display Size should be W*H in mm: 50*20So*203. DisplayType: Graphic/Standard/ Custom LCD.4. Current : 0-5 mAmp for cannuls mode and 0-6- mA for percutaneous nerve mapping mode.5. Max Voltage Output : 60-90 V6. Max Output load (Impedance) Nominal/Max: 12/17kOhm (5mAmp)7. Impulse Duration : 0.05,0.1,0.2,0.3,0.5, 1.0 ms.8. Stimulus Frequency : 1Hz, 2Hz.9. Should have a

		1
current.		
10.		
11.		
12.Should have a		
display of set		
current.		
13.Should have a		
display of Impulse		
Duration.		
14.Alarm signal :		
Should be alerted by		
a change in beep		
tone.		
15.Menu for setup and		
freatures should be		
there.		
16.PERCUTANEOUS		
NERVE MAPPING:		
Facility should be		
there in the		
standard		
equipment.		
17.Remote		
Control:Optional .		
18.Size Up to H * W* D		
(mm) : 160*90*40.		
19. Stimulation		
switchable between		
cannula and		
percutaneous		
stimulating handle		
-		
for nerve mapping.		
20.Should have low		
power consumption		
without power		
shutoff.		
21.Coded connector		
Phy on the case.		
22. Should have inbuilt		
battery.		
	I	1

1) Taxes, if any (Kindly mention the rate.)

2)Delivery Schedule – Within 15 days from the date of issue of PO

3)Price should be for- Destination basis

4) LD @ 0.5% per week maximum upto 10% will be deducted for delay in supply.5)Quotation Name/No. should be mentioned on top of envelops.

Sr.Administrative Officer

AIIMS, Raipur (C.G.)