



No. AIIMS, RPR/Store/Purchase/Surgical Micromotor for Dentistry/2015/

Date: 01/01/2016

Inviting Quotations for Supply of Surgical Micromotor for Dentistry AIIMS, Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of Surgical Micromotor Handpiece for Dentistry AIIMS, Raipur and should be submitted to office of Store Officer up to 3:00 pm on 18th January, 2016. The quotations will be opened on the same day. As per detailed below:-

S.No	Item Description	Preferred Make	Qty.	Unit Price
1.	Surgical Micromotor Handpiece US – FDA/CE – European for the product and model Manufacturer/Supplier should have ISO certification for quantity standards. Surgical Micro motor piece (scratch resistant, titanium body)	NSK/ Kavo/ W&H	5	
	Specification:			
	Surgical handpieces without light			
	Transmission ratio:	1:1		
	Spray:	External		
	Chucking System:	Level Chuck		
	Rotary Instrument:	For surgical burs and cutters Ø 2.35 mm (also Stryker System)		
	Maximum Length:	45 mm – 70 mm		
	Maximum drive speed:	40,000 min. ⁻¹ - 50,000 min. ⁻¹		
	Dismountable:	Yes		
	The unit should have high torque even at slow speed to enact smooth bone and tooth structure cutting. The speed should adjustable from 1000 rpm to 40,000 rpm.			

1. Taxes, if any (Kindly Mention).
2. Document relating to registration of firm i.e. should be submitted along with quotation.
3. Delivery Schedule – within 7 to 15 days from the date of issue of PO.
4. Price should be for Destination basis.
5. LD @ 0.5% of PO value per week will be deducted for delay of supply of material subject to maximum up to 10% of PO value.
6. Quotation No/Name and Due date of opening must be written on top of envelop.

Store Officer
AIIMS, Raipur (C.G.)

