



अखिलभारतीयआयुर्विज्ञानसंस्थान(छत्तीसगढ़)रायपुर ,
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh,
Raipur-492 099 (CG)

Quotation Notice

Admin / Store/Purchase/Hospital Equipment / 2015-AIIMS.RPR/ Dated: 17.05.2016

Inviting quotation for Supply of Equipment's for Hospital, AIIMS, Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of serum Equipment's for Hospital, AIIMS, Raipur and should be submitted to office of Store Officer up to 3:00 pm on 23.05.2016. The quotations will be opened on the same day at 3:30pm. As per detailed below:-

S.No	Name of Instrument	Quantity	Unit rate
		Required	
1	BP instrument, with Adult cuff, make Heine or other similar quality.	15 pieces	
	Technical Specification for B.P apparatus:		
(i)	The unit shall be CE certified.		
(ii)	The company shall have a local service centre based at Raipur or anywhere in Chhattisgarh. Please provide address of service centre.		
(iii)	The company/Supplier shall have experience in supplying/sale of above items in govt.sector.		
(iv)	Should be aneroid type.		
(v)	Should have ameasuring range from 0 to 300 mmHg,		
(vi)	Should be provided with adult arm cuff.		
(vii)	The dial manometer markings and graduations should be permanent and clearly visible and filled with pigments.		
(viii)	Movement mechanism- Brass		
(ix)	Connection: brass,nickel plated for 3-4 mm rubber hose.		
(x)	Dial - Aluminium		

(xi)	Window lenses – Clear plastic.		
(xii)	All plastic parts, if any used should not crack, flake, peel or disintegrate in normal use.		
(xiii)	Should provide a carry bag to keep the whole system safe and sound. All parts should be replaceable in case of breakage.		
(xiv)	Shock Proof.		
(xv)	Large Diameter Scale should easy to read.		
(xvi)	Large insufflations Bulb: Large Bulb Volume for fast inflation of the cuff.		
(xvii)	Warranty should be at least 2 year and should cover all accessorie in warranty.		
(xviii)	User/technical/maintenance manual should be supplied.		
(Xix)	Should provide demonstration before supply.		
2	Stethoscope , adult	15 pieces	
	Technical Specification of Stethoscope :		
(i)	Product should be CE certified.		
(ii)	Chrome-plated,dual -head chest piece.		
(iii)	48mm chest piece with diaphragm.		
(iv)	Should provide extra rubber ear plug.		
(v)	Should provide demonstration before supply.		
(vi)	Ear plug should be siliconised.		
3	Weighing Machine	20 pieces	
	Technical specification for Weighing Machine.		
(i)	Should be CE certified.		
(ii)	Weighing scale should be suitable for weighing up to 120kg weight.		
(iii)	Weighing scale should be having graduation of 500gm.		
(iv)	Base,top cover should be made of steel approximately 1mm thick.		
(v)	The design of scale should be portable.		
(vi)	The scale mechanism protected against rust & corrosion and sprayed in enamel colours having attractive finish.		

(vii)	There should be arrangement for zero adjustment.		
(viii)	The reading scale should be clearly visible having plastic magnifying lens.		
(ix)	Should provide demonstration before supply.		

1. Taxes, if any (Kindly Mention).
2. Document relating to registration of firm i.e. should be submitted along with quotation.
3. Delivery Schedule - within 15 days from the date of issue of PO.
4. Price should be for-Destination basis.
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation Name/No. should be mentioned on top of envelopes.

Sr.Administrative Officer
AIIMS, Raipur (CG)