



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

H. Store/Purchase/N. C.-Anaesth /2016-AIIMS RPR

Date: 06/12/2016.

Inviting Quotations for Purchase of Non Consumable for Anaesthesia dept. in major O.T, for the smooth functioning of MOT, at AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of Non consumable items for major O.T. at AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer Room No. – 51, 2nd floor Ayush Building, Gate No. 1 up to 15/12/2016 before 3:00 pm**. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.No	Complete Description of items (specification model, catalogueNo.)Use separate sheet if required& signed by indenter and HOD.	Unit	Quantity Required	Make /model/ brand	Unit rate in Rs.
1	Ambu bag size adult, Detail specification as per (annexure A).	Nos	02		
2	Ambu bag size paediatric Detail specification as per (annexure A).	Nos	02		
3	Reservoir bag size 0.5 lit, Detail specification as per (annexure B).	Nos	03		
4	Reservoir bag size 1 lit. Detail specification as per (annexure B).	Nos	03		
5	Reservoir bag size 1.5 lit. Detail specification as per (annexure B).	Nos	03		
6	Reservoir bag size 2 lit.Detail specification as per (annexure B).	Nos	03		
7	Disposable breathing circuit Size Paediatric	Set	10		
8	Laryngoscope with blade adult. Detail specification as per (annexure C).	Set	03		
9	Laryngoscope with blade paediatric. Detailspecification as per (annexure C).	Set	03		

Note: Firm to submit sample along with quotation .sample shall be returned after finalization of tender.

Terms & Condition

1. Firm to mention Manufactures name/Make/Brand name in their quotation.
2. Taxes should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. Tin number and relevant document should be submitted along with quotation.
4. Products Quoted should be certified from ISI/GMP/FDA/CE0434 as applicable, copy of certificate should be submitted with quotation.

5. Supply should be done within 15 days after Placement of PO & delivery challan to be submitted with supply.
6. Price should be FOR Destination basis (i.e. concerned department).
7. 100 % Payment will be released after certification on original invoice from department of anesthesia.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply shall be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms, and to increase/decrease the number of required quantity.
11. All other terms & condition as per GCC applicable.
12. Material to be delivered at **Gate No-01, Hospital store, AYUSH PMR Building, 2nd Floor, Room No.-56.**
13. Validity of quotation should be 90 days from the date of opening

Stores Officer (H)
AIIMS, Raipur (C.G.)

ANNEXURE - A

Technical specification of ambu/resuscitator Bag for Adult patients

1. CE European or USFDA certified, self-inflating bag.
2. It shall be made of high translucent liquid silicone which shall be free from latex.
3. It shall have a detachable, reusable reservoir bag of not less than 2.5 liter capacity, detachable transparent silicone mask of size 4 and 5.
4. Capacity of bag shall not be less than 1600 ml. Bag shall have a texture that provides firm grip. It shall be sturdy to bear wear and tear.
5. It shall have facility for oxygen supplementation.
6. One way valve of fish mouth-type, which minimizes possibility of wrong assembly. The casing, which houses the fish mouth valve shall be crystal clear to allow unimpeded observation of the valve operation and airflow to the patient.
7. Shall not allow build up pressure to more than 70 cm H₂O.
8. Shall be able to wash with detergent and complete disinfection by Cidex solution and shall be able to be sterilized properly by autoclaving at 134°C.
9. One sample of ambu/resucitator bag has to be deposited at the time of submission of bid.
10. Demonstration of bag can be done at AIIMS Raipur

ANNEXURE- B

Technical specification of reservoir/breathing bag

1. CE European or USFDA certified, antistatic, reservoir/ breathing bag of capacity 0.5, 1.0, 1.5 and 2.0 liter capacity.
2. It shall be designed to provide optimal feedback of accurate pressure assessment. It shall conform to ASTM standard of <1.5 l bag, pressure shall be not less than 30 cm of H₂O and over 50 cm of H₂O when the bag is expanded to 4 times its capacity. For bags >1.5 l capacity pressure shall be not less than 35 cm of H₂O and over 60 cm of H₂O when the bag is expanded to 4 times its capacity.
3. Preferable green in colour, ellipsoid in shape, made of latex free synthetic material.
4. Its neck shall have 22 mm female connector, tail opposite to the neck.
5. It shall be non-sterile and reusable.
6. It shall be washable with normal water and detergents.
7. One sample of the reservoir bag has to be deposited at the time of submission of bid.

Sr no.	Capacity of bag (liters)
1	0.5
2	1.0
3	1.5
4	2.0

ANNEXURE – C

Technical specification of paediatric and adult and paediatric laryngoscope

1. CE European or USFDA certified direct laryngoscope handle with detachable blades.
2. Portable, shall not be bulky.
3. Equipment shall be sturdy, high quality stainless steel, matt/glossy finish, resistant to corrosion, resistant to breakage and damage on fall on hard surface or hit against wall, shall bear drop test of 2 meter.
4. Handle of standard size and pencil light handle. It shall have shell and battery with LED/halogen lamp technology. Battery shall be 2.5 v rechargeable /non rechargeable with a capacity of not less than 2 Hrs. of illumination. Handle and blade shall be interchangeable.
5. Classic Macintosh blade size 0,1,2,3 and 4 made of high quality stainless steel matt/glossy finish, light source with LED/halogen lamp technology. Blade shall be smooth, smooth edges to avoid injury, will be maintenance free.
6. Classic Miller blade size 00, 0, 1, 2 made of high quality stainless steel matt finish, light source with LED/halogen lamp technology. Blade shall be smooth, smooth edges to avoid injury, maintenance free.
7. Light source in oropharynx should be from diffuse LED with balanced illumination. Temperature of light source or blade should not exceed > 40°C even on prolonged use.
8. Warranty period of the laryngoscope shall be **5 years**.
9. Demonstration of equipment at AIIMS Raipur.

Laryngoscope with blade adult		
Sr. no.	Equipment name	type
1.	Standard handle	
2.	Pen light handle	
3.	Classic Macintosh blade	0
		1
		2
		3
		4
4	Rechargeable battery	1 set

Laryngoscope with blade pediatrics		
Sr. no.	Equipment name	type
1	Standard handle	
2.	Pen light handle	
3.	Classic miller blade	00
		0
		1
		2
4	Rechargeable battery	1 set