



No. AIIMS/R/Cantal Store/Tender/Psychiatry/01/2016/A

Raipur, Dtd:-07/03/2017

### Corrigendum

<b>Page/Clause/ Point</b>	<b>Existing Sentence</b>	<b>Amendment</b>
Page No 15 of 20 Point No. 10	Should have stimulus current from 500-900mA, Frequency 10-120Hz, Pulse width – 0.1 -1 msec , Stimulus duration of 0.14 to 8 sec	Should have stimulus current from 500-900mA, Frequency 20-120Hz, Pulse width – 0.3 -1 msec , Stimulus duration of 0.05 to 8 sec
Page No 16 of 20 Point No. 28	The system delivery should include installation on turnkey basis with dedicated onsite training to end users by the authorized representatives from the principal manufacturers.	“System delivery should include installation on turnkey basis with dedicated onsite training to Doctors/Technicians/Staff should be provided by authorized representatives from the principal manufactures/authorized representatives form factory trained engineer at free of cost for smooth operation and maintenance of the equipment to the satisfaction of the user department”

All other terms and condition will remain unchanged.

**Stores Officer**  
**AIIMS Raipur (CG)**