



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Corrigendum for Craniofacial Distraction kit

Tender Enquiry No. Store/Tender/Craniofacial_Distraction_Kit/1 /2015 Dt. 18.06.2015

<u>Page no/ Clause no</u>	<u>Old</u>	<u>Updated</u>
Page no. 26 Sr. no. 1	<ul style="list-style-type: none">• Should include for both left and right sides. (2 distractors left and right form a set)• Should be able to attain distraction of at least 40 mm	<ul style="list-style-type: none">• Should include one set of Pediatric Distractor and one set of adult distractor. Each set comprises of left and right distractors.• Should be able to attain distraction of at least 25mm with pediatric distractor and 30mm with adult distractor

Store Officer
AIIMS, Raipur