



Tender enquiry no. Store/Tender/Refrigerated_Blood_Bag_Centrifuge/2/2015

Dt. 24/11/2015

Corrigendum

Page No/ Clause no/ point no	Existing sentence	To be read as
Page no. 18 Point no. 10	Provide a set of equipments for providing calibration leg thermometer and routine preventive Maintenance as per manufacturer documentation in service/technical manual. Should provide Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Extra set of buckets for the blood bags should be provided with the equipment, along with balancing weights.	Should calibrate the equipment on installation with details of routine preventive maintenance. Should provide Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out. Extra set of plastic buckets for the blood bags should be provided with the equipment, along with balancing weights.
Page no. 18 Point no. 13	CE/FDA/ BIS certification specific for the product should be submitted, non-submission shall amount to the rejection of the equipment from the competitive bid.	CE/FDA and BIS certification specific for the product should be submitted, non-submission shall amount to the rejection of the equipment from the competitive bid.

Store Officer
AIIMS, Raipur