



Corrigendum

MULTIPARA MONITOR (HIGH END) (04 Nos.) Schedule-A

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Last date of submission/ opening where mentioned in tender document	Last date of submission is 19.04.2016 at 3:00 PM and opening of technical bid 19.04.2016 at 3:30 PM	Last date of submission is 25.04.2016 at 3:00 PM and opening of technical bid 25.04.2016 at 3:30 PM
2	Page no. 15 Point no. 1	All the material /equipment should be European CE & US FDA.	All the material /equipment should be European CE & US FDA 510k.
3	Page no. 18 Point no. 9	<ul style="list-style-type: none">Should provide mainstream modules-2nos and microstream - 2 nos.	<ul style="list-style-type: none">Should provide mainstream modules-2nos and microstream/ low flow side stream 0.1-5ml/min - 2 nos.
4	Addendum as point no. 24 on Page no. 19	Nil	24. Central Station (01 no.):for 8 Bed <ul style="list-style-type: none">Should have capability of monitoring at least eight beds Wired/wirelessly on central station with at least two waveforms from each bedside being viewed. Additional parameters to be displayed in numeric form.Should have dual displays, one for continuous display of all beds and other for trend review, full disclosure etc.Size of both the displays should be at least 21" flat screen TFTShould have facility for alarm review capability of a minimum of two waveforms per eventShould have storage of minimum of 40 eventsShould provide multilead arrhythmia analysis..Laser printer should also be included for printing of data including 'holter like' full disclosure reports and vital sign data from bedside.Should have 24 hour full disclosure for at least three waveforms.Should have facility for remote access of patient data including waveforms and numerics.Remote patient data should be viewed at Consultants residence or at Chamber through HIS and static IP (HIS and static IP shall be



Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
			<p>provided by the hospital for the purpose).</p> <ul style="list-style-type: none">• Suitable rated UPS of 30 minutes backup to be provided for the Complete System i.e. monitors and central station.• It should be networked with other central stations i.e. CTVS ICU, CTVS step down ICU, CCU, CCU step down ICU.• It should have HL7 compliant and be compatible with any HIS system• Should have valid US FDA certification.

NEONATAL MULTIPARA MONITOR -HIGH END (2 Nos.) Schedule-B

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page no. 23 Point no. 11	Microstream end tidal capnography (EtCO2)	Microstream end tidal capnography (EtCO2) / Low flow side stream < 1ml/min.

Store Officer,
AIIMS, Raipur