



Tender enquiry no. STORE/Tender/Hospital Furniture /3/2015 date 02/09/2015

Corrigendum

Page No/ Clause no/ point no	Existing sentence	To be read as
Page no. 3 Point no. 3	<p>Size shall be more than 900mm L x500mm W x 1500mm H.</p> <p>The crash cart should be made of 25.4mmx18G Stainless steel tubular frame work.</p> <p>Shall have Epoxy / Anti-Microbial powder paint inside and ...</p>	<p>Size shall be more than 900mm L x500mm W x 1500mm H.</p> <p>Shall have Epoxy / Anti-Microbial powder paint inside and ...</p>
Page no. 4 Point no. 2	<p>Soiled linen trolley-</p> <p>Overall approx size: 910mm H x 510mm dia. SS tubular framework fitted with three swivel castors, 100mm dia. Framework made of 25.4mm dia x18G verticals, 10</p> <p>Upper ring made of 19mm dia x 18G tube and support stays of 15.8mm dia. X 18G stainless steel tubes. Supplied with canvas bag. All SS components should be of 304 qualities.</p>	<p>Soiled linen trolley-</p> <p>Overall approx size: 910mm H x 510mm dia. SS tubular framework fitted with three swivel castors, 100mm dia. Framework made of 25.4mm dia x18G verticals, Upper ring made of 19mm dia x 18G tube and support stays of 15.8mm dia. X 18G stainless steel tubes. Supplied with canvas bag. All SS components should be of 304 qualities.</p>