

Tender Enquiry No. Store/ Tender/Ultra Deep Freezer/1/2015Dated 11/07/2015

Corrigendum for "Ultra Deep Freezer for Microbiology Department"

Page No./Clause No.	Existing Parameter	Amendment
Page no. 19 Point no. 11	Inner door : 5 independent interior compartment doors to reduce cold air loss, improve temperature recovery after door closing and to help maintain cabinet temperature during door opening.	Inner door : 4-5 independent interior compartment doors to reduce cold air loss, improve temperature recovery after door closing and to help maintain cabinet temperature during door opening.
Page no. 19 Point no. 12	Shelves : 4 sturdy, solid stainless steel shelves.	Shelves : 3-4 sturdy, solid stainless steel shelves.

Store Officer AIIMS, Raipur