



Date:- 23-12-2016

Corrigendum

No. AIIMS/Store/Tender/Pharmaceutical Refrigerator/2016/01

Page/Clause/ Point	Existing Sentence	To be read as
As mentioned in page no. 16 of Tender Document	<p>Pharmaceutical refrigerator Quantity: 04 Nos.</p> <ol style="list-style-type: none">1. Temperature Range :2°C to 12°C; +/-0.5°C2. Volume Capacity :370 Lt3. Inner Cabinet :Stainless Steel 304 Grade4. Outer Cabinet :Powder Coated CRC Sheet5. Temperature Control : Microprocessor with PT-100 Sensor6. Display: 4 inch attractive LCD display.7. Alarms: Power failure visual alarm, Audio Visual alarm in case of door open for over one minute, Audio visual temperature variation alarm. Alarms for high/low set parameters8. Illumination : 8 Watts Fluorescent Tube9. Electrical : 220-240 Volts, 50 Hz Single Phase.10. Insulation : 70 mm minimum for Body and 80 mm for Door, CFC free polyurethane foam11. Powerful fan motor for forced air circulation to maintain uniform conditions inside chamber12. Unique design of thermal barrier for better energy efficiency13. Hermetically sealed Compressor with CFC free refrigerant14. Adjustable trays15. Electrical circuit breaker	<p>Pharmaceutical refrigerator Quantity: 04 Nos.</p> <ol style="list-style-type: none">1. Temperature Range : 2°C to 12°C; +/-0.5°C2. Volume Capacity : 370 Lt3. Inner Cabinet : Stainless Steel 304 Grade4. Outer Cabinet : Powder Coated CRC Sheet5. Temperature Control : Microprocessor with PT-100 Sensor6. Display : Approx. 4 inch attractive LCD display.7. Alarms : Power failure visual alarm, Audio Visual alarm in case of door open for over one minute, Audio visual temperature variation alarm. Alarms for high/low set parameters8. Illumination : 8 Watts Fluorescent Tube9. Electrical : 220-240 Volts, 50 Hz Single Phase.10. Insulation : 70 mm minimum for Body and 80 mm for Door, CFC free polyurethane foam or double glass door insulation.11. Powerful fan motor for forced air circulation to maintain uniform conditions inside chamber12. Unique design of thermal barrier for better energy efficiency13. Hermetically sealed Compressor with CFC free refrigerant14. Adjustable trays15. Electrical circuit breaker



	<p>16. Time delay for compressor switch on 17. Overload cut off relay for compressor 18. Intelligent Controller helps maintain temperature in case of sensor failure 19. Battery Backup for Temperature Controller 20. Auto tuning of controller 21. Self-Diagnosis for errors 22. Histogram format of 24 hours temperature recording 23. Power failure and resumption recorded with date and time. 24. Model : CC-12 plus (LCD Version). 25. Warranty one year.</p>	<p>16. Time delay for compressor switch on 17. Overload cut off relay for compressor 18. Intelligent Controller helps maintain temperature in case of sensor failure 19. Battery Backup for Temperature Controller 20. Auto tuning of controller 21. Self-Diagnosis for errors 22. Histogram format of 24 hours temperature recording 23. Power failure and resumption recorded with date and time. 24. Warranty one year.</p>
As mentioned in Tender Document	Last date of submission of tender on 16-12-2016 at 3.00 PM and opening of tender on 16-12-2016 at 3.30 PM	Last date of submission of tender on 28-12-2016 at 3.00 PM and opening of tender on 28-12-2016 at 3.30 PM

Points to be added

Point No. 25	It should be a Vertical Pharmaceutical refrigerator.
--------------	--

**Sr. AO/Stores Officer
AIIMS Raipur (CG)**