



No. Admin/store/purchase/ Vertical Autoclave / 2015-AIIMS- RPR

18th March, 2015

INVITING QUOTATIONS FOR Supply of Vertical Autoclave for Department of Microbiology, AIIMS, Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of Vertical Autoclave for Department of Microbiology, AIIMS, Raipur and should be submitted to office of Store Officer up to 3:00 pm on 26th March, 2015. The quotations will be opened on the same day. As per detailed below:-

S. No.	Items name	Specification	Qty	Unit Rate in ₹
1.	Vertical Autoclave	<ul style="list-style-type: none">• Approximate Size: 457X 711 mm (≥110 litre capacity)• Fully stainless steel double walled with chamber & jacket of SS 304 & ceramic wool insulation in between SS 304 boiler tank.• Lid should have joint less silicon gasket• Should be equipped with dial pressure gauge, water level indicator• Should have working pressure of 15-20 psi and 121°C working temperature• Should have steam release valve, and safety valve.• Should be fitted with ISI marked Heaters• Should be operated in mains 220 – 240 V AC 50/60 Hz input power supply.• Should have an ISI mark.• Pedal lifting & handle to lift the lid.	One	

1. Taxes, if any (Kindly Mention).
2. Delivery Schedule – within 10 days from the date of issue of PO.
3. Price should be for Destination basis.
4. LD @ 0.5% of PO value per week for delay of supply of material subject to maximum up to 10% of PO value.

(Store Officer)
AIIMS, Raipur (CG)

