



## Corrigendum

Subject: This office tender enquiry number: ADMIN/Tender/OT/1/2013/ Dated 10-09-2013

For supply, installation and commissioning of anaesthesia work station at All India Institute of Medical Sciences Raipur (CG)

Various queries raised by bidders during pre-meeting held at Committee room 1st floor medical college building, AIIMS Raipur, has been examined by technical and commercial committee and amendment is issued here under.

Page no /Clause no / Point no	Existing sentence	To be read as
Page 2; Table Row vi; Column 3	17.10.2013, 1400 hours	<b>24.10.2013 1400 hours</b>
Table Row vii; Column 3	17.10.2013,1430hrsIST	<b>24.10.2013,1430hrsIST</b>
Clause 7 (Tender issued on 10-09-2013)	7. Should be able to hold three seletatec vaporizers (Isoflurane, Sevoflurane & Desflurane) simultaneously. Vaporizers should be maintenance free. Cost of vaporizers to be quoted separately. The anesthesia machine should provide for desflurane compensation.	7. Should be able to hold <b>any two-</b> (Isoflurane, Desflurane Sevoflurane) <b>selectatec</b> vaporizers simultaneously. Vaporizers should be maintenance free. Two anesthesia machine should be provided with sevoflurane vaporizer and the other 2 should be provided with isoflurane vaporizer Cost of vaporizers (Isoflurane, Desflurane Sevoflurane) should be quoted separately also. The anesthesia machine should provide for desflurane compensation.
Clause 13 c (corrigendum issued on 01-10-2013)	Monitor should have 12" color <b>touch</b> screen and color coding for displayed parameters	Monitor should have 12" color screen and color coding for displayed parameters
Add Clause 19 (Corrigendum issued on 01-10-2013)	The Anesthesia work station should be provided with anesthesia management system with a keyboard and a mouse	<b>The Anesthesia work station system should automatically record parameters displayed on patient monitor in graphical as well as in digital format. The system should also record parameters displayed on ventilator and respiratory gas module. The resolution of such recorded parameters should be user selectable and configurable. The Anesthesia work station should be provided with a keyboard and a mouse for manually entering user defined data. The system should provide a facility to lock the configuration. The system should be capable to generate user configurable reports.</b>

Page no /Clause no / Point no	Existing sentence	To be read as
		<b>The system should have facility to backfill / edit the data up to a user defined time; thereafter, the system should not allow editing of the data.</b>
<b>Monitor</b>		
Clause 2 (corrigendum issued on 01-10-2013)	2. Should have atleast 17" or more TFT color monitor with 12 or more waveforms displayed at a time. The waveforms should be color coded	2. Should have atleast <b>18"</b> or more TFT color monitor with 12 or more waveforms displayed at a time. The waveforms should be color coded
Page 29, clause 46.1 Add para (d) (Tender issued on 10-09-2013)		The purchaser will have a right to scrutinize validity / authenticity of the documents submitted by the tenderers

Administrative Officer  
AIIMS, Raipur