



**Tender enquiry no. ADMIN/Tender/Blood Bank Equipment/1/2014 dt. 05/12/2014**

## **Corrigendum for Blood Bank Equipment**

### **SCHEDULE – C AUTOCLAVE (Qty- one)**

<b>Page No/ Clause no/ point no</b>	<b>Existing sentence</b>	<b>To be read as</b>
Page no. 18, 19 & 20	<p>1. Chamber dimension 660 X 950 X 1524mm-Volume 900-1000 Litre.</p> <p>2. Fully automatic double door/double jacketed made 316L/ 316Ti Stainless Steel.</p> <p>3. Front loading</p> <p>4. Nickel with clad steel with sliding door.</p> <p>5. Control display with battery backup &amp; digital thermal printer with 200 cycles inbuilt Memory.</p> <p>6. Accessory must include loading carts and transfer Carriages.</p> <p>7. Electrical service 440v 50Hz 3 phase.</p>	<ul style="list-style-type: none"><li>- <b>550x750 mm (Diameter x Height). Basket Dimension 525 x 600 (Diameter x Height). Capacity 187 Lts.</b></li><li>- <b>Vertical Jacketed autoclave.</b></li><li>- <b>Pressure range 5-40</b></li><li>- <b>Inner chamber size:-550 - 650 x 350 - 450 x 350 - 450 mm</b></li><li>- <b>Joint less gasket</b></li><li>- <b>Additional accessory to be quoted separately.</b></li><li>- <b>Gasket 2 in number.</b></li><li>- <b>Heating coil 2 in number.</b></li><li>- <b>Stainless steel chamber 4</b></li><li>- <b>Stainless steel tray 2.</b></li><li>- <b>Touch screen panel LED/LCD, digital temperature and pressure display.</b></li><li>- <b>Should meet the requirements of ASME Boiler and Pressure Vessel Code.</b></li><li>- <b>Power input 220-240v Ac; European CE certified.</b></li><li>- <b>Sterility assurance consumables products and Compatible validation process. For two years included in the quotation. Bowle Dick Test Packs=100 no's; Biological Indicators=500 no's, Chemical Indicators TST Emulating; Biological Indicator activator and Incubator Machine with each unit.</b></li></ul>

Store Officer  
AIIMS, Raipur