



Tender enquiry no. ADMIN/Tender/Supply of ENT Instruments/1/2014

Corrigendum for ENT Instruments

S. No.	Clause	Existing Parameter (mentioned in last amendment)	Amendment
1.	Page No. 25, Schedule C, Point No. C (viii)	Full high definition 2 channel, simultaneous recording device, along with 500 GB internal memory, MPEG-4, AVC / H-264	Full High Definition single/dual channel recording device along with 1 TB internal memory MPEG-2/MPEG-4 recording window based (XP 4 above)
2.	Page No. 26, Schedule-C, point no. C (xii)	Trolley should be of same manufacturer as Camera System with recording.	Equipment Cart from same manufacturer, wide, high, rides on 4 antistatic dual wheels equipped with locking brakes, mains switch on cover, central beam with integrated electrical subdistributors with minimum 12 sockets, grounding plugs, Dimensions in mm (w x h x d): Equipment cart: minimum 800-850 x 1450-1500 x 700-750, With 3 Nos. of Shelf: 600-650 x 25-30 x 500-530, with one Drawer Unit with Lock, Camera Holder & isolation transformer 2000 VA, 220-240 AC, 50 Hz

Administrative Officer
AIIMS, Raipur