



Tender enquiry no. ADMIN/Tender/ICU\_Ventilator\_for\_HDU/1/2014

**Corrigendum for “ICU Ventilator for HDU”**

**ICU Ventilator (10 Nos)**

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	As mentioned in tender document	Last Date of submission 14-10-2014 at 03:00 PM. Technical Bid will be opened on the same day at 03:30 PM	<b>Last Date of submission 20-10-2014 at 03:00 PM. Technical Bid will be opened on the same day at 03.30 PM.</b>
2	Page No. 13 & 14	USFDA, European CE and ISO 9001 Certified	<b>USFDA and European CE</b>
3	Page No. 21 point no. 4 (e)	Advanced mode like Pressure Regulated volume control mode	Advanced mode like Pressure Regulated volume control mode/ <b>Volume guaranteed ventilation</b>
4	Page No. 22 point no. 7	Should have built-in ultrasonic nebulizer	<ul style="list-style-type: none"> <li>• Ultrasonic/Pneumatic nebulizer should be capable to give particle size of &lt; 3 μ.</li> <li>• Should be integrated / inline / inbuilt.</li> </ul>
5	Page No. 22 point no. 5 (j)	Inspiratory and Expiratory flow and pressure Trigger Sensitivity	Trigger Sensitivity <ul style="list-style-type: none"> <li>• Should have separate inspiratory and expiratory flow sensor.</li> <li>• Inspiratory breath in spontaneous made should be flow triggered with or without pressure triggering.</li> </ul>
6	Addendum Page no. 17	-	<ul style="list-style-type: none"> <li>• Filters – HME x 100</li> <li>• Expiratory valve/cassettes – 2 sets (Reusable )</li> <li>• Flow Sensors –10</li> <li>• Support arm for the breathing circuit from the same manufacturer.</li> <li>• Ventilator trolley from the same manufacturer.</li> <li>• O<sub>2</sub> hose – 2 nos,</li> <li>• Air hose – 2 nos.</li> <li>• Humidifier – 810 (Reusable chambers) Adult – 3 Pediatric – 2</li> <li>• Disposable Chambers &amp; Circuits Adult – 10 Pediatric – 10</li> <li>• Test Lung Adult – 1 Pediatric – 1</li> <li>• Nebulization kit Adult – 10 Pediatric – 10</li> </ul>

Deputy Director (Admin)  
 AIIMS, Raipur