



Tender enquiry no. ADMIN/Tender/Syringe_Infusion_Pumps/1/2014

Corrigendum for Syringe Infusion Pumps

Syringe Infusion Pumps (100 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page no. 12 & 13	European CE & US FDA/EN ISO Certified	European CE or US FDA
2	Page no. 20	Syringe size: 5, 10, 20, 50ml sizes	Syringe size: 10, 20, 30, 50ml sizes
3	Page no. 20	Flow rate range: 0.1 to 1500ml/hr	Flow rate range: 0.1 to 1200ml/hr or more B/W
4	Page no. 20	Alarms: Occlusion alarm with user-adjustable occlusion pressures Occlusion pressure must be displayed on screen Syringe near-empty alarm Low battery Internal malfunction Syringe wrongly fitted Power supply failure	Alarms: Occlusion alarm with user-adjustable occlusion pressures Occlusion pressure must be displayed on screen Syringe near-empty alarm Low battery Internal malfunction Syringe wrongly fitted Power supply failure Occlusion level alarms – 3 levels
5	Page no. 21	Display: Flow rate in large LCD display	Display: Flow rate in LCD display
6	Page no. 21	Audible alarm: Sound intensity user-adjustable Rechargeable, maintenance free battery Life when fully charged: up to 10 hrs	Audible alarm: Sound intensity user-adjustable Rechargeable, maintenance free battery Life when fully charged: up to 10 hrs at least 4 hrs
7	Page no. 21	Comprehensive warranty and AMC would include all parts (plastic, metallic, glass, batteries and rubber etc.) (Without any exclusion) except the consumable accessories.	Deleted
8	Page no. 22	Comprehensive warranty and AMC would include periodic checking and periodic calibration of all parameters strictly as per manufacturer's recommendations (at least every 3 month) and any spares or standards required for that.	Warranty and AMC would include periodic checking and periodic calibration of all parameters strictly as per manufacturer's recommendations (at least every 3 month) and any spares or standards required for that.
9	Page no. 22	Should have resident service engineer in Patna and should be available 24 x 7 to solve the complaints.	Should have resident service engineer in Raipur and should be available 24 x 7 to solve the complaints.



Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
10	Page no. 22	If machine is not made fully functional within 24 hours of the lodgement of complaint (even for single parameters), a standby machine of similar configuration must be provided. Failure to do so will invite an increase in the period of comprehensive warranty or AMC by 7 times of the downtime period. (i.e. 7 days for every 1 day of downtime).	If machine is not made fully functional within 24 hours of the lodgement of complaint (even for single parameters), a standby machine of similar configuration must be provided. Failure to do so will invite an increase in the period of warranty or AMC by 7 times of the downtime period. (i.e. 7 days for every 1 day of downtime).
11	Addendum Page no. 23	-	IP65 or IP55 Body should not allow ingress of liquid. Certificate should be provided.
12	Addendum Page no. 23	-	Should have Drug Library

Administrative Officer
AIIMS, Raipur