



**Tender enquiry no. ADMIN/Tender/Bronchoscope/1/2014**

**Corrigendum for supply of "Bronchoscope"**

Sr. No.	Page No/ Clause no/ point no	Corrigendum dated 01-10-2014	Amendment
1	Annexure-I <b>Bronchoscope (Adult)</b>	1. Insertion Tube outer diameter: less than 5.20 mm. Working length up to <b>540 mm</b>	1. Insertion Tube outer diameter: less than 5.20 mm. Working length <b>atleast 540 mm</b>
2	Annexure-II <b>Bronchoscope (Paediatric)</b>	2. Insertion Tube outer diameter: less than 4.20 mm. Working length <b>upto 540 mm</b> 3. Instrument Cannel Working channel inner diameter <b>2.0 mm</b> or wider. Minimum visible distance 3 mm from distal end	2. Insertion Tube outer diameter: less than 4.20 mm. Working length <b>atleast 540 mm</b> 3. Instrument Cannel Working channel inner diameter <b>1.5 mm</b> or wider. Minimum visible distance 3 mm from distal end
3	As mentioned in tender document	Last Date of submission 08/10/2014 at 03:00 PM	<b>Last Date of submission 15/10/2014 at 03:00 PM</b>

Administrative Officer  
AIIMS, Raipur