



Tender enquiry no. ADMIN/Tender/Multipara_ & Neonatal_Multipara_Monitor/1/2015 Date 04-02-2015

Corrigendum

MULTIPARA MONITOR (HIGH END) (14 Nos.) Schedule-A

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page no. 23 No. 1	Individual Monitors : <ul style="list-style-type: none"> ▪ Medical grade, TFT Flat screen, slim size, at least 19" display 	Deleted
	Page no. 23 General	Monitor must be able to connect to central monitoring station and should use single network for all kind of networking with the central station or other hospital information system (HIS). It should have dual display, one for continuous display of all beds and other for trend review, full disclosure etc.	Monitor must be able to connect to central monitoring station and should use single network for all kind of networking with the central station or other hospital information system (HIS).
2	Page no. 24 No. 2	Parameters monitored : <ul style="list-style-type: none"> ▪ EtCO₂ ▪ Minimally invasive cardiac output ▪ BIS ▪ NMT 	Parameters monitored : <ul style="list-style-type: none"> ▪ EtCO₂ ▪ Minimally invasive cardiac output ▪ BIS ▪ NMT Rate to be quoted separately
3	Page. No. 16, 26 &. 27		
4	Page. No. 26	Comprehensive warranty and CAMC would include all parts-plastic, metallic, glass, batteries and rubber except the consumable accessories listed below against item no. 22. Comprehensive warranty and CAMC would include periodic checking and periodic calibration of all parameters strictly as per manufacturer's recommendations (at least every 3 month) and any spares, gases or standards required for that.	Comprehensive warranty and CAMC would include periodic checking and periodic calibration of all parameters strictly as per manufacturer's recommendations (at least every 3 month) and any spares, gases or standards required for that.
5	Page. No. 27 & 28	Comprehensive warranty and CAMC would include all parts-plastic, metallic, glass, batteries and rubber (without any exclusion) except the consumable accessories listed above.	Deleted.
6	Addendum as point no. 11 on Page. No. 13		All parameters in the monitors should be individually USFDA approved.
7	Addendum as point no. 12 on		Monitor should have separate transport module with screen and



आरोग्यम् सुखं सम्पदा

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
	Page. No. 13		data storage and able to monitor temperature, ECG, SPO2, NIBP & atleast one IBP

NEONATAL MULTIPARA MONITOR -HIGH END(2 Nos.) Schedule-B

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Addendum as point no. 33 on Page. No. 34		Monitor should have separate transport module with screen and data storage and able to monitor temperature, ECG, SPO2, NIBP & atleast one IBP

MULTIPARA MONITOR (8 Nos.) Schedule-C

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page no. 35 Sr. No. 3	Individual Monitors : <ul style="list-style-type: none">Medical grade, TFT Flat screen, slim size, at least 17" display	Individual Monitors : <ul style="list-style-type: none">Medical grade, TFT Flat screen, slim size, at least 19" display
2	Addendum as point no. 20 on Page no. 35		Monitor should be upgradable to HIS
3	Addendum as point no. 21 on Page no. 35		Monitor should have separate transport module with screen and data storage and able to monitor temperature, ECG, SPO2, NIBP & atleast one IBP

Store Officer
AIIMS, Raipur