



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Tender enquiry no. ADMIN/Tender/Bronchoscope/1/2014

Corrigendum for supply of "Bronchoscope"

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	As mentioned in tender document	Bronchoscope Quantity 03	Bronchoscope Quantity 03 (Adult – 02 + Paediatric – 01)
2	Page no. 23 Annexure-I	Technical Specification for Bronchoscope Annexure-I	Revised Technical Specification for Bronchoscope (Adult) attached as Annexure-I
3	Addendum as Annexure-II	-	Revised Technical Specification for Bronchoscope (Paediatric) attached as Annexure-II
4	Addendum as Annexure-III	-	Revised Financial Bid attached as Annexure-III

Administrative Officer
AIIMS, Raipur



Annexure-I

BRONCHOSCOPE FOR ADULT (02 Nos.)

Sr. No.	Name of Equipment	Specifications Required for the Equipment to be purchased
1.	Bronchoscope (Adult)	<ol style="list-style-type: none">Optical system Field of view at least 100°. Depth of field 3~50 mmInsertion Tube outer diameter: less than 5.20 mm. Working length up to 540 mmInstrument Cannel Working channel inner diameter 2.0 mm or wider. Minimum visible distance 3 mm from distal endBending Section Angulations range UP 180°, DOWN 130°Compatible with diagnostic and therapeutic high frequency treatment devices like electro surgical proceduresEnhance mucosal imaging either by digital filter based contrast enhancement or narrow band imaging.Light Source Xenon or LED light at least 300W with scope compatibility with lamp life of at least 500 hoursVideo processor with scope compatibility (01 No). Video processor should have HDTV signal output, Narrow band imaging facilities/I-scan/FICE/SPICE.Medical grade monitor-at least 19 inches (01 No)Bronchoscopy video recording & reporting system, compatible software for still and live recording and report generation / 1 TB memory / minimum OS with Window-08Power Capable of operating on 220V: 50Hz ACIntegrated with the entire system on a single trolleyACCESSORIES <p>The standard set of Bronchoscope should contain the following accessories along with or otherwise these should be quoted separately. Accessories need to be supplied and approved by the original manufacturer of video bronchoscope:-</p> <ol style="list-style-type: none">Cleaning brushes 2Bite block 2Leakage tester 1Cleaning and maintenance kit 1 <p>14. Terms and conditions: Three years warranty with five years CMC. Approval of the equipment would be subject to the working demonstration at AIIMS Raipur. The company should give a certificate that the model quoted is the most recent one and should have spares and accessories available for atleast next 5 -7 years.</p> <ol style="list-style-type: none">Should supply with good quality trolley.



Annexure-II

BRONCHOSCOPE FOR PAEDIATRIC (01 Nos.)

Sr. No.	Name of Equipment	Specifications Required for the Equipment to be purchased
2.	Bronchoscope (Paediatric)	<ol style="list-style-type: none">Optical system Field of view at least 100°. Depth of field 3~50 mmInsertion Tube outer diameter: less than 4.20 mm. Working length up to 540 mmInstrument Cannel Working channel inner diameter 2.0 mm or wider. Minimum visible distance 3 mm from distal endBending Section Angulations range UP 180°, DOWN 130°Compatible with diagnostic and therapeutic high frequency treatment devices like electro surgical proceduresEnhance mucosal imaging either by digital filter based contrast enhancement or narrow band imaging.Light Source Xenon or LED light at least 300W with scope compatibility with lamp life of at least 500 hoursVideo processor with scope compatibility (01 No). Video processor should have HDTV signal output, Narrow band imaging facilities / I-scan/FICE/SPICE.Medical grade monitor-at least 19 inches (01 No)Bronchoscopy video recording & reporting system, compatible software for still and live recording and report generation / 1 TB memory / minimum OS with Window-08Power Capable of operating on 220V: 50Hz ACIntegrated with the entire system on a single trolleyACCESSORIES The standard set of Bronchoscope should contain the following accessories along with or otherwise these should be quoted separately. Accessories need to be supplied and approved by the original manufacturer of video bronchoscope:-<ol style="list-style-type: none">Cleaning brushes 2Bite block 2Leakage tester 1Cleaning and maintenance kit 1Terms and conditions: Three years warranty with five years CMC. Approval of the equipment would be subject to the working demonstration at AIIMS Raipur. The company should give a certificate that the model quoted is the most recent one and should have spares and accessories available for atleast next 5 -7 years.Should supply with good quality trolley.



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Annexure-III

**“Bronchoscope”
AIIMS, Raipur
FINANCIAL BID**

(In sealed Cover-II super scribed “Financial Bid”)

S. No.	Item Description	Qty of units	Unit Price in INR	Custom duty in INR (if applicable)	Taxes (if applicable) VAT / Sales Tax/etc in INR	Service Tax (if applicable) in INR	Total Unit Price (including all taxes)	Total price of (Unit price x Quantity)
1	Bronchoscope (Adult)	02						
2	Bronchoscope (Paediatric)	01						
	CMC Total for Five Years (Details is to be mentioned in separate Table)							
	Grand Total*:							

The grand total * will be the deciding factor for L1 provided all the other conditions mentioned in the tender document GCC, SCC and the technical conditions) are fulfilled.

DATE:

SIGNATURE WITH NAME AND SEAL