



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

**Tender Enquiry No. ADMIN/Tender/Dentistry Equipment/1/2013**

**Corrigendum**

<b>S No</b>	<b>Page no of Tender Document</b>	<b>Point</b>	<b>As in the present Tender</b>	<b>To Be changed to</b>
1	24	General Specifications		6. All the instruments/equipment should be provided with the detailed data sheet along with the original <a href="#">brochure</a> of the quoted model (Photocopies/Print outs are not accepted)
2	28	A 61	Interior furnishing with the requisite furniture, storage, cupboards, etc. for the Operatory area Size of 12 X 12 per dental Chair. Will also form an essential component of this turnkey project.	Interior furnishing with the requisite furniture, storage, cupboards, etc. for the Operator area Size of 12 feet X 12 feet per dental Chair. Will also form an essential component of this turnkey project.
3	29	B 22	Sterilizer should have inbuilt water reservoir with storage capacity up to 5 Litres.	Sterilizer should have inbuilt water reservoir with storage capacity from 3 Litres to 5 Litres.
4	34	F2	Simple, flat control panel and LCD screen	Endomotor with cord should be Simple, flat control panel and LCD screen

**Administrative Officer**  
**AIIMS, Raipur**