



- 8 -

अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road,

Raipur-492 099 (CG)

[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS-RPR/Store/LPC/Microbiology/2016/

Date: 16.09.2016

Inviting Quotations for Purchase of Non Consumable Item (i.e. LED Digital Dry Bath) for Department of Microbiology, AIIMS Raipur.

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of Non Consumable Items for Department of Microbiology, AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Stores Officer** up to 23.09.2016 before 03:00 pm. The quotations will be opened on the same day at 03:30pm. Details of item are given as under:-

S.no.	Name of the Item	Quantity Required	UNIT RATE IN Rs.	Total amount in Rs
01	LED Digital Dry Bath (Hot Block)	01 No.		

**Technical specification of Digital Dry bath (Heat Block)**

1. Microprocessor digital control panel.
2. Heating temperature range +30- 120°C.
3. Temperature resolution 0.1°C.
4. Temperature uniformity  $\pm 0.2^\circ\text{C}$
5. Control accuracy  $\pm 0.5^\circ\text{C}$ . Should provide precise, accurate control over user desired temperature.
6. Advanced internal temperature sensing probe for outstanding temperature accuracy and control.
7. Should be ideal for molecular biology work of DNA denaturation.
8. Should be provided with single blocks of stainless steel which should have  $\geq 20$  places for holding both 1.5 ml and 2 ml microcentrifuge tubes.
9. All the samples should receive the same temperature treatment regardless of their position in the block.


10. Should have timer.
11. 4.3' touch screen display, operating voltage 200-240 V/ frequency 50Hz.
12. Easy user calibration.
13. Time setting range 1-99h59min.
14. Operation type Continuous/timed operation
15. In-built over temperature protection.
16. Protection class according to DIN EN60529.
17. CE/ISO certified.
18. One year warranty.
19. All technical specification will be evaluated with on-site satisfactory demonstration of the equipment with compliance to all the enlisted technical features.

**Term & Conditions:**

1. Rate should be mentioned in word and figures both.
2. Taxes, if any (Should be clearly mentioned).
3. Delivery Schedule – within 21 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (i.e. Concerned Department).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of envelop.
7. VAT Reg. (i.e. Form 11) must be attached with the quotation.
8. 100% payment against receipt and acceptance of material.
9. Brand and make should be clearly mentioned in offer. *without this offer may be summarily rejected*
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firm.
11. Validity of offer should not be less than 90 days.
12. *Installation and commissioning will be done by supplier (if applicable)*

  
16.09.16



  
16/09/16  
Store Officer  
AIIMS, Raipur (C.G.)