



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/HS/Dentistry/17/16867/LPC

Date: 16.01.2017

Inviting Quotations for Purchase of Capital Asset items for the Department of Dentistry OPD, AIIMS Raipur.

### QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase

of LED Gooseneck Light for department of Dentistry OPD. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer Room No. – 51, 2nd floor Ayush Building, Gate No. 1 up to 19.01.2017** before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

| S.No | NAME  | Qty. | UNIT | MAKE | UNIT RATE IN Rs. |
|------|---|------|------|------|------------------|
| 1    | LED gooseneck light<br>Specification: <ul style="list-style-type: none"><li>Operate on 220-240 V AC with the ability to tolerate moderate voltage fluctuation.</li><li>Weight approx. 5kg.</li><li>Should be mounted on lockable wheels for portability.</li><li>Base should be narrow to permit use in room of compact dimensions.</li><li>No stroboscopic effect when taking pictures with LED light as source of illumination.</li><li>Power cord about 10 feet length with earthing if metal body.</li><li>Flexible neck should permit light to be about 4 feet above floor level.</li><li>Should permit spray-and-wipe disinfection.</li><li>Must be water resistant to withstand minor spills of water.</li><li>Stem/neck should be flexible to permit easy and precise manipulation of direction of light beam and sufficiently rigid to hold this position without sagging</li><li>Silent operation.</li><li>Lamp life 15000 hours or more.</li><li>Should have a warranty of 5 years.</li><li>Illumination intensity 40000 Lux or higher.</li><li>Cold LED lamp of approximately 3 watt rating with no heat in direction of light beam.</li><li>Should have a brightness control.</li><li>On/off switch should be durable and be capable of thousands of operations without fault.</li><li>Beam of light should have an adjustable spot diameter.</li><li>Should have an attractive and modern design.</li></ul> | 02   | Nos. |      |                  |

**Note:-5 Years warranty**

## **Terms & Condition**

1. Firm to mention Manufactures name/Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. Tin number and relevant document should be submitted along with quotation.
4. Supply should be done within 15 days after Placement of PO.
5. Price should be FOR Destination basis (i.e. concerned department).
6. Payment will be released after certification from Dentistry department.
- 7. Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
10. Material to be delivered at **Gate No-01, Hospital Store, AYUSH PMR Building, 2<sup>nd</sup> Floor, Room No. - 56.**
- 11. Validity of quotation should be 90 days from the date of opening**

**Stores Officer (H)**  
AIIMS, Raipur  
(C.G.)