



No.: AIIMS-R/Str/LPC/Bair_Hug_Temp_Mngt_Unit/2016/1/036L

Date: 03/06/2016

QUOTATION NOTICE

Inviting Quotations for procurement of "Bair Hugger Therapy Temperature Management Unit" for Department of Trauma & Emergency at AIIMS Raipur.

Sealed quotations are invited from intending registered Stockiest/Distributors having TIN and relevant documents for "Bair Hugger Therapy Temperature Management Unit". Quotation should be submitted to office of the Store Officer up to 08.06.2016 before 3:00 pm. The quotations will be opened on the same day at 3.30 pm. Details of item are given as under:

S. no.	Particular	Qty.	Unit Price in ₹	VAT/ Tax	Total in ₹ (incl. of tax)
1	<p>Bair Hugger Therapy Temperature Management Unit</p> <ul style="list-style-type: none">▪ Dimensions: (13"-15"h×14"-15"w×13"-14"d)▪ Weight: (15-16 lb),▪ Operating Temperatures:<ul style="list-style-type: none">High: 43°±1.5°C(109.4°±2.7°F)Med: 38°±1.5°C(100.4°±2.7°F)Low: 32°±1.5°C(89.6°±2.7°F)Ambient: (Room Temperature)▪ Fan Speeds and relative Noise:<ul style="list-style-type: none">High Fan Setting: 37 CFM (17.5 L/s) 53 dBALow Fan Setting: 32 CFM (15.1 L/s) 48 dBA▪ Leakage current: Meets UL 60601-1 and IEC 60601-1 requirements. CAN/CSA-C22.2, No. 601.1▪ Safety Systems: Audible and visual alarms;<ul style="list-style-type: none">Over-temperature ≤ 56°C, 53°C ± 3°C typical▪ Filter: High-efficiency 0.2 μm filter▪ Device Ratings:<ul style="list-style-type: none">110-120 VAC, 50/60 Hz, 11.7 Amperes220-240 VAC, 50/60 Hz, 7.2 Amperes100 VAC, 50/60 Hz, 14.5 Amperes▪ Compatibility: Compatible with all Bair hugger blanket styles-pediatric to adult and bair paws gowns <p>Specification of Bair Hugger Therapy full body postoperative blankets (Adult-1 & Pediatric-1)</p> <ol style="list-style-type: none">1. Underbody design Should provide full, unrestricted access to the patient and accommodates lateral, supine and lithotomy position	01			



<ol style="list-style-type: none">2. Fluid outlets should minimize pooling of fluids on the surface of the blanket3. Pass-through slit design allows for flexible patient positioning and accommodates the use of a draw sheet.4. Generously-sized peritoneal cut-out at the base of the blanket Should provide clinicians unobstructed access to patient5. The blanket draws up next to patients when inflated to surround them in warmth6. Blankets should be soft, radiolucent, latex-free materials. Foot drape minimizes the risk of thermal injury to the feet and lower leg area7. Integrated tuck flaps at shoulders help maintain blanket position				
--	--	--	--	--

Terms & Condition

1. Taxes, if any (Kindly mention in above table)
2. Document relating to registration of firm i.e. Tin number should be submitted along with quotation.
3. Supply should be done within 2 weeks after making instruction from concerned authority of store.
4. Price should be for –Destination basis.
5. Payment will be release after certified from store department and concerned department.
6. Quotation Name/No. must be mentioned on top of envelops.
7. **Warranty period should be 2 years.**


Store Officer,
AIIMS Raipur
Store Officer
AIIMS, Raipur (C.G.)