



आरोग्यम् सुखं सम्पदा

अखिल भारतीय आयुर्विज्ञान संस्थान(छत्तीसगढ़) रायपुर ,
All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh,
Raipur-492 099 (CG)

No. AIIMS- RPR/Store/P/Pathology/Centrifuge/2016

Dated: 16/11/2016

Notice for inviting Quotation for supply of Laboratory Centrifuge at AIIMS, Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of the following items to be used in Pathology & Lab Medicine at AIIMS, Raipur and should be submitted to **Medical College Building, office of Store Officer, Gate No. 5, AIIMS Raipur** up to 3:00 pm on 22nd November 2016, item description as per detailed below.

| Sl. No. | Item Description | Unit | Qty. | Unit Rate in Rs. |
|----------------|--|-------------|-------------|-------------------------|
| 1. | <u>Laboratory Centrifuge</u> Specifications for Table Top Laboratory Centrifuge <ol style="list-style-type: none">Should be Table Top Version with swing out rotor head.Should have a minimum tube capacity of 16 x 15 ml.Should be made up of strong fabricated & corrosion resistant steel and have stainless steel centrifuge chamber, easy to clean.Should have control panel-for start/stop switch, dynamic brakes, step less speed regulator with zero start switch protective fuses.Should have smooth & soft start. Inverter fault detection with auto shutdown.Should have safety lid interlock to prevent lid opening during centrifugation & automatic door opening on completion.Should have maintenance free brushless induction motor with frequency drive.Should have Microprocessor controller with digital display.Max. Speed rang: RPM – 100 to 4500 or above, accuracy 1 RPM, RCF (g) – 3485 or above.Power supply: power input: 220-240 V AC, 50Hz.Should have 7 segment LED display of speed, digital countdown timer (Digital timer range 0-59 minutes) & continuous run.Should have last set parameters recall, useful for repetitive analysis.Should have imbalance detection & centrifugation stop with display of error.Should have motor overload protection.Should have gas hinge to prevent door falling.Should have emergency lid lock release.Dimension should be approximately (W x D x H) 380x470x300 MM.Should have ISO 9001:2008 & ISO 13485:2003 certification.Should have WHO, Good Manufacturing Practices & CE certification.Calibration should be performed on supply of instruments & Calibration Certificate provided.Company should provide 2 years warranty. | Set | 01 | |

OTHER COMMERCIAL TERMS

1. Rate should be mentioned in word and figures both.
2. Taxes, if any (Should be clearly mentioned).
3. Delivery Schedule – within 15 days from the date of issue of PO.
4. Price should be F.O.R. Destination basis (i.e. Concerned Department).
5. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Quotation No/Name and Due date of opening must be written on top of the envelop.
7. VAT Reg. (i.e. From 11) must be attached with the quotation.
8. 100% payment against received and acceptance of material.
9. RTGS detail required for payment purpose.
10. Brand and make should be clearly mentioned in offer.
11. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firm.
12. Validity of offer should not be less than 90 days.

**Stores Officer
AIIMS, Raipur (CG)**