

अखिल भारतीय आयुर्विज्ञान संस्थान(छत्तीसगढ़) रायपुर, All India Institute of Medical Sciences, Raipur (Chhattisgarh) Tatibandh, Raipur-492 099 (CG)

AIIMS- RPR/ Store/Purchase/ Non-Consumable /2016

Dated: 21/04/2016

Notice for inviting quotation for supply of Non-Consumable Item.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of the following items to be used in Department of Pathology & Lab Medicine, at AIIMS, Raipur and should be submitted to office of Store Officer up to 3:00 pm on 28th April 2016, item description as per detailed bellow.

SI. No.	Item Description	Unit	Required Quantity	Unit Rate in ₹
1.	 MAGNETIC STIRRER- SINGLE PLACE Should have no moving parts Should have digital display with real time and speed. Should have adjustable Speed from 50 to 1200 rpm at every 50 rpm Should be microprocessor controlled with last run memory feature Should have Stirring cap. 650 ml or more – set up place of 100 mm Should have pulse mode to reserve the direction every 30 secs for proper mixing Should be IP 65 complaint design (Water spill resistant) Should have corrosion resistant SS top surface. Should be supplied with 1 (one) high strength stir bar. Company should provide 2 years warranty 	Set	01	
2.	 MAGNETIC STIRRER- FOUR PLACE Should have no moving parts. Should have Digital display with adjustable speed & time. Should have stirring cap. 400 ml - set up place diameter of 85 mm. Should have software driven gradual acceleration. 	Set	01	

				1
	 Should have speed setting from 25 to 1000 RPM & time from 1 to 99 min & continuous mode. Should have pulse mode to reserve the direction every 30 secs for proper mixing. Should have microprocessor controlled with last run memory feature. Should be corrosion resistant SS top surface. Should be IP 65 complaint design (Water spill resistant) Should be supplied with 1 (one) high strength stir bar Company should provide 2 years warranty 			
3.	 Medical Centrifuge Should Have Fixed Angle Rotor (Cap. 6*10/15ml). Should Have Brushless DC motor for maintenance free long life. Should be Microprocessor based Centrifuge with large digital display. Should have Precise RPM setting from 500-6000 RPM in steps of 100 RPM. Should have Timer setting 1 to 30 minutes & infinite mode. Should accommodate most common collection tubes should come with adaptors for smaller 5ml, 7ml, 9/10 ml tubes. Should Have Imbalance detection safety with auto cut off. Should Have Lid Lock safety feature during operation - Lid should release automatically after run completion. Should have one switch change over from RPM to RCF. Company should provide 2 years warranty. 	Set	02	

OTHER COMMERCIAL TERMS

- 1. Taxes, if any (Kindly Mention).
- 2. Delivery Schedule within 10 days from the date of issue of PO.
- 3. Price should be for Destination basis.
- 4. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
- 5. Quotation No/Name and Due date of opening must be written on top of the envelop.

Sr. Administrative Officer AIIMS, Raipur (CG)