



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

H.Store/LPC/Suction machine/2016/AIIMS-RPR/01

Date: 01/11/2016

Inviting Quotations for Purchase of Suction machine for 200 beds IPD at A, A1 block AIIMS Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of suction machine for IPD at A, A1 block AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Stores Officer, Room no. 51, Ayush PMR Building, 2nd floor, Gate no.-1 AIIMS, Tatibandh, Raipur up to 07/11/2016 before 03:00 pm.** The quotations will be opened on the same day at 03:30pm. Details of item are given as under:-

S.No	Name of Item	Required Quantity	Unit Price ₹	Vat/Tax %	Total Value ₹
01	Suction Machine (Description attached in Annexure A)	08			

Note: - Sample of machine to be provided at the time of submission of offer.

Terms & condition

01. Firm to mention Manufactures name/make/ brand Name in their quotation.
02. Taxes, if any (kindly mention in above table) should be clearly mentioned in the offer.
03. Document relating to registration of firm i.e. Tin number/all relevant document should be submitted along with quotation.
04. Supply should be done within 15 days after Placement of PO.
05. Price should be FOR Destination basis (i.e. concerned department)
06. Payment – 100% payment will be released after certification from Medical Superintendent.
07. Quotation name /No. & due date of opening must be mentioned on top of envelops.
08. LD @ 0.5% of delayed supply per week or part of week for delay of supply of Material subject to maximum up to 10% of delayed supply should be deducted.
09. AIIMS Raipur reserves the right to increase/decrease the number of required quantity, on one more firms.
10. All other terms & condition as per GCC applicable.
11. **Material to be delivered at Gate No-01, Hospital Store, AYUSH PMR Building, 2nd Floor, Room No- 56.**
12. Validity of the offer should not be less 90 days from the date of opening.

Stores Officer (H)
AIIMS, Raipur (C.G)

Annexure-A

Technical Specification for Suction Machine:

1. The company shall be ISO certified.
2. The unit shall be CE/FDA certified unit.
3. The company shall have a local service centre based at Raipur or anywhere in Chhattisgarh. Please provide address of service centre.
4. The company/supplier shall have experience in supplying/sale of above items in govt. sector.
5. Absolutely noiseless.
6. Complete plastic panels, hand switch and foot switch for operation easily.
7. Heavy duty motorised pump-for fast vacuum built up.
8. Over flow safety device over flow check valve and removable filter to protect the pump from moisture.
9. Selected negative adjustment and overflow protective mechanism, safe and reliable.
10. Machine should have 2 polycarbonate jars (autoclavable) with over flow protection system with bacterial filter.
11. Operating voltage 110-230 volts Ac.
12. Should be UPS inverter friendly.
13. Max vacuum approx. 715 \pm 10mm Hg regulable through vacuum control knob.
14. Max vacuum 20lits/min. vacuum gauge bourdon type 3" 0-760mm Hg calibration.
15. Suction jars of capacity 2000-2500ml made up of polycarbonate.
16. Motor $\frac{1}{4}$ Hp vacuum pump rotatory type overall size L330 X W405mm X H730mm (approx.).
17. Demonstration and training should be provided by the supplier at the time of installation.
18. Warranty should be at least 2 years.
19. Additional fuse should provide.
20. The machine should be portable on castors/mobile trolley.
21. At least 5 spare filter with each machine.
22. 1 polycarbonate jar extra with each machine.
23. Sample of machine should be provided at the time of quotation submission.