



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

Admin/ Store/Purchase/Capital Asset/Microbiology /2016-AIIMS.RP Date: 12.04.2016

Inviting Quotations for Purchase of Capital Asset items for department of Microbiology, AIIMS Raipur.

**QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of Capital Asset items for department of Microbiology, AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer** up to 19/04/2016 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.No.	Complete Description of items	Quantity Required	Purpose	UNIT RATE IN Rs.
1	Membrane Filter assembly <b>Specification</b> <ul style="list-style-type: none"><li>Made of borosilicate glass with fritted glass filter base.</li><li>The base design should have an integral vacuum connection located above the filtrate drip to prevent contamination of the vacuum line from droplets.</li><li>Anodised aluminium spring clamp</li><li>The filter size is 47 mm</li><li>Should be able to hold membrane filter of 47 mm diameter.</li><li>Should be autoclavable</li><li>Compatible for filtration of water samples &amp; other liquids.</li><li>Funnel capacity 1000 ml.</li><li>The funnels should have graduations/scale to measure the quantity.</li><li>Flask capacity 1000ml.</li></ul>	01	Required for water sampling and processing under Infection control section	

<b>S.No.</b>	<b>Complete Description of items</b>	<b>Quantity Required</b>	<b>Purpose</b>	<b>UNIT RATE IN Rs.</b>
2.	Vacuum Pump	01	Required for water sampling and processing under Infection control section	
	<p style="text-align: center;"><b>Specification</b></p> <ul style="list-style-type: none"> <li>• Should be compact and light.</li> <li>• Should have Aluminium die cast construction.</li> <li>• Should be Oil Free , Piston Powered.</li> <li>• Should be noise free.</li> <li>• Should be Compatible to establish connection with glass membrane filter assembly.</li> </ul>			

- 1) Taxes, if any (Kindly mention the rate.)
- 2) Delivery Schedule – Within 15-21 days from the date of issue of PO
- 3) Price should be for– Destination basis
- 4) LD @ 0.5% per week maximum upto 10% will be deducted for delay in supply.
- 5) Quotation Name/No. should be mentioned on top of envelopes.

Store Officer  
AIIMS, Raipur (C.G.)