



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर(छत्तीसगढ़)

All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Admin/ Store/Purchase/Non-Consumable/Anaesthesiology /2016-AIIMS.RPR
Date: 20.07.2016

Inviting Quotations for Purchase of Non-Consumable items for department of Anaesthesiology, AIIMS Raipur.

QUOTATION NOTICE


Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of Non-Consumable Items for department of Anaesthesiology, AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Store Officer** up to 25/07/2016 before 3:00 pm. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

S.No.	Item Description	Qty Required	Purpose	Unit Rate
1	Peripheral Nerve Stimulator (PNS) Specification 1. Amplitude settings: Digital and Analog dial display of the stimulating voltage. 2. Display Size should be W*H in mm: 50*20 3. Display Type: Graphic/Standard/ Custom LCD. 4. Current : 0-5 mAmp for cannuls mode and 0-6- mA for percutaneous nerve mapping mode. 5. Max Voltage Output : 60-90 V 6. Max Output load (Impedance) Nominal/Max: 12/17kOhm (5mAmp) 7. Impulse Duration : 0.05,0.1,0.2,0.3,0.5, 1.0 ms. 8. Stimulus Frequency : 1Hz, 2Hz. 9. Should have a display of patient current.	One	Teaching and training purpose.	

P.T.O.

	<p>10. Should have a display of set current.</p> <p>11. Should have a display of Impulse Duration.</p> <p>12. Alarm signal : Should be alerted by a change in beep tone.</p> <p>13. Menu for setup and features should be there.</p> <p>14. PERCUTANEOUS NERVE MAPPING: Facility should be there in the standard equipment.</p> <p>15. Remote Control: Optional .</p> <p>16. Size Up to H * W* D (mm) : 160*90*40.</p> <p>17. Stimulation switchable between cannula and percutaneous stimulating handle for nerve mapping.</p> <p>18. Should have low power consumption without power shutoff.</p> <p>19. Coded connector Phy on the case.</p> <p>20. Should have inbuilt battery.</p>			
--	---	--	--	--

- 1) Taxes, if any (Kindly mention the rate.)
- 2) Delivery Schedule - Within 15 days from the date of issue of PO
- 3) Price should be for- Destination basis
- 4) LD @ 0.5% per week maximum upto 10% will be deducted for delay in supply.
- 5) Quotation Name/No. should be mentioned on top of envelopes.



Sr. Administrative Officer
AIIMS, Raipur (C.G.)