



No. AIIMS, RPR/Store/Purchase/Suction Machine/2015/

Date: 19.01. 2016

Inviting Quotations for Supply of Suction Machine for Ward & Hospital, AIIMS, Raipur.

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockiest/ Distributors having TIN/ relevant documents for supply of Suction Machine for Ward/Hospital, AIIMS, Raipur and should be submitted to office of Store Officer up to 3:00 pm on 28.1.2016. The quotations will be opened on the same day at 3:30 pm. As per detailed below:-

S.N.	Item Description	Qty. Req.	Unit	Unit Price
1	Suction Machine (Electrically Operated) Should be FDA/CE certified	06	Each	
	<u>Technical Specification for Suction Machine</u>			
	The company shall be ISO certified.			
	The unit shall be CE/FDA certified unit.			
	The <u>company</u> shall have a local service centre based at Raipur or anywhere in Chhattisgarh, so please provide address of service centre.			
	The company/ supplier shall have experience in supplying/sale of above items in government sector.			
	Absolutely noiseless.			
	Complete plastic panels, hand switch and foot switch for operation easily.			
	Heavy duty motorised pump-for fast vacuum built up.			
	Over floe safety device over floe check valve and removable filter to protect the pump from moisture.			
	Selected negative adjustment and overflow protective mechanism, safe and reliable.			
	Machine should have two polycarbonate jars (Autoclavable) with over flow protection system with bacterial filter.			
	Operating voltage 110-230 volts Ac.			
	Should be UPS inverter friendly.			
	Max vacuum approx. 715 ± 10 mm Hg regulable through vacuum control knob.			
	Max vacuum 20Ltrs. /minute vacuum gauge bourdon type 3'' 0-760mm Hg calibration.			
	Suction jars of capacity 2000-2500ml made up of polycarbonate.			
	Motor $\frac{1}{4}$ Hp vacuum pump rotatory type overall size: (L) 330 x (W)405mm x (H)730mm (approx.)			
	Demonstration and training should be proved by the supplier at the time of installation.			
	Warranty should be at least 2 years, against manufacturing defects.			
	Additional fuse should provide.			
	The machine should be portable on castors/mobile trolley.			

	At least five spare filter with machine.
	One polycarbonate jar extra with each machine.

Terms & Conditions

1. Taxes, if any (Kindly Mention).
2. Demonstration is must before supply.
3. Document relating to registration of firm i.e. should be submitted along with quotation.
4. Delivery Schedule – within 7 to 15 days from the date of issue of PO.
5. Price should be for Destination basis.
6. LD @ 0.5% of delayed supply per week will be deducted for delay of supply of material subject to maximum up to 10% of delayed supply.
7. Quotation No/Name and Due date of opening must be written on top of envelop.

Store Officer
AIIMS, Raipur (C.G.)