



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर(छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

AIIMS/R/HS/Radio/2017/435/LPC/2

Date: 11/04/2017

Inviting Quotations for Purchase of Suction machine for Radio diagnosis department, at AIIMS Raipur.

### **QUOTATION NOTICE**

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Purchase of suction machine for Radio diagnosis department AIIMS Raipur. The quotation with copy of certificate of TIN & other documents should be submitted to office of **Stores Officer, Room no. 51, Ayush PMR Building, 2<sup>nd</sup> floor, Gate no.-1 AIIMS, Tatibandh ,Raipur up to 18/04/2017 before 03:00 pm**. The quotations will be opened on the same day at 03:30pm. Details of item are given as under:-

S.No	Name of Item	Required Quantity	Unit Price ₹	Vat/Tax %	Total Value ₹
01	Suction Machine (Description attached in Annexure A)	06			

**Note: - Sample of machine to be provided at the time of submission of offer.**

#### **Terms & condition**

01. Firm to mention Manufactures name/make/ brand Name in their quotation.
02. Taxes, if any (kindly mention in above table) should be clearly mentioned in the offer.
03. Document relating to registration of firm i.e. Tin number/all relevant document should be submitted along with quotation.
04. Supply should be done within 15 days after Placement of PO.
05. Price should be FOR Destination basis (i.e. concerned department)
06. Payment – 100% payment will be released after certification from Department of Radio diagnosis .
07. Quotation name /No. & due date of opening must be mentioned on top of envelops.
08. LD @ 0.5% of delayed supply per week or part of week for delay of supply of Material subject to maximum up to 10% of delayed supply should be deducted.
09. AIIMS Raipur reserves the right to increase/decrease the number of required quantity, on one more firms.
10. All other terms & condition as per GCC applicable.
11. **Material to be delivered at Gate No-01, Hospital Store ,AYUSH PMR Building, 2<sup>nd</sup> Floor, Room No- 56.**
12. Validity of the offer should not be less 90 days from the date of opening.

**Stores Officer (H)**  
AIIMS, Raipur (C.G)

## **Annexure-A**

### **Technical Specification for Suction Machine:**

1. Volt- 230 VAC.
2. Rating of motor should be with oil emulsion
3. Suction bottle should be made of polycarbonate material with a capacity-2 x 2000 ml minimum ( with safe valve)
4. Gauge-0 to 760 mm Hg
5. Pump -oil lubricate rotary pump.
6. Suction tubing- ID 7mm, 5m long and Non collapsible
7. Should have air tight lids.
8. Should have noise less operation
9. Should provide filter to absorb moisture & water particles entering into the rotor.
10. Should have an external provision for topping up lubricant.
11. Should be well designed cabinet made of mild steel powder coated.
12. Should bear ISI mark.