



Date: - 09/11/2016

Corrigendum

Tender Enquiry No.: Store_/Tender_RC_/CSSD_/1/2016

Page/Clause / Point	Existing Sentence	To be read as
Page no. 16 & 17. Point no. 6.	Alkaline Detergent Product should have biodegradable, ecofriendly, non-irritant to skin, saliva and eyes. Safe for surgical instruments. Chemical should have good cleaning quality along with multi enzymatic solution. Chemical should be low foaming based and will be used to both of ultrasonic and washer disinfectant.	Alkaline Detergent Product should have biodegradable, ecofriendly, non-irritant to skin, saliva and eyes. Safe for surgical instruments. Alkaline Detergent should be multi enzymatic solution of four enzymes i.e. protease, amylase, lipase and cellulase with pH range not exceeding 9 . Chemical should be low foaming based and must be compatible to both ultrasonic and automated washer disinfectants. Dilution factor need to be mentioned per liter for highly soiled items.
Page no. 17 Point no. 10.	Biological Indicator It can used for organism test, can be available in self-contained vials. Organism pollution rate should be 10. The organism group should have Geobacillious Stearothermophilus with proper indication of lot no. expiry date, processed chemical indicator and certified according by ISO and EN.	Biological Indicator It can be used for organism test, can be available in self-contained vials. Organism pollution rate should be 10⁶ . The organism group should have Geobacillious Stearothermophilus with proper indication of lot no. expiry date, processed chemical indicator and certified according by ISO and EN.
Page no. 1 Serial no. 5. and as mentioned in tender document	Last Date of Submission of Tender 11.11.2016 at 3:00 PM	Last Date of Submission of Tender 18.11.2016 at 3:00 PM & Opening at same date at 3:30 PM

Store Officer
AIIMS Raipur (CG)