



Tender enquiry no. Store/Tender/Arterial Blood Gas, Shadowless Lamp & Trans Venous Pacemaker/3/2015

Corrigendum for
“Arterial Blood Gas, Shadowless Lamp & Trans Venous Pacemaker”

Schedule-A Arterial Blood Gas Analysis Machine (ABG Machine)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page No. 18 Point No. 4	Sample volume-less than 200 micro litre.	Sample volume-less than 100 micro litre.
2	Page No. 18 Point No. 4	Fast analysis time – less than 150 sec.	Fast analysis time – less than 120 sec.
3	Page No. 18 Point No. 11	Built in auto Quality control facility.	Deleted.

Schedule-B Shadowless Lamp Stand Model

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page No. 19	<ul style="list-style-type: none">▪ One main halogen bulb and one reserve halogen bulb.▪ Reserved halogen bulb switched ON automatically in case of failure of main bulb.▪ Halogen bulb positioned at the centre of the heat-absorbing colour correcting filter.▪ Replacement of halogen bulb quick and without using any tool.	<ul style="list-style-type: none">▪ One main LED bulb and one reserve LED bulb.▪ Reserved LED bulb switched ON automatically in case of failure of main bulb.▪ LED bulb positioned at the centre of the heat-absorbing colour correcting filter.▪ Replacement of LED bulb quick and without using any tool.

Store Officer
AIIMS, Raipur



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in