



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Tender enquiry no. AIIMS-RPR/Store/Tender/Bronchoscope/2016/1

date 10-06-2016

Corrigendum

| Clause | Existing Parameter | To be read as |
|--------------------------|---|---|
| In last corrigendum | Last date of submission of technical bid is 13-06-2016 at 3:00 and opened in same date at 3:30 PM. | Last date of submission of technical bid is 16-06-2016 at 3:00 and opened in same date at 3:30 PM. |
| <u>Schedule-B</u> | | |
| In last corrigendum | Compatible 300 Watt LED/Halogen light source with coloured temperature around 6000 kelvin and Led lamp as auxiliary / back up. | Compatible 300 Watt Xenon light source with coloured temperature around 6000 kelvin and LED/Halogen lamp as auxiliary / back up. |
| <u>Schedule-D</u> | | |
| In last corrigendum | Insertion Tube outer diameter: less than 4.20 mm. Working length atleast 540 mm | Insertion Tube outer diameter: less than 5.20 mm. Working length atleast 540 mm |

Store officer
AIIMS, Raipur