



Tender enquiry no. AIIMS/Store/Tender/ Hospital/Furniture/2016/1

date 25-04-2016

Corrigendum

Clause	Existing Parameter	Amendment
As mentioned in tender document	Last date of submission of technical bid is 27-04-2016 at 3:00 and opened in same date at 3:30 PM.	Last date of submission of technical bid is 06-05-2016 at 3:00 and opened in same date at 3:30 PM.
Page no. 7 point no. 32	In case asked tenderer must personally give the demonstration of the manikin/simulator to the competent authority of the institute and in that case all the demonstration expenses will be borne by the supplier. In case, the supplies are found defective at any stage, the tenderer will have to replace defective material with standard material. Any cost incurred by the institute to meet the eventuality of defective supplies shall also be recoverable from the tenderer. Institute may also take action under Clause 4 for Terms & conditions of the tenders well.	In case asked tenderer must personally give the demonstration of all items to the competent authority of the institute and in that case all the demonstration expenses will be borne by the supplier. In case, the supplies are found defective at any stage, the tenderer will have to replace defective material with standard material. Any cost incurred by the institute to meet the eventuality of defective supplies shall also be recoverable from the tenderer. Institute may also take action under Clause 4 for Terms & conditions of the tenders well. If necessary plant visit will be done by committee, expenses will be borne by the supplier.
Page no. 19 point no. i)	ISO certified / product should be ISI marked from the Manufacturer Company for applicable schedules.	ISO 9001:2008, ISO 14001:2004 certified manufacturer and product should be ISI or European CE marked from notified body. (applicable for all furniture items)
Addendum as point no. 1 n. in page no. 19		Certificate for 304 grade stainless steel where ever applicable.
Addendum as point no. 1 o. in page no. 19		Certificate for thickness of powder coating where ever applicable.
Page no. 24 point no. i)	<u>Fowler Bed, Four sections with mattress –</u> <ul style="list-style-type: none"> Back-rest and knee rest should be maneuverer by the screw mechanism, welded with approx. 30-36 mm dia x16 G ERW M. S. Tube for linear movement in a approx. 32-38mm dia x 16 GERW M. S. covers. 	<u>Fowler Bed, Four sections with mattress –</u> <ul style="list-style-type: none"> Back-rest and knee rest should be maneuverer by the screw mechanism (with torque release) welded with approx. 30-36 mm dia x16 G ERW M. S. Tube for linear movement in a approx. 32-38mm dia x 16 GERW M. S. covers.

Clause	Existing Parameter	Amendment
	<ul style="list-style-type: none"> • Detachable Head and Leg Bows: Should be made from stainless steel 30-34 mm dia x 18G tube with laminated panels of approx. size 810-820mm L x 140-150 mm wide x14-16 mm thick on stainless steel bracket. • The base frame should be fitted with non-rusting swivel castor wheels 125-130 mm diameter, 2 with brake, 2 without brakes, cross legs. • A mattress thickness 100mm minimum suitable for the bed made thick PU foam 40 density suitable mattress covered with good quality Rexine. • Provision of MS/SS two hooks to attach bags • Bed head panel with O2 outlet-1 no & High pressure flexible tube 3mtr & provision for one side fit in oxygen flow meter, one side fit in O2 outlet.(for 10 % of bed) 	<ul style="list-style-type: none"> • Detachable Head and Leg Bows: Should be made from stainless steel (grade 304) 30-34 mm dia x 18G tube with laminated panels of approx. size 810-820mm L x 140-150 mm wide x14-16 mm thick on stainless steel bracket. • The base frame should be fitted with non-rusting swivel twin wheel castor wheels 125-130 mm diameter, 2 with brake, 2 without brakes, cross legs. • A mattress thickness 100mm minimum suitable for the bed made thick PU foam 40 density suitable mattress covered with good quality PVC Rexine. • Provision of SS two hooks to attach bags • This point is deleted. <p>Note: All other points will be remain same.</p>
Addendum as new bullet in Page no. 24 point no. i)		<ul style="list-style-type: none"> • Suitable good quality buffers at four corners of bed.
Addendum as compliance sheet		Compliance sheet is attached as annexure-I
Page no. 18 Form D point no. 2	2. Demand Draft for EMD for below mentioned schedule(s) refundable EMD is enclosed. List as per tender	2. Demand Draft for EMD for below mentioned schedule(s) refundable EMD is enclosed. Revised List is attached as annexure-II
Addendum as point no. 40. c) viii Page no.9		viii. Small Scale Industrial (SSI) are exempted to submit the EMD (copy of registration must be provided along with)
Addendum as point no. 41. a) viii Page no.9		viii. Small Scale Industrial (SSI) are exempted to submit the tender fee (copy of registration must be provided along with)

Store Officer,
AIIMS, Raipur

S.N.	Items with specification	Compliance Yes / No	Deviation
1	<u>Bed side locker:</u> Specification as per tender / corrigendum.		
2	<u>BEDSIDE SCREEN (Ward Screen) Stainless Steel with Single Curtain, three section:</u> Specification as per tender / corrigendum.		
3	<u>Fowler Bed, Four sections with mattress -</u> Specification as per tender / corrigendum.		
4	<u>Crash cart –</u> Specification as per tender / corrigendum.		
5	<u>Dressing trolley-</u> Specification as per tender / corrigendum.		
6	<u>Over bed table/Food trolley:</u> Specification as per tender / corrigendum.		
7	<u>Stainless steel operation instruments trolley(Without bucket ring)-</u> Specification as per tender / corrigendum.		
8	<u>Kick Bucket:</u> Specification as per tender / corrigendum.		
9	<u>Medicine trolley-</u> Specification as per tender / corrigendum.		
10	<u>Mayo operation trolley-</u> Specification as per tender / corrigendum.		
11	<u>Patient examination table-</u> Specification as per tender / corrigendum.		
12	<u>Stool SS, fixed-</u> Specification as per tender / corrigendum.		
13	<u>Stool SS, Round & revolving top, adjustable for heights –</u> Specification as per tender / corrigendum.		
14	<u>Soiled linen trolley-</u> Specification as per tender / corrigendum.		
15	<u>WHEEL CHAIR (SS) –</u> Specification as per tender / corrigendum.		
16	<u>Wheel chair folding-</u> Specification as per tender / corrigendum.		

Annexure – II

Schedule	Bed side locker	EMD	DD no.	Date	Bank
1	Bed side locker	₹ 28,350/-			
2	BEDSIDE SCREEN (Ward Screen) Stainless Steel with Single Curtain, three section	₹ 47,250/-			
3	Fowler Bed, Four sections with mattress	₹ 220,500/-			
4	Crash cart	₹ 7,500/-			
5	Dressing trolley	₹ 13,500/-			
6	Over bed table/Food trolley	₹ 45,000/-			
7	Stainless steel operation instruments trolley(Without bucket ring)	₹ 7,650/-			
8	Kick Bucket	₹ 1,800/-			
9	Medicine trolley	₹ 19,800/-			
10	Mayo operation trolley	₹ 3,000/-			
11	Patient examination table	₹ 15,000/-			
12	Stool SS, fixed	₹ 15,000/-			
13	Stool SS, Round & revolving top, adjustable for heights	₹ 7,500/-			
14	Soiled linen trolley	₹ 2,400/-			
15	WHEEL CHAIR (SS)	₹ 4,500/-			
16	Wheel chair folding	₹ 4,050/-			