



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
**All India Institute of Medical Sciences, Raipur (Chhattisgarh)**  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

**Tender enquiry no. Store/Tender/Medical Gas Cylinder/ 1 / 2015 dated 07.05.2015**

### Corrigendum

| Clause                                     | Old | To be read as  |
|--|-----|--|
| Addendum as<br>Page no. 11<br>Point no. 18 | -   | The bidder can quote for all or any one gas.   |
| Addendum as<br>Page no. 11<br>Point no. 19 | -   | Gas supplies made to emergency service 108 is acceptable for performance certificate provided total no. of cylinders supplied every year is more than 600.   |
| Addendum as<br>Page no. 11<br>Point no. 20 | -   | For N2O and Carbon dioxide, the rate should be quoted on per Kg basis.   |
| Addendum as<br>Page no. 11<br>Point no. 21 | -   | For O2, the cylinder pressure should be > 135 bar in winter and > 140 bar in summer. Cylinders showing less pressure shall be returned and the vendor shall replace them with adequately filled cylinders at their own cost. |

Store Officer  
AIIMS, Raipur