

(3rd Call)

Notice Tender Invited for

"Multipara & Neonatal Multipara Monitor"

At

All India Institute of Medical Sciences, Raipur

Sr. No.	Description	Start Date & Time
1.	NIT No.	AIIMS-RPR/Store/Tender/Multipara & Neonatal Multipara Monitor/2016/1
2.	NIT issue date	22-03-2016
3.	Pre-bid Meeting	05-04-2016-at-03:00 PM
4.	Venue	Committee Hall, 1 st floor, Medical College Building, AIIMS, Tatibandh, Raipur-492099
5.	Last Date of submission	19-04-2016- at 03:00 PM
6.	Open EMD & Technical-bid	19-04-2016- at 03:30 PM
7.	Venue	Store Officer, Medical College Building, 2 nd floor, AIIMS, Tatibandh, Raipur-492099
8.	Tender document cost	₹ 5,700/- (Including VAT)
9.	EMD Amount	For Schedule-A : ₹ 2,16,000/- (Rupees Two Lac Sixteen Thousand Only) For Schedule-B : ₹ 88,500/- (Rupees Thousand Eight Thousand Five Hundred Only)



**All India Institute of Medical Sciences
Tatibandh, Raipur – 492099, Chhattisgarh**

Tele: 0771- 2971307, email: store@aiimsraipur.edu.in

Website: www.aiimsraipur.edu.in, www.tenders.gov.in



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर, छत्तीसगढ़
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
Website : www.aiimsraipur.edu.in
 Tele: 0771- 2971307, e-mail: store@aiimsraipur.edu.in

Sub.: Invitation of sealed tender for Supply of “**Multipara & Neonatal Multipara Monitor**” as per details and specifications shown in the Annexure-I.

Dear Sir/Madam,

ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS) RAIPUR invites sealed tender for Supply of “**Multipara & Neonatal Multipara Monitor**” as per details and specifications shown in the **Annexure-I** on the following terms & conditions:

Schedule	Name of the Item/Equipment	Quantity	EMD
A	Multipara Monitor (High End)	4 Nos.	₹2,16,000/-
B	Neonatal Multipara Monitor -High End	2 Nos.	₹88,500/-

- If the supplier/firm is manufacturer/authorized dealer/sole distributor of any item, the Certificate to this effect should be attached.
- The tender documents are to be in two parts as Technical Offer and Financial-offer:
 - The Technical offer should include the detailed specifications of main item/equipment and its accessories. All items should be numbered as indicated in the **Annexure-I**.
 - The financial offer should include the cost of main item/equipment and its accessories numbered as in Annexure-I. If there is any separate cost for installation etc. that should be quoted separately. The total cost should be quoted in words as well as figures (typed or printed).- Amendment should be avoided.-Amendments, if any, should be duly initialled, failing which the offers are liable to be rejected.
 - If bidder is participated in multiple schedules, the two parts of the offer should be placed in separate sealed envelopes for Schedule-A & Schedule-B separately clearly marked “Technical Offer” & “Financial Offer” that should be put in different envelopes duly labelled with name of schedule. These four envelopes along with envelope for EMD and Tender Fee marked “Tender Fee & EMD” same as above (total **Six envelopes**) must be enclosed in one bigger envelope duly sealed and superscribed with tender number, name of the item/equipment {**Multipara & Neonatal Multipara Monitor**} and tender due date must be forwarded to the undersigned so as to reach him on or before the due date.
 - The Financial Offer must be mentioned in the prescribed format as per the **Annexure-II (as applicable A or B)** only. If the financial offer is not in the prescribed format, it will be rejected.

Incomplete tenders, amendments and additions to tender after opening or late tenders are liable to be ignored and rejected.

3. Fax and Email quotation are not acceptable.
4. Quotations should be valid for 180 days from the tender due date. The quotation should clearly indicate the period of delivery, warranty terms etc.-A minimum of five years warranty is required from the date of commissioning.
5. Relevant literature pertaining to the items quoted with full specifications (and drawing, if any) should be sent along with the Quotations, wherever applicable.
6. Tenderer must provide evidence of having supplied to at least 5 reputed government/reputed private organizations in India including a least one government institution.
7. Tenderer must provide evidence of having supplied government hospital / reputed private hospital organizations in India similar nature of items of at least ₹24,00,000/- for schedule-A and ₹9,83,000/- for schedule-B of contract value in the last three years.
8. The firm should be registered and should have the average annual turnover of the bidder in the last three financial years of `36,00,000/- for schedule-A and `14,75,000/- for schedule-B. Copies of authenticated balance sheet for the past three financial years should be submitted.
9. The tender document must be accompanied by copy of PAN, Certificate of firm/company registration, TIN/VAT registration (Sales tax) and service tax registration.
10. The Institute is exempted to pay of "Service Tax" as per the Notification No.: 25/2012-ST, dated 20th June, 2012 issued by Govt. of India, Ministry of Finance.
11. The quotations should be given for the items in the same order as in the tender document.
12. The quantity shown against each item is approximate and may vary as per demand of the Institute at the time of placing order.
13. The Vendor must be able to provide the product/items within specified time period as prescribed in the Purchase Order. Failing the EMD will be forfeited. Furthermore on completion of the stipulated time period, Purchase Order will be cancelled and award will be given to another qualified bidder with the negotiated terms & conditions.
14. Imported supplies should be delivered by air and quoted for CIP Raipur.
15. (a)-In case of payment (for import item) through Letter of Credit (LC), 90% (Ninety Percent) of the payment will be released after completion of the supply. The balance 10% (Ten Percent) of the payment will be released after satisfactory installation of the item / equipment.
(b) If the Payment (for import item) in the form of foreign currency draft / TT, the Institute will release the payment within 30 working days after delivery and successful installation of the item/equipment at AIIMS Raipur Premises.
(c)-Advance Payment (for indigenous supplies) is not allowed in any case. 100% payment will be released after completion of the supply and successful installation.

16. AIIMS Raipur is exempted from payment of Excise Duty and is eligible for concessional rate of Custom Duty. Necessary certificate will be issued on demand. AIIMS Raipur will not make necessary arrangements for the clearance of imported goods at the Airport.
17. In the event of any dispute or difference(s) between the vendor AIIMS Raipur and the vendor(s) arising out of non-supply of material or supplies not found according to the specifications or any other cause whatsoever relating to the supply or purchase order before or after the supply has been executed, shall be referred to the concerned authority of AIIMS Raipur who may decide the matter himself or may appoint arbitrator(s) under the arbitration and conciliation Act 1996. The decision of the arbitrator shall be final and binding on both the parties.
18. The place of arbitration and the language to be used in arbitral proceedings shall be decided by the arbitrator.
19. All disputes shall be subject to Raipur Jurisdiction only.
- 20. AIIMS Raipur reserves the rights to accept/reject any offer in full or in part or accept any offer other than the lowest offer without assigning any reason thereof. Any offer containing incorrect and incomplete information shall be liable for rejection.**
21. The Tender/Bid will open on 19-04-2016 at 03:30 PM at AIIMS Raipur Premises.
 - a) The suppliers or their authorized representative may also be present during the opening of the Technical offer, if they desire so, at their own expenses.
 - b) Only those financial offers will be opened whose technical offers are found suitable by the expert committee appointed for the concerned instrument/equipment.
 - c) No separate information shall be given to individual bidders. In incomparable situation, the committee may negotiate price with the technically and financially qualified bidder before awarding the offer.
 - d) The-Tender-Committee-reserves-its-right-to-select-or-reject-any-or-all-of-the items/equipment mentioned above without assigning any reasons.
22. In case the supplier requires any elucidation regarding the tender documents, they are requested to contact to the Store Officer, AIIMS Raipur through e-mail stores@aiimsraipur.edu.in on or before 08-04-2016 at 3:00 PM.
23. A demand draft/Pay Order of ₹5,700/- (Including VAT) towards non-refundable tender fee and Earnest Money Deposit (EMD) in form of demand draft/BG/FDR/Pay Order of ₹2,16,000/- for Schedule-A and ₹88,500/- for Schedule-B towards refundable EMD from a Schedule bank in favor of "AIIMS Raipur" payable at Raipur placed in a separate envelope marked "Tender Fee & EMD" should accompany tender bid documents. Both the demand drafts should be valid for 90 days (Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India). Without the Tender Fee and EMD the bid will not be considered.

The EMD of the successful bidder will be returned to them without any interest after completing the successful installation. The earnest money

of unsuccessful bidders will be returned to them without any interest within thirty (30) working days after awarding the offer.

Tender fee & EMD is exempted for SSI/NSIC Registered firms and will be considered only if an attested copy is enclosed along with tender documents.

24. All tender documents should have to be forwarded through speed post or registered post, courier, Hand Delivery on / before 19-04-2016 at 3.00PM to Store Office, Medical College Building, 2nd floor, AIIMS, Tatibandh, Raipur-492099 (C.G.).
25. Copies of original documents defining the constitution or legal status, place of registration and principal place of business of the company or firm or partnership, etc.
26. **Award of Contract**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

 - a. Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.
 - b. The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.
27. Normal comprehensive warranty/ guarantee of 5 years shall be applicable to the supplied goods.
28. Rates quoted should be inclusive of packing, forwarding, postage and transportation charges etc.
29. The competent authority reserves all rights to reject the goods if the same are not found in accordance with the required description/specifications/quality.
30. The detailed description, specifications and quantity of the goods required is given in Annexure-I. The quantity mentioned in Annexure-I will be subject to variation.
31. Each bidder shall submit only one quotation.
32. Conditional Quotation(s) will not be accepted.

We look forward to receiving your quotations and thank you for your interest in this project.

Note: Quotation(s) without E.M.D. will be out rightly rejected.

**Stores Officer,
AIIMS, Raipur**

Other Terms & Conditions:

1. Pre-Qualification-Criteria:

- a. Bidders should be the manufacturer / authorized dealer. Letter of Authorization from Manufacturer on the same and specific to the tender should be enclosed.
- b. An undertaking from the original Manufacturer is required stating that they would facilitate the bidder on a regular basis with technology/product updates and extend support for the warranty as well.

2. Performance Guarantee Bond:

- a. Performance Guarantee Bond is mandatory.
- b. Successful supplier/ firm should submit performance guarantee as prescribed in favour of "AIIMS, Raipur" and to be received in the Store Office, 2nd Floor, Medical College Building, Tatibandh, Raipur (C.G) Pin-492 099 before the date of commencement of supply or 30 days from the date of acceptance of the purchase order, whichever is earlier. The performance guarantee bond to be furnished in the form of Bank Guarantee as per given proforma of the tender documents, for an amount covering 10% of the value of quantity of material on landed cost basis.
- c. The Performance Guarantee should be established in favour of "AIIMS Raipur" through any Schedule Bank Situated at Raipur with a clause to enforce the same on their local branch at Raipur.
- d. Validity of the performance guarantee bond shall be for a period of 60 days beyond of entire warranty period from the date of issue of installation & commissioning.

3. Delivery & Installation:-The successful bidders should strictly adhere to the following delivery schedule supply, installation & commissioning should be effected within 6 weeks from the date of supply order and this clause should be strictly adhere to failing which administrative action as deemed fit under rules will be taken against the defaulter. Otherwise LD will be imposed as per clause no. 4. Purchase order will be placed as required by consignee.

4. Penalty:-If the suppliers fails to deliver and place any or all the Equipment or perform the service by the specified date as mention in purchase order, penalty at the rate of 0.5% per week of delayed value of goods subject to the maximum of 10% of delayed goods value will be deducted, afterwards another penalty may be imposed.

5. Training and Demonstration: Suppliers need to provide adequate training and demonstration at AIIMS Raipur to the nominated person of AIIMS Raipur at their cost. AIIMS Raipur will not bear any training or living expenditure in this regard. The Supplier should arrange for regular weekly visit to the AIIMS, Raipur campus by its technical team and assist in maintenance of the item/equipment within warranty period. Assistance limited to locking companies with manufacturer will not be considered sufficient.

6. **Installation & Warranty Declaration:**-Suppliers must give the comprehensive onsite warranty as required from the date of successful installation of item/equipment against any manufacturing defects. In the installation report the model number of instrument and all spares parts / accessories numbers should be in the line of purchase order. And suppliers must be written in the warranty declaration that “everything to be supplied by us hereunder shall be free from all defects and faults in material, workmanship and shall be of the highest quality and material of the type ordered, shall be in full conformity with the specification and shall be completed enough to carry out the experiments, as specified in the tender document.” If any item covered under warranty fails, the same shall be replaced free of cost including all the applicable charges (shipping cost both ways). **Installation must be done within stipulated time period from the date of delivery of the item/equipment as specified in the purchase order.**
7. **Validity of the bids:** The bids shall be valid for a period of 180 days from the date of opening of the tender. This has to be so specified by the tenderer in the commercial bid which may be extended, if required.
8. **Right of Acceptance:** AIIMS, Raipur reserves the right to accept or reject any or all tenders/quotations without assigning any reason there of and also does not bind itself to accept the lowest quotation or any tender. AIIMS, Raipur also reserves the rights to accept all the equipment/instruments in the given tender or only part of it in any given schedule without assigning any reason.
9. **Communication of Acceptance:** AIIMS, Raipur reserves all right to reject any tender including of those tenderers who fails to comply with the instructions without assigning any reason whatsoever and does not bind itself to accept the lowest or any specific tender. The decision of this Institute in this regard will be final and binding.
10. **Guarantee / Warranty, Service, Maintenance:** The tenderers must quote for 5 years onsite warranty from the date of completion of the satisfactory installation as certified by the stipulated committee. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders should submit their quote for subsequent 5 years on site CMC (include free labour, repair other services & spare parts). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be made taking into account on basic price and post warranty CMC. The Rate Contracting Authority reserves the right to award CMC (include free labour, repair other services & spare parts) shall be quoted for equipment’s costing above ₹5.00 lacs. So the price of CMC should be quoted according to the cost of equipment. The amount of CMC would be released to the supplier on successful completion of the maintenance of that particular year duly certified by the user department.

The supplier will ensure regular maintenance service by the appropriate engineer having the technical know-how of the equipment. The supplier shall also ensure the presence of resident engineer in the geographical

location of this city of Raipur so that he attends the call without loss of time.

11. **Force Majeure:** If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, exception, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party shall be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance and deliveries have been so resumed or not shall be final and conclusive.

Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, AIIMS, Raipur party may, at least option to terminate the contract.

12. **Insolvency etc.:** In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Raipur shall have the power to terminate the contract without any prior notice.
13. **Breach of Terms and Conditions :** In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order / job without assigning any reasons thereof and nothing will be payable by AIIMS, Raipur. In that event the security deposit shall also stand forfeited.-
14. **Subletting of Work:** The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Raipur, which will be at liberty to refuse if thinks fit. The tender is not transferable. One tenderer shall submit only one tender.
15. **Right to call upon information regarding status of work:** The AIIMS, Raipur will have the right to call upon information regarding status of work/job at any point of time.
16. **Terms of payment:**
1. The payment would be made for actual supply taken and no claim in this regard should be entertained. 90% payment will be made on receiving of goods satisfactorily with approved quality & ordered quantity. And balance 10% will be paid after successful installation, commissioning and / or report from the user department.
 2. No payment shall be made for rejected Stores. Rejected equipment's must be removed by the supplier within two weeks of the date of issue of rejection

advice at their own cost & replace immediately. In case these are not removed these will be auctioned at the risk and responsibility of the suppliers without notice.

For payments supplier should submitted following documents:

1. Two copies of Invoice
2. Packing List/ Delivery Challan
3. Certificate of Origin
4. Any other documents, if required.
5. Installation Report for final 10% Payments.

17. Compulsory Enlistment of Indian Agents

As per the Compulsory Enlistment Scheme of the Department of Expenditure, Ministry of Finance, it is compulsory for Indian agents who desire to quote directly on behalf of their foreign manufacturers/principals, to get themselves enlisted with the Department of Expenditure, through the Central Purchase Organization (e.g. DGS&D).

The compulsory enlistment of Indian Agents under the scheme of Ministry of Finance is simpler and differs from the registration of Indian Agents with the Central Purchase Organization (e.g. DGS&D) described in the earlier paragraphs.

The registration of the foreign manufacturer is not a must for enlisting the Indian Agent under this scheme. No Inspection Report in respect of the foreign manufacturer/principal is necessary.

The enlistment under the scheme is not equivalent to the Registration with DGS&D. Such firms do not enjoy the same status as that of DGS&D registered suppliers. A note to this effect is given in the Enlistment Letter to the firm.

18. Octroi and Local Taxes

The goods supplied against contracts placed by Ministry / Department are generally exempted from levy of Town Duty, Octroi Duty, Terminal Tax and other Levies of local bodies. The suppliers should be informed accordingly by incorporating suitable instructions in the tender enquiry document and in the resultant contract. Wherever required, the suppliers should obtain the exemption certificate from the purchase organization to avoid payment of such levies and taxes. In case, where the municipality or the other local bodies insist upon such payments (in spite of purchase organization's exemption certificate), the supplier should make the payment to avoid delay in supplies and forward the receipt of the same to the purchase organization for reimbursement and, also, for further necessary action by the purchase organization.

19. Custom Duty on Imported Goods

In respect of imported goods, the tenderers shall also specify separately the total amount of custom duty included in the quoted price. The tenderers should also indicate correctly the rate of custom duty applicable for the

goods in question and the corresponding Indian Customs Tariff Number. Where customs duty is payable, the contract should clearly stipulate the quantum of duty payable etc. in unambiguous terms.

20. Fall Clause :

1. Prices charged for supplies under Rate Contract by the supplier should in no event exceed the lowest prices at which he offers to sell or sells the stores of identical description to any other State Government/DGS&D/Public Undertaking during the period of the contract.
2. If at any time during the period of contract, the prices of tendered items is reduced or brought down by any law or Act of the Central or State government, the supplier shall be bound to inform Purchasing Authority immediately about such reduction in the contracted prices, in case the supplier fails to notify or fails to agree for such reduction of rates, the Purchasing authority will revise the rates on lower side. If there is a price increase for any product after quoting the rates, the bidder will have to supply the item as per quoted rates. This office will not accept any higher rates after wards.
3. If at any time during the period of contract, the supplier quotes the sale price of such goods to any other State Govt./DGS&D and Public Undertakings at a price lower than the price chargeable-under the rate contract he shall forthwith notify such reduction to Purchasing Authority and the prices payable under the rate contract for the equipment's supplied from the date of coming into force of such price stands correspondingly reduced as per above stipulation.

Any deviation in the material and the specifications from the accepted terms may liable to be rejected and the suppliers need to supply all the goods in the specified form to the satisfaction / specifications specified in the Purchase order and demonstrate at the their own cost.

**Store Officer,
AIIMS Raipur**

Form A

PARTICULARS-TO-BE-FILLED BY-THE BIDDER

1. Name of the Supplier -----: _____
2. Complete Address of the Supplier : _____

3. Availability for demonstration of instruments at AIIMS Raipur. Yes/No [Please √]
4. Cost of the Tender enclosed: Yes/No [Please √] If yes,
 - a.) Name of the Bank :-- _____
 - b.)- Amount-in- (₹.) : _____
 - c.) Demand Draft No. : _____
5. Earnest Money Deposit enclosed: Yes / No [Please √] if Yes,
 - a.) Name of the Bank : _____
 - b.)- Amount in (₹.) : _____
 - c.) Demand Draft No. : _____
 - d.) Last Validity date of the enclosed DD: _____
6. Communication details of the concerned contact person to whom all references shall be made regarding this tender enquiry.

[**NOTE:** Any changes after submission of Tender documents kindly update AIIMS Raipur]

- a.) Full Name : _____
- b.) Complete Postal Address: _____
- c.) Telephone No. : _____
- d.) Fax No. : _____
- e.) Mobile No. : _____
- f.) E-mail : _____
- g.) Website Address : _____

Note: - Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India.

Form-B

**PARTICULARS FOR REFUND OF EMD TO SUCCESSFUL/UNSUCCESSFUL BIDDER
RTGS / National Electronic Fund Transfer (NEFT) Mandate Form**

1	Name of the Bidder	
2	Permanent Account No (PAN)	
3	Particulars of Bank Account	
	a)-Name of the Bank	
	b)-Name of the Branch	
	c)-Branch Code	
	d)-Address	
	e)-City Name	
	f)- Telephone No	
	g)-NEFT/IFSC Code	
	h)-RTGS Code	
	i)- 9 Digit MICR Code appearing on the	
	j)- Type of Account	
	k)-Account No.	
4	Email id of the Bidder	

Form-C

CHECKLISTS FOR INSTRUMENTS AND EQUIPMENTS ITEMS FOR OBSTETRICS & GYNAECOLOGY

Checklist – Tender Fee & EMD			
Sr. No.	Particulars	Documents to be attached wherever applicable	
		Yes	No
1.	Demand Draft for ₹5,700/- (Including Tax) towards non-refundable tender fee is enclosed. **Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India. (Tender fee is exempted for SSI/NSIC Registered Firms)		
2.	Demand Draft for ₹2,16,000/- for Schedule-A Demand Draft for ₹88,500/- for Schedule-B towards refundable EMD is enclosed. **Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India (EMD is exempted for SSI/NSIC Registered Firms)		
3.	Envelope is marked as “Tender Fee and EMD”		

Checklist – Technical Offer			
Sr. No.	Particulars	Documents to be attached wherever applicable	
		Yes	No
1.	a) Attached documents as required in the tender document (i.e. Supplier/firm is manufacturer/authorized dealer/sole distributor certificate,		
	b) Authorization certificate from the manufacturer in case of dealer / distributor.		
	c) Copy of PAN,		
	d) Certificate of firm/company registration,		
	e) TIN/VAT registration (Sales tax),		
	f) Income Tax Return of last three years,		
	g) Tenderer must provide experience/ supplied as per the clause 7.		
	h) Annual Turnover & balance sheet of last three years duly certified by C.A. as per the clause 8.		
	i) Certificate of European CE and USFDA from the Manufacturer Company.		
	j) Certificate of IEC 60601-1-1 medical safety standard		
	k) Tender document duly seal and sign by the tenderer.		
	l) Tenderer must provide a certificate on letter head that proprietor/firm has never been black listed by any organization.		
	m) Certificate of proof of manufacturing submitted by the tenderer.		
	n) Copy of Service Tax Registration.		
	o) Relevant brochure/catalogue pertaining to the items quoted with full specifications etc.		
2.	Technical Specifications Compliance Report		
3.	Duly filled Form – A & Form –B		
4.	Envelope is marked as “Technical Offer”		

Checklist – Financial Offer			
Sr. No.	Particulars	Documents to be attached wherever applicable	
		Yes	No
1.	Financial Offer of the item/equipment as per proforma available as Annexure-II (A or B as applicable)		
2.	Envelope is marked as “Financial Offer”		

Checklist – Master Envelope			
Sr. No.	Particulars	Documents to be attached wherever applicable	
		Yes	No
1.	Master envelope containing Envelopes of “Technical Offer”, “Financial Offer” and “Tender Fee and EMD” are superscribed with Tender Number, Name of item/equipment and Tender Due Date		

Note:

1. Bidders/Vendors who do not fulfil any or all of the above conditions or provide incomplete information in any respect are liable to be rejected.
2. The Director, AIIMS reserves the right to select any or all items together from the eligible & most responsive bid. The quantities mentioned are indicative only. The final decision as regards to selection of bids depends upon the actual requirement, and will be decided by the competent authority.

Signature of Vendor with Seal:-
Address :-

Name of Vendor :-
Date :-

PARTICULARS FOR PERFORMANCE GUARANTEE BOND

(To be typed on Non-judicial stamp paper of the value of Indian Rupees of Two Hundred) (TO BE-ESTABLISHED THROUGH ANY-OF-THE-SCHEDULED BANK-(WHETHER SITUATED AT RAIPUR OR OUTSTATION) WITH A CLAUSE TO ENFORCE THE SAME ON THEIR LOCAL BRANCH AT RAIPUR. BONDS ISSUED BY CO- OPERATIVE BANKS ARE NOT ACCEPTED.)

To,
The Director
All India Institute of Medical Sciences (AIIMS),
Tatibandh, GE Road,Raipur-492 099 (CG)

LETTER OF GUARANTEE

WHEREAS All India Institute of Medical Sciences (AIIMS) Raipur (Buyer) have invited Tenders vide Tender No..... Dt..... for purchase of AND WHEREAS the said tender document requires the supplier / firm (seller) whose tender is accepted for the supply of instrument / machinery, etc. in response thereto shall establish an irrevocable Performance Guarantee Bond in favour of "AIIMS Raipur"- in- the- form- of- Bank- Guarantee- for ₹..... [10% (ten percent) of the purchase value]-which will be valid for entire warranty period from the date of installation & commissioning, the said Performance Guarantee Bond is to be submitted within 30 (Thirty) days from the date of Acceptance of the Purchase Order.

NOW THIS BANK HEREBY GUARANTEES that in the event of the said supplier/firm (seller) failing to abide by any of the conditions referred to in tender document / purchase order/ performance of the instrument / machinery, etc. this Bank shall pay to All India Institute of Medical Sciences (AIIMS) Raipur on demand and without protest or demur---₹ (Rupees.....).

This Bank further agrees that the decision of All India Institute of Medical Sciences (AIIMS) Raipur (Buyer) as to whether the said supplier/firm (Seller) has committed a breach of any of the conditions referred in tender document / purchase order shall be final and binding.

We, (name of the Bank & branch) hereby further agree that the Guarantee herein contained shall not be affected by any change in the constitution of the supplier/firm (Seller) and/ or All India Institute of Medical Sciences (AIIMS) Raipur (Buyer).

Notwithstanding anything contained herein:

a. Our liability under this Bank Guarantee shall not exceed ₹..... (Indian Rupees only).

b.This Bank Guarantee shall be valid up to(date) and

c. We are liable to pay the guaranteed amount or any part thereof under this bank guarantee only and only if AIIMS Raipur serve upon us a written claim or demand on or before..... (date)

This Bank further agrees that the claims if any, against this Bank Guarantee shall be enforceable at our branch office at situated at..... (Address of local branch).

Yours truly,

Signature and seal of the Guarantor

Name of the Bank:.....

Complete Postal Address:

ANNEXURE-I

TECHNICAL SPECIFICATION FOR MULTIPARA & NEONATAL MULTIPARA MONITOR

General Specifications:

1. All the material /equipment should be European CE & US FDA.
2. All the Electronic equipment's should comply with Electrical safety conforms to standards for electrical safety IEC 60601-1.
3. All the equipment's power input should be 220- 240V AC, 50Hz fitted with Indian plug.

SCHEDULE – A: MULTIPARA MONITOR (HIGH END) (4 Nos.)

General:

- Parameters – ECG, heart rate, respiratory rate, SpO2, NIBP, Dual IBP, EtCO2, Dual Temperature.
- Should be Modular and capable of being connected to a central station.
- The equipment should come with all standard accessories required to run all parameters.
- Waveform display: at least 12 channels at a time, user selectable.
- Digital display: Heart rate, respiratory rate, oxygen saturation, temperature, Blood Pressure (systolic, diastolic, mean).
- System should be compatible with HIS.
- System should be US FDA & European CE approved for quality assurance.
- Monitor must be able to connect to central monitoring station and should use single network for all kind of networking with the central station or other hospital information system (HIS). It should have dual display, one for continuous display of all beds and other for trend review, full disclosure etc.

Individual Monitors:

- Wall-mountable and pivotable.
- Medical grade, TFT Flat screen, slim size, at least 19" display.
- Flexible display mode for various monitoring needs.
- Should be knob control and touch screen.
- Viewing angle at least 90o.
- Bigger font of numeric should also be available.
- Should have standby mode for temporary leaving of patients.
- Adjustable contrast and brightness.

Should have facility to record data for individual patients identifiable by their name and hospital registration number

The following modules complete with their accessories for each monitor:

- Heart rate
- ECG
- Respiration
- Oxygen saturation
- Temperature: 2 ports
- Non-invasive Blood pressure
- Invasive blood pressure 2 ports
- EtCO2

2. The following modules are also required in the quantity mentioned

Parameters monitored: Cost of following modules should be quoted separately.

- EEG – 2
- NMT – 1
- Minimal invasive Cardiac Output monitor – 2
- BIS - 1

3. Heart rate/ECG:

- 10-lead selectable ECG
- Built in arrhythmia monitoring in all leads
- Inbuilt ST segment analysis and arrhythmia detection facility
- Display of 2 ECG leads simultaneously at a time
- Heart rate range 20-250 bpm
- Accuracy ± 5 bpm
- Display sweep speeds 12.5, 25 mm/sec (user adjustable)
- Averaging time: user selectable up to 8 seconds
- ECG amplitude user adjustable
- Defibrillator protected

4. Respiratory rate

- Measured by transthoracic impedance using the same ECG lead
- Range 0 to 150 breaths/min
- Accuracy ± 2 bpm
- Display sweep speeds 6.25, 12.5 & 25 mm/sec (user adjustable)
- Averaging time: user selectable up to 8 seconds
- User selectable apnea alarm time

5. Oxygen Saturation

- Masimo-SET/FAST Technology to take care of low perfusion states and motion artifacts
- Dual wavelength LED pulse oximetry
- Range 1 to 100%
- SpO₂ accuracy : ± 2 % (70-100% range)
- Averaging time: user selectable up to 8 seconds
- Plethysmographic waveform display
- System must display perfusion index (PI %) from SpO₂ as an indication of pulse strength at the sensor site.

6. Temperature

- 2-channels
- Continuous digital display of two site temperatures
- Range: 25-50 degree Celsius
- Resolution: ± 0.1 degree Celsius
Accuracy: + 0.1 degree Celsius
- Should provide skin temperature probe and rectal temperature probe for every monitor

7. Noninvasive Blood pressure:

- Capable of measuring blood pressure including in neonates weighing 500 g
- Microprocessor software with unit in mmHg
- Oscillometric technique
- Manual, auto and time limited stat modes
- User selectable automatic time intervals
- Display systolic, diastolic and mean BP
- Blood pressure range

Systolic : 30-250 mm Hg

Diastolic : 10-200 mm Hg

Mean: 20 – 100 mmHg

- Pulse rate range : 20-250 bpm
- Cuff : auto deflate with over pressure protection
- Should automatically establish zero reference after each reading

8. Invasive Blood pressure

- At least 2 channels
- Compatible with reusable and disposable pressure transducers
- Transducer should allow continuous infusion into the artery through an infusion pump while pressure is simultaneously displayed
- Input through pressure transducer that functions through a fluid filled catheter system
- Unit: mmHg
- Should allow continuous fluid infusion into the artery /vein through an infusion pump without any volume limitation
- Compatible with arterial BP, CVP and intracranial pressure monitoring catheters with user selectable display names
- Range 0-350 mmHg
- Display resolution: ± 1 mmHg
- Accuracy ± 1 mmHg
- Digital waveform display
- User selectable pressure channel display
- Should be Compatible with ICP monitoring system
- Name of manufacturer of compatible ICP monitoring system should be provided with documentary evidence.
- Monitor must measure dynamic preload parameters such as pulse pressure variation (PPV) and systolic pressure variation (SPV) from arterial waveform and display on main screen.)

9. EtCO2

- Should provide mainstream modules-2nos and microstream – 2 nos.
- Warm up time: capnogram displayed in less than 15 sec
- Measurement method: infrared spectrum
- Unit: %, mmHg, kPa
- CO2 Accuracy: ± 2.0 mmHg
- Measurement range: 0-150 mmHg
- Resolution: 1 mmHg
- Resp rate accuracy ± 1 rpm
- Range of resp rate measurement: 0-150/ min
- Should also be able to monitor Etco2 by non-invasive (nasal) way from the same module.

10. EEG

- EEG module should be interchangeable with any of the monitors Input: 6 lead
- Sensitivity: 20, 50,100,250/ μ V/cm
- Display: Left or right sided waveform
- Small, cup type EEG electrodes and lead set kit suitable for neonatal use with a low impedance

11. User selectable alarms

- High and low heart rate
- High and low respiratory rate
- Apnea with adjustable time 5-20 seconds
- High and low saturation
- High and low SBP

- High and low DBP
- High and low MAP
- Probe failure
- Poor signal
- Power failure
- Audio & visual alarms with message with separate alarm indicator with 360^oviewing.

12. Trends

- Memory storage : at least 72 hours
- Data display interval : not more than 20 sec
- Display range : last ½ hour to 72 hours
- Graphical and tabular format of display of variables
- Full waveform disclosure of 24 hrs.

13. Power

- 220/240 V 50/60 Hz AC
- Rechargeable internal / external battery with a backup of at least 1 hr

14. **Manuals:** One set of operator & service manuals with each monitor

15. Comprehensive warranty period: 5 years and CMC for 5 years after warranty

16. Availability of spares for at least 10 years after date of installation

17. **Warranty and CAMC:** Comprehensive warranty and CAMC would include all parts-plastic, metallic, glass, batteries and rubber except the consumable accessories listed below against item no. 22. Comprehensive warranty and CAMC would include periodic checking and periodic calibration of all parameters strictly as per manufacturer's recommendations (at least every 3 month) and any spares, gases or standards required for that.

18. Should have online and telephonic registration of the complaints

19. Should have resident service engineer in Raipur and should be available 24 x 7 to solve the complaints.

20. Down time of the equipment will start from the time of lodgment of first complaint.

21. If machine is not made fully functional within 24 hours of the lodgment of complaint, a standby machine of similar configuration must be provided. Failure to do so will invite an increase in the period of comprehensive warranty or CAMC by 7 times of the downtime period (i.e. 7 days for every 1 day of downtime).

22. Should be able to provide training of all healthcare personnel during first 1 month of installation in the group of 5 trainees per session.

23. Essential Accessories for 1 monitor.

The following quantities are to be supplied for every monitor with the initial order:

ECG/Respiration

- ECG patient cable : 2
- Reusable ECG lead wires set : 2 sets

Oxygen saturation

- Patient extension cables: 2
- Reusable neonatal wrap around probes: 4
- Reusable pediatric clip/ finger probes: 4

Temperature

- Reusable surface (skin) temperature probes: 2
- Reusable rectal temperature probes: 2

NIBP

- Patient extension cable – 2
- Disposable NIBP cuffs suitable for neonates, children and adults (<1000g, 1000-2000g, >2000g, pediatric, adolescent and adult): 5 no of each size

IBP

- Transducer connecting cables: 2
- Reusable pressure transducer: 5
- Disposable pressure dome kit with neonatal flush device: 10
- Quotation must include a compliance statement and in addition, each of the above points is marked in the technical brochure. Points not covered in the brochure must be specifically addressed in a separate certificate.

Manuals:

- One set of operator & service manuals with each monitor
- Should be able to provide training of all healthcare personnel during first 1 month of installation in the group of 5 trainees per session.
- FDA (USA) and CE certificate must be enclosed.
- Onsite physical demonstration of the machine with all standard actual accessories of the same make which the firm intends to supply will be mandatory if demanded by the technical committee.

The Machine with all the accessories has to be left in the unit for demonstration on real patients for at least 1 month.

Prices of the above consumable accessories should be quoted separately, which will be fixed for the duration of comprehensive warranty and CAMC (i.e., for 10 years)

Quoting Price and Rates:

- Rates should be quoted in full (indicating rates of items quoted, VAT, CST, Octroi, Insurance, Packing, freight excise) and separately.
- The total cost of accessory items along with the cost of machine will be considered for comparative purpose among companies.
- If the firm does not list any item in the tender, no payment will be made for the items required for complete operation of the machine.
- The firm will be obliged to supply any omitted items for the entire duration of warranty free of cost and no representation on this will be entertained. If the company fails to quote for any item, it has to provide the item free of cost for three years.

24. General

- Should be able to remotely access the patient monitoring system via internet
- System should be compatible with HIS and be HL-7 compliant
- Ready to run web based applications like HIS, PACS, RIS, LIS etc on
- Patient monitor screen itself without need of additional server/PC/gateway hardware and software as standard supply.
- License should be provided for at least 2 users.
- Should provide with sturdy good quality stand to mount on wall.

SCHEDULE – B : NEONATAL MULTIPARA MONITOR -HIGH END (2 Nos.)

- Basic parameters (ECG/heart rate, O2 saturation, NIBP, respiration & temperature) – in all monitors
- Invasive BP module: in all monitors
- Microstream end tidal CO2 module: in all monitors

1. General:

- Upgradable Modular system, capable of being connected to a central station
- The equipment should come with all standard accessories required to run all parameters.
- **Waveform display:** at least 12 channels, user selectable Digital display: Heart rate, respiratory rate, oxygen saturation, temperature, Blood Pressure (systolic, diastolic, mean), EtCO2(whenever applicable).
- Should be upgradable for measurement of cardiac output/SvO2
- Should be possible to move the Microstream end tidal CO2 and EEG modules from one monitor to another.
- Should be able to remotely access the patient monitoring system via internet
- System should be compatible with HIS and be HL-7 compliant
- Ready to run web based applications like HIS, PACS, RIS, LIS etc on patient monitor screen itself without need of additional server/PC/gateway hardware and software as standard supply.

2. Individual Monitors:

- Wall mountable and pivotable
- Medical grade, TFT Flat screen, slim size, at least 19” display
- Screen resolution at least 1280x1024 pixels
- Clear bright color display with large character size
- Should have knob control and touch screen
- Viewing angle at least 90o
- Adjustable contrast and brightness
- Ability to zoom any parameters
- Ability to adjust individual alarms
- Ability to change color of the trace by user
- RS 232C interface for data communication
- Internal at least 1 hour backup time
- Store and review trends for at least 72 hours
- License should be provided for at least 2 users.

3. Parameters monitored:

- The following modules complete with their accessories:
- Heart rate/ECG
- Respiration
- Oxygen saturation
- Temperature
- Non-invasive Blood pressure
- Invasive blood pressure
- EEG (1 monitor), ‘micro stream’ EtCO2 (2 monitors)

4. Heart rate/ECG:

- Derived 12 lead ECG
- Built in arrhythmia monitoring in all leads
- Inbuilt ST segment analysis and arrhythmia detection facility
- Display of 2 ECG leads simultaneously at a time
- Heart rate range 20-240 bpm

- Accuracy + 5 bpm
- Display sweep speeds 12.5, 25 mm/sec (user adjustable)
- Averaging time: user selectable up to 8 seconds
- ECG amplitude user adjustable
- Defibrillator protected

5. Respiratory rate

- Measured by transthoracic impedance using the same ECG lead
- Range 0 to 150 breaths/min
- Accuracy + 2 bpm
- Display sweep speeds 6.25, 12.5 & 25 mm/sec (user adjustable)
- Averaging time: user selectable up to 8 seconds
- User selectable apnea alarm time

6. Oxygen Saturation

- Masimo-SET/FAST technology to take care of low perfusion states and motion artefacts.
- Dual wavelength LED pulse oximetry
- Range 1 to 100%
- SpO2 accuracy : + 2 % (70-100% range)
- Averaging time: user selectable up to 8 seconds
- Plethysmographic waveform display

7. Temperature

- Continuous digital display of two site temperatures
- Range: 25-50 degree Celsius
- Resolution: ± 0.1 degree Celsius
- Accuracy: + 0.1 degree Celsius

8. Non-invasive Blood pressure:

- Capable of measuring blood pressure in neonates weighing 400g to 5000g
- Microprocessor software with unit in mmHg
- Oscillometric technique
- Manual, auto and time limited stat modes
- User selectable automatic time intervals
- Display systolic, diastolic and mean BP
- Blood pressure range
- Systolic BP : 30-150 mm Hg
- Diastolic BP : 10-100 mm Hg
- Mean: BP 20 – 100 mmHg
- Pulse rate range : 20-240 bpm
- Cuff : auto deflate with over pressure protection
- Should automatically establish zero reference after each reading

9. Invasive Blood pressure

- At least 2 channels
- Compatible with reusable and disposable pressure transducers
- Transducer should allow continuous infusion into the artery through an infusion pump while pressure is simultaneously displayed Input through pressure transducer that functions through a fluid filled catheter system
- Unit: mmHg
- Should allow continuous fluid infusion into the artery /vein through an infusion pump without any volume limitation
- Compatible with arterial BP, CVP, pulmonary artery pressure and intracranial
- pressure monitoring catheters

- Range 0-300 mmHg
- Display resolution: ± 1 mmHg
- Accuracy ± 1 mmHg
- Digital waveform display
- User selectable pressure channel display

10. EEG

- EEG module should be interchangeable with any of the monitors Input: 6 lead
- Sensitivity: 20, 50,100,250/ μ V/cm
- Display: Left or right sided waveform
- Small, cup type EEG electrodes and lead set kit suitable for neonatal use with a low impedance

11. Microstream end tidal capnography (EtCO2)

- Module should be interchangeable with any of the monitors
- Microstream technology with neonatal mode
- Display of both waveform and numerical values
- Should be usable in intubated as well as non-intubated neonates
- Measured parameters: EtCO₂, CO₂ waveform, Respiratory rate
- CO₂ range – 0-150 mmHg
- Respiratory rate – 0-150 breaths / min
- Display – in mmHg and %
- Accuracy - ± 2 mmHg (0-38 mmHg)

12. User selectable alarms

- High and low heart rate
- High and low respiratory rate
- Apnea with adjustable time 5-20 seconds
- High and low saturation
- High and low SBP
- High and low DBP
- High and low MAP
- High and low EtCO₂
- Probe failure
- Poor signal
- Power failure
- Audio & visual alarms with message

13. Trends

- Memory storage : at least 72 hours
- Data display interval : not more than 20 sec
- Display range : last ½ hour to 72 hours
- Graphical and tabular format of display of variables

14. Power:

- 220/240 V
- 50/60 Hz AC
- Rechargeable internal/external battery with a backup of at least 1hr

15. Manuals: Operator & service manuals

16. Should be approved & certified by FDA (USA) and European CE (Certificate to be submitted)

17. Manufacturer should have ISO certification for quality standards.

18. Shall meet the safety requirements as per IEC 60601-2-27:1994—Medical electrical equipment—Part 2: Particular requirements for the safety of electrocardiographic monitoring equipment.

19. Availability of spares for at least 10 years after date of installation

20. Warranty and CMC would include the periodic calibration of all parameters strictly as per manufacturer's recommendations and any spares or standards required for that.

21. Onsite physical demonstration of the monitor with all the requested modules will be mandatory.
22. Should have local service facility and should have the necessary equipment's to carry out preventive maintenance test.
23. **Essential Accessories**
 - The following quantities are to be supplied with the initial order:
 - ECG/Respiration per monitor.
 - ECG patient cable : 2
 - ECG lead wires set (at least 3 leads) : 2 sets
 - Disposable small ECG electrodes for preterm neonates: 50 sets (set of 3)
 - Oxygen saturation
 - Patient extension cables: 2
 - Reusable neonatal wrap around probes: 2
 - Reusable ear probe – 2
 - Temperature
 - Reusable surface (skin) temperature probes: 2
 - NIBP
 - Patient extension cable – 2
 - Disposable NIBP cuffs of sizes suitable for neonates (<1000g, 1000-2000g, and >2000g): 10 no of each size (total 30 no).
- IBP**
 - Transducer connecting cables: 2
 - Reusable pressure transducer: 4
 - Disposable pressure dome kit with neonatal flush device: 50
 - Microstreamcapnography (EtCO₂)
 - Disposable EtCO₂ sensors: Through ET tube: 50
- EEG**
 - Patient cable: 2
 - Small, cup type EEG electrodes suitable for neonates with lead set kit: 2
24. **Warranty and CAMC:** Comprehensive warranty and CAMC would include all parts-plastic, metallic, glass, batteries and rubber except the consumable accessories listed below against item no.
25. Comprehensive warranty and CAMC would include periodic checking and periodic calibration of all parameters strictly as per manufacturer's recommendations (at least every 3 month) and any spares, gases or standards required for that.
26. Should have online and telephonic registration of the complaints
27. Should have resident service engineer in Raipur and should be available 24 x 7 to solve the complaints.
28. Down time of the equipment will start from the time of lodgment of first complaint.
29. If machine is not made fully functional within 24 hours of the lodgment of complaint, a standby machine of similar configuration must be provided. Failure to do so will invite an increase in the period of comprehensive warranty or CAMC by 7 times of the downtime period (i.e. 7 days for every 1 day of downtime).
30. Should be able to provide training of all healthcare personnel during first 1 month of installation in the group of 5 trainees per session.
31. Prices of above consumables should be quoted separately and frozen for the period including warranty and CMC period.
32. Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility or should comply with 89/366/EEC; EMC-directive.
33. The unit shall be capable of being stored continuously in ambient temperature of 0-50deg C and relative humidity of 15-90%.
34. The unit shall be capable of operating continuously in ambient temperature of 10 -40 deg C and relative humidity of 15-90%.

35. Should have local service facility .The service provider should have the necessary equipment's recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual.
36. User/Technical/Maintenance manuals to be supplied in English.
37. Certificate of calibration and inspection.
38. List of important spare parts and accessories with their part number and costing.
39. Log book with instruction for daily, weekly, monthly and quarterly maintenance checklist.
40. The job description of the hospital technician and company service engineer should be clearly spelt out.
41. Should provide with sturdy good quality stand to mount on wall.

Note:

- **5 YEARS WARRANTY WITH QUOTE FOR NEXT 5 YEARS CMC IS REQUIRED INCLUDING ALL ACCESSORIES.**
- **PLEASE QUOTE SEPARATE PRICES FOR ACCESSORIES WHICH SHOULD BE FREEZED FOR NEXT 5 YEARS.**

(Signature & Stamp of the bidder)

Note- Please sign each page of document including terms & conditions & tender

ANNEXURE-II(A): FINANCIAL OFFER FOR INDIGINEOUS SUPPLIES

Ref. No. & Date : -

Tender No. : -

Due Date : -

PRICE – BID**Schedule-A (Multipara Monitor)**

S. No.	Description of Item & Specification (Model No. if any)	Qty. In Units	Make & Model No.	Unit Price in ₹	Discount (%)	Excise Duty / Custom Duty	CST /VAT (%)	Total Price in ₹
1	Multipara Monitor	4						
2	Web Browsing/ Gateway/ Telemetry/ Distant Access	1						
3	Minimal Invasive Cardiac Output Module	2						
4	NMT Module	1						
5	BIS Module	1						
6	EEG Module	2						
7	Central Station	1						
Total								

Schedule-B (Neonatal Multipara Monitor)

S. No.	Description of Item & Specification (Model No. if any)	Qty. In Units	Make & Model No.	Unit Price in ₹	Discount (%)	Excise Duty / Custom Duty	CST /VAT (%)	Total Price in ₹
1	Neonatal Multipara Monitor	2						
2	Web Browsing/ Gateway/Telemetry /Distant Access	1						
3	EEG Module	1						
Total								

- **Delivery Mode** : Delivery at AIIMS Raipur, at site only
- Total-bid-price-should-be-inclusive-of-all-taxes-and-levies,-transport,-loading, unloading, cartridge charges, etc.
- **Warranty Period** : 5 Years
- **Delivery Period** :days.
- **Quotation Validity Date:** - Minimum 180 Days from the date-of-Submission of quotation/tender.
- **Payment Term** : Payment within 30 working days from the date of submission clear of bill with acceptance certificate from the concerned dept./Sect./MMD.
- **The comparison of prices will be made schedule wise.**

Sign of bidder :-

Date :-

Name of the bidder :-

Firm's Name :-

Sign & Seal of Bidder

ANNEXURE-II (B) : FOR-IMPORTED-SUPPLIES

Ref. No. & Date : -
 Tender No. : -
 Due Date : -

PRICE – BID**Schedule-A (Multipara Monitor)**

Sr. No.	Description of Item & Specification	Qty. in Units	Make & Model No.	Unit Price	Price Basis at AIIMS, Raipur	Total Price
					CIP	
1	Multipara Monitor	4				
2	Web Browsing/ Gateway/ Telemetry/ Distant Access	1				
3	Minimal Invasive Cardiac Output Module	2				
4	NMT Module	1				
5	BIS Module	1				
6	EEG Module	2				
7	Central Station	1				
Total						

Schedule-B (Neonatal Multipara Monitor)

Sr. No.	Description of Item & Specification	Qty. In Units	Make & Model No.	Unit Price	Price Basis at AIIMS, Raipur	Total Price
					CIP	
1	Neonatal Multipara Monitor	2				
2	Web Browsing/ Gateway/ Telemetry/ Distant Access	1				
3	EEG Module	1				
Total						

1. **Delivery Period** :
2. **Warranty Period** : 5 Years.
3. **Terms of payment:**
 - (a) 100% payment by Letter of Credit (90% payment will be released on receipt of documents without any discrepancies and balance 10% will be paid after satisfactory installation and commissioning).
 - (b) Payment by wire transfer (on request) within 30 days from the date of supply and installation of item/equipment.
4. Validity of the bid 180 days from the date of opening of the tender.
5. **The comparison of prices will be made schedule wise.**

Signature :
 Name :
 Business Address :
 Place :
 Date :
 Affix Rubber Stamp :

Note: Price Bid should be submitted in given format only. For additional information/extra items above format may be typed and used.

PRICE SCHEDULE FOR
COMPREHENSIVE MAINTENANCE CONTRACT (C.M.C) AFTER
EXPIRY OF WARRANTY

(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)

Sr No	SME Code No.	Name of the Equipment	For Sixth year with spare parts & labour	For Seventh year with spare parts & labour	For Eighth year with spare parts & labour	For Ninth year with spare parts & labour	For Tenth year with spare parts & labour
(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)

Place:

Date:

Signature

Name (Capital Letters):.....

Designation :.....

Declaration by the Bidder:

1. This-is-to-certify-that-I/We-before-signing-this-tender-have-read-and-fully-understood all the terms and conditions contained in Tender document regarding terms & conditions of the contract, rules regarding High Frequency Ventilator & Oscillator. I/we agree to abide them.
2. No other charges would be payable by Client and there would be no increase in rates during the Contract period.

Place:.....

(Signature of Bidder with seal)

Date:.....

Name :

Seal :

Address :