

(2nd Call)
Notice Tender Invited
for
"Supply of Surgical Loups"

At

All India Institute of Medical Sciences, Raipur

Sr. No.	Description	Start Date & Time
1.	NIT No.	Store/Tender/Surgical_Loups/2 / 2017
2.	NIT issue date	11-01-2017
3.	Pre-bid Meeting	25-01-2017 at 03:00 PM
4.	Venue	Venue: Committee Hall, 1 st floor, Medical College Building, AIIMS, Tatibandh, Raipur-492099
5.	Last Date of submission	10-02-2017 at 03:00 PM
6.	Open EMD & Technical bid	10-02-2017 at 03:30 PM Venue: Store Officer, Medical College Building, 2 nd floor, AIIMS, Tatibandh, Raipur-492099
7.	Venue	Store Officer, 2 nd floor, Medical College Building, AIIMS, Tatibandh, Raipur-492099
8.	Tender document cost	₹ 5,725/- (Inclusive VAT)



आरोग्यम् सुखं सम्पदा

All India Institute of Medical Sciences
Tatibandh, Raipur – 492099, Chhattisgarh
Tele: 0771- 2971307, email: store@aiimsraipur.edu.in
Website: www.aiimsraipur.edu.in, www.tenders.gov.in



अखिलभारतीयआयुर्विज्ञानसंस्थान, रायपुर, छत्तीसगढ़
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
Website : www.aiimsraipur.edu.in

Tele: 0771- 2971307, e-mail: store@aiimsraipur.edu.in

Tender No.: Store/Tender/Surgical_Loups/2/2017,

Dt: 11.01.2017

Last date : on/before 10.02.2017 at 03:00 PM

Sub.: Invitation of sealed tender for “**Supply of Surgical Loups**” in various department at AIIMS, Raipur as per details and specifications shown in the Annexure-I.

Dear Sir/Madam,

ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS) RAIPUR invites sealed tender for “**Supply of Surgical Loups**” in various department at AIIMS, Raipur as per details and specifications shown in the **Annexure-I** on the following terms & conditions:

Name of Department	Schedule	Description of item	Qty.	EMD in ₹
Burn & Plastic Surgery	1	Surgical Loup: high end with illumination accessories	1	13,500.00
	2	Surgical Loupes: without illumination accessories	1	7,500.00
Otorhinolaryngology	3	Surgical Loup: high end with illumination accessories	1	18,000.00
	4	Surgical Loupes: without illumination accessories	1	9,000.00
Pediatric Surgery	5	Magnifying Loup	1	4,500.00
	6	Magnifying Loup with LED Light	1	10,500.00
Trauma & Emergency	7	Surgical Loup: high end with illumination accessories	1	22,500.00
	8	Surgical Loupes: without illumination accessories	1	7,500.00

1. If the supplier / firm is manufacturer / authorized dealer / sole distributor of any item, the Certificate to this effect should be attached.
2. The tender documents are to be in two parts as Technical Offer and Financial offer:
 - a) The Technical offer should include the detailed specifications of main item/equipment and its accessories. All items should be numbered as indicated in the **Annexure-I**.
 - b) The financial offer should include the cost of main item/equipment and its accessories numbered as in Annexure-II (A or B). If there is any separate cost for installation etc. that should be quoted separately. The total cost should be quoted in words as well as figures (typed or printed). Amendment should be

avoided. Amendments, if any, should be duly initialled, failing which the offers are liable to be rejected.

- c) The two parts of the offer should be placed in separate sealed envelopes clearly marked "Technical Offer" & "Financial Offer". These two envelopes along with envelope for EMD and Tender Fee marked "Tender Fee & EMD" (total three envelopes) must be enclosed in one bigger envelope duly sealed and superscribed with tender number, name of the item/equipment **{Supply of Surgical Loups at AIIMS, Raipur}**; and tender due date must be forwarded to the undersigned so as to reach him on or before the due date.
- d) The Financial Offer must be mentioned in the prescribed format as per the **Annexure-II (as applicable A or B)** only. If the financial offer is not in the prescribed format, it will be rejected.

Incomplete tenders, amendments and additions to tender after opening or late tenders are liable to be ignored and rejected.

3. Fax and Email quotation are not acceptable.
4. Quotations should be valid for 180 days from the tender due date. The quotation should clearly indicate the period of delivery, warranty terms etc. as per tender document.
5. Relevant literature pertaining to the items quoted with full specifications (and drawing, if any) should be sent along with the Quotations, wherever applicable.
6. Tenderer must provide evidence of having supplied government hospital / reputed private hospital organizations in India similar nature of items of ₹ 1,50,000/- for schedule-1, ₹ 83,500/- for schedule-2, ₹ 2,00,000/- for schedule-3, ₹ 1,00,000/- for schedule-4, ₹ 50,000/- for schedule-5, ₹ 1,17,000/- for schedule-6, ₹ 2,50,000/- for schedule-7, ₹ 83,500/- for schedule-8 of contract value in the last three years.
7. The firm should be registered and the average annual turnover of the bidder in the last three financial years should not be less than 50% of contract value. Copies of authenticated balance sheet for the last three financial years should be submitted.
8. The tender document must be accompanied by copy of PAN, Certificate of firm/company registration, TIN / VAT registration (Sales tax).
9. The quotations should be given for the items in the same order as in the tender document.
10. The quantity shown against each item is approximate and may vary as per demand of the Institute at the time of placing order.
11. The Vendor must be able to provide the product / items within specified time period as prescribed in the Purchase Order. Failing the EMD will be forfeited. Furthermore on completion of the stipulated time period, Purchase Order will be cancelled and award will be given to another qualified bidder with the negotiated terms & conditions.
12. Imported supplies should be delivered by air and quoted for CIP Raipur.
- (a) In case of payment (for import item) through Letter of Credit (LC) in case of foreign tenderer, 90% (Ninety Percent) of the payment will be released after completion of the supply. The balance 10% (Ten Percent) of the payment will be released after satisfactory installation of the item / equipment.

(b) If the Payment (for import item) in the form of foreign currency draft / TT, the Institute will release the payment within 30 working days after delivery and successful installation of the item / equipment at AIIMS Raipur Premises.

(c) Advance Payment (for indigenous supplies) is not allowed in any case. 100% payment will be released after completion of the supply and successful installation.

For 90% Payment following documents to be submitted:

- a. Two copies of Invoice
- b. Packing list / Delivery Challan
- c. Certificate of Origin
- d. Any other documents, if applicable.

For 10% Payment following documents to be submitted:

- a. Warranty certificate
- b. Installation Report
- c. Any other documents, if applicable.

13. AIIMS Raipur is exempted from payment of Excise Duty and is eligible for concessional rate of Custom Duty. Necessary certificate will be issued on demand. AIIMS Raipur will not make necessary arrangements for the clearance of imported goods at the Airport.
14. In the event of any dispute or difference(s) between the vendee AIIMS Raipur and the vendor(s) arising out of non-supply of material or supplies not found according to the specifications or any other cause whatsoever relating to the supply or purchase order before or after the supply has been executed, shall be referred to the concerned authority of AIIMS Raipur who may decide the matter himself or may appoint arbitrator(s) under the arbitration and conciliation Act 1996. The decision of the arbitrator shall be final and binding on both the parties.
15. The place of arbitration and the language to be used in arbitral proceedings shall be decided by the arbitrator.
16. All disputes shall be subject to Raipur Jurisdiction only.
- 17. AIIMS Raipur reserves the rights to accept/reject any offer in full or in part or accept any offer other than the lowest offer without assigning any reason thereof. Any offer containing incorrect and incomplete information shall be liable for rejection.**
18. The Tender / Bid will open on 10/02/2017 at 03:30 PM at AIIMS Raipur Premises.
 - a) The suppliers or their authorized representative may also be present during the opening of the Technical offer, if they desire so, at their own expenses.
 - b) Only those financial offers will be opened whose technical offers are found suitable by the expert committee appointed for the concerned instrument / equipment.
 - c) No separate information shall be given to individual bidders. In incomparable situation, the committee may negotiate price with the technically and financially qualified bidder before awarding the offer.
 - d) The Tender Committee reserves its right to select or reject any or all of the items / equipment mentioned above without assigning any reasons.

19. In case the supplier requires any elucidation regarding the tender documents, they are requested to contact to the Store Officer, AIIMS Raipur through e-mail stores@aiimsraipur.edu.in on or before 25.01.2017 at 3:00 PM.
20. A demand draft/Pay Order of ₹ 5,725/- (Cost ₹ 5000 + VAT@14.5% ₹ 725= ₹ 5,725/-) towards non-refundable tender fee and Earnest Money Deposit (EMD) in form of demand draft/BG/FDR/Pay Order towards refundable EMD from a Schedule bank in favor of “AIIMS Raipur” payable at Raipur placed in a separate envelope marked “Tender Fee & EMD” should accompany tender bid documents. Both the demand drafts should be valid for 90 days (Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India). Without the Tender Fee and EMD the bid will not be considered. The EMD of the successful bidder will be returned to them without any interest after completing the successful installation. The earnest money of unsuccessful bidders will be returned to them without any interest within thirty (30) working days after awarding the offer.
21. All tender documents should have to be forwarded through speed post or registered post, courier, Hand Delivery on / before 10/02/2017 at 3.00PM to Store Office, Medical College Building, 2nd floor, AIIMS, Tatibandh, Raipur-492099.

**Stores Officer,
AIIMS, Raipur**

Other Terms & Conditions:

1. Pre-Qualification Criteria:

- a. Bidders should be the manufacturer / authorized dealer. Letter of Authorization from Manufacturer on the same and specific to the tender should be enclosed.
- b. An undertaking from the original Manufacturer is required stating that they would facilitate the bidder on a regular basis with technology/product updates and extend support for the warranty as well.

2. Performance Guarantee Bond:

- a. Performance Guarantee Bond draft is given on page no. 13.
- b. The successful tenderer will be required to furnish a Performance Security Deposit of 10% of tender amount in the form of Demand Draft, Fixed Deposit Receipt or Bank Guarantee from any Scheduled Bank duly pledged in the name of the "All India Institute of Medical Sciences, Raipur". The security deposit can be forfeited by order of this Institute in the event of any breach or negligence or non-observance of any condition of contract or for unsatisfactory performance or non-observance of any condition of the contract.
- c. Performance Security shall be submitted in the form of Demand Draft, Bank Guarantee or Fixed Deposit Receipt issued by any Scheduled Bank. Performance Security will valid till 60 days after completion of contractual obligations (including warranty period, if applicable) under the contract.

3. Delivery & Installation: The successful bidders should strictly adhere to the following delivery schedule supply, installation & commissioning should be effected within 6 weeks from the date of supply order and this clause should be strictly adhere to failing which administrative action as deemed fit under rules will be taken against the defaulter. Otherwise LD will be imposed as per clause no. 4. Purchase order will be placed as required by consignee.

4. Penalty: If the suppliers fails to deliver and place any or all the Equipment or perform the service by the specified date as mention in purchase order, penalty at the rate of 0.5% per week of the delayed value of goods subject to the maximum of 10% of delayed goods value will be deducted.

5. Training and Demonstration: Suppliers need to provide adequate training and demonstration at AIIMS Raipur to the nominated person of AIIMS Raipur at their cost. AIIMS Raipur will not bear any training or living expenditure in this regard. The Supplier should arrange for regular weekly visit to the AIIMS, Raipur campus by its technical team and assist in maintenance of the item/equipment within warranty period. Assistance limited to locking companies with manufacturer will not be considered sufficient.

6. **Installation & Warranty Declaration:** Suppliers must give the comprehensive onsite warranty as per annexure-I is required from the date of successful installation of item/equipment against any manufacturing defects. In the installation report the model number of instrument and all spares parts / accessories numbers should be in the line of purchase order. And suppliers must be written in the warranty declaration that “everything to be supplied by us hereunder shall be free from all defects and faults in material, workmanship and shall be of the highest quality and material of the type ordered, shall be in full conformity with the specification and shall be completed enough to carry out the experiments, as specified in the tender document.” If any item covered under warranty fails, the same shall be replaced free of cost including all the applicable charges (shipping cost both ways).
7. **Validity of the bids:** The bids shall be valid for a period of 180 days from the date of opening of the tender. This has to be so specified by the tenderer in the commercial bid which may be extended, if required.
8. **Right of Acceptance:** AIIMS, Raipur reserves the right to accept or reject any or all tenders/quotations without assigning any reason there of and also does not bind itself to accept the lowest quotation or any tender. AIIMS, Raipur also reserves the rights to accept all the equipment/instruments in the given tender or only part of it in any given schedule without assigning any reason.
9. **Communication of Acceptance:** AIIMS, Raipur reserves all right to reject any tender including of those tenderers who fails to comply with the instructions without assigning any reason whatsoever and does not bind itself to accept the lowest or any specific tender. The decision of this Institute in this regard will be final and binding.
10. **Guarantee / Warranty, Service, Maintenance:** The tenderers must quote onsite warranty as per **Annexure-I** from the date of completion of the satisfactory installation as certified by the stipulated committee. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected.

The supplier will ensure regular maintenance service by the appropriate engineer having the technical know-how of the equipment. The supplier shall also ensure the presence of resident engineer in the geographical location of this city of Raipur so that he attends the call without loss of time.
11. **Force Majeure:** If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, exception, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party shall by reason of such event be entitled to terminate this contract nor shall either

party have any claim for damages against other in respect of such non-performance or delay in performance and deliveries have been so resumed or not shall be final and conclusive.

Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, AIIMS, Raipur party may, at least option to terminate the contract.

12. **Insolvency etc.:** In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Raipur shall have the power to terminate the contract without any prior notice.
13. **Breach of Terms and Conditions :** In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order / job without assigning any reasons thereof and nothing will be payable by AIIMS, Raipur. In that event the security deposit shall also stand forfeited.
14. **Subletting of Work:** The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Raipur, which will be at liberty to refuse if thinks fit. The tender is not transferable. One tenderer shall submit only one tender.
15. **Right to call upon information regarding status of work:** The AIIMS, Raipur will have the right to call upon information regarding status of work/job at any point of time.
16. **Terms of payment:**
 1. The payment would be made for actual supply taken and no claim in this regard should be entertained. 70% payment will be made on receiving of goods satisfactorily with approved quality & ordered quantity. And balance 30% will be paid after successful installation, commissioning and / or report from the user department.
 2. No payment shall be made for rejected Stores. Rejected equipment's must be removed by the supplier within two weeks of the date of issue of rejection advice at their own cost & replace immediately. In case these are not removed these will be auctioned at the risk and responsibility of the suppliers without notice.

For 70% Payment following documents to be submitted:

 - a. Three copies of Invoice
 - b. Packing list / Delivery Challan
 - c. Certificate of Origin
 - d. Any other documents, if applicable.

For 30% Payment following documents to be submitted:

- a. Warranty certificate
- b. Installation Report
- c. Any other documents, if applicable.

17. Compulsory Enlistment of Indian Agents

As per the Compulsory Enlistment Scheme of the Department of Expenditure, Ministry of Finance, it is compulsory for Indian agents who desire to quote directly on behalf of their foreign manufacturers/principals, to get themselves enlisted with the Department of Expenditure, through the Central Purchase Organization (e.g. DGS&D).

The compulsory enlistment of Indian Agents under the scheme of Ministry of Finance is simpler and differs from the registration of Indian Agents with the Central Purchase Organization (e.g. DGS&D) described in the earlier paragraphs.

The registration of the foreign manufacturer is not a must for enlisting the Indian Agent under this scheme. No Inspection Report in respect of the foreign manufacturer/principal is necessary.

The enlistment under the scheme is not equivalent to the Registration with DGS&D. Such firms do not enjoy the same status as that of DGS&D registered suppliers. A note to this effect is given in the Enlistment Letter to the firm.

18. Octroi and Local Taxes

The goods supplied against contracts placed by Ministry / Department are generally exempted from levy of Town Duty, Octroi Duty, Terminal Tax and other Levies of local bodies. The suppliers should be informed accordingly by incorporating suitable instructions in the tender enquiry document and in the resultant contract. Wherever required, the suppliers should obtain the exemption certificate from the purchase organization to avoid payment of such levies and taxes. In case, where the municipality or the other local bodies insist upon such payments (in spite of purchase organization's exemption certificate), the supplier should make the payment to avoid delay in supplies and forward the receipt of the same to the purchase organization for reimbursement and, also, for further necessary action by the purchase organization.

19. Custom Duty on Imported Goods

In respect of imported goods, the tenderers shall also specify separately the total amount of custom duty included in the quoted price. The tenderers should also indicate correctly the rate of custom duty applicable for the goods in question and the corresponding Indian Customs Tariff Number. Where customs duty is payable, the contract should clearly stipulate the quantum of duty payable etc. in unambiguous terms.

20. Customs Duty Reimbursement:

The supplier will pay the Customs duty wherever applicable, which will be reimbursed by purchaser as per documentary evidence. The Customs duty exemption certificate will be provided to the supplier as and when required.

21. Fall Clause :

1. Prices charged for supplies under Contract by the supplier should in no event exceed the lowest prices at which he offers to sell or sells the stores of identical description to any other State Government /DGS&D/Public Undertaking during the period of the contract.
2. If at any time during the period of contract, the prices of tendered items is reduced or brought down by any law or Act of the Central or State government, the supplier shall be bound to inform Purchasing Authority immediately about such reduction in the contracted prices, in case the supplier fails to notify or fails to agree for such reduction of rates, the Purchasing authority will revise the rates on lower side. If there is a price increase for any product after quoting the rates, the bidder will have to supply the item as per quoted rates. This office will not accept any higher rates after wards.
3. If at any time during the period of contract, the supplier quotes the sale price of such goods to any other State Govt./DGS&D and Public Undertakings at a price lower than the price chargeable under the rate contract he shall forthwith notify such reduction to Purchasing Authority and the prices payable under the rate contract for the equipment's supplied from the date of coming into force of such price stands correspondingly reduced as per above stipulation.

Any deviation in the material and the specifications from the accepted terms may liable to be rejected and the suppliers need to supply all the goods in the specified form to the satisfaction / specifications specified in the Purchase order and demonstrate at the their own cost.

**Store Officer,
AIIMS Raipur**

FormA

PARTICULARS TO BE FILLED BY THE BIDDER

1. Name of the Supplier : _____
2. Complete Address of the Supplier : _____

3. Availability for demonstration of instruments at AIIMS Raipur. Yes/No [Please √]
4. Cost of the Tender enclosed: Yes/No [Please √] If yes,
 - a.) Name of the Bank : _____
 - b.) Amount in (Rs.) : _____
 - c.) Demand Draft No. : _____
5. Earnest Money Deposit enclosed: Yes / No [Please √] if Yes,
 - a.) Name of the Bank : _____
 - b.) Amount in (Rs.) : _____
 - c.) Demand Draft No. : _____
 - d.) Last Validity date of the enclosed DD: _____
6. Communication details of the concerned contact person to whom all references shall be made regarding this tender enquiry.

[**NOTE:** Any changes after submission of Tender documents kindly update AIIMS Raipur]

- a.) Full Name : _____
- b.) Complete Postal Address: _____
- c.) Telephone No. : _____
- d.) Fax No. : _____
- e.) Mobile No. : _____
- f.) E-mail : _____
- g.) Website Address : _____

Note: - Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India.

Form B

PARTICULARS FOR REFUND OF EMD TO SUCCESSFUL/UNSUCCESSFUL BIDDER
RTGS / National Electronic Fund Transfer (NEFT) Mandate Form

1	Name of the Bidder	
2	Permanent Account No (PAN)	
3	Particulars of Bank Account	
	a) Name of the Bank	
	b) Name of the Branch	
	c) Branch Code	
	d) Address	
	e) City Name	
	f) Telephone No	
	g) NEFT/IFSC Code	
	h) RTGS Code	
	i) 9 Digit MICR Code appearing on the	
	j) Type of Account	
	k) Account No.	
4	Email id of the Bidder	

Form C**CHECKLISTS FOR SUPPLY OF SURGICAL LOUPS FOR VARIOUS DEPARTMENT
AT AIIMS RAIPUR**

Checklist – Tender Fee & EMD			
Sr. No.	Particulars	Documents to be attached wherever applicable	
		Yes	No
1.	Demand Draft for ₹ 5,725/- (Cost ₹ 5000 + VAT@14.5% ₹ 725= ₹ 5,725/-) towards non-refundable tender fee is enclosed. **Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India		
2.	Demand Draft schedule wise towards refundable EMD is enclosed. **Demand Drafts must be complied with CTS 2010 standards prescribed by Reserve Bank of India		
Checklist – Technical Offer			
Sr. No.	Particulars	Documents to be attached wherever applicable	
		Yes	No
1.	1) Attached documents as required in the tender document (i.e. Supplier/firm is manufacturer/authorized dealer/sole distributor certificate, 2) Authorization certificate from the manufacturer in case of dealer / distributor. 3) Copy of PAN, 4) Certificate of firm/company registration, 5) TIN/VAT registration (Sales tax), 6) Income Tax Return of last three years, 7) Tenderer must provide experience / supplied as per the clause 6. 8) Annual Turnover & balance sheet of last three years duly certified by C.A. as per the clause 7. 9) Certificate of European CE, USFDA as per Annexure-I . 10) Certificate of warranty on letter head as per Annexure-I . 11) Tender document duly seal and sign by the tenderer. 12) Tenderer must provide a certificate on letter head that proprietor /firm has never been black listed by any organization. 13) Relevant brochure/catalogue pertaining to the items quoted with full specifications etc. 14) Company/Manufacturer/Firm should have branch office within Chhattisgarh Provide address in detail. If not undertaking for full support & maintenance on letter head. 15) Undertaking as per Clause 1(b) of page no. 06.		
2.	Technical Specifications Compliance Report		
3.	Duly filled Form – A, Form –B & Form-C		

PARTICULARS FOR PERFORMANCE GUARANTEE BOND

(To be typed on Non-judicial stamp paper of the value of Indian Rupees of Two Hundred) (TO BE ESTABLISHED THROUGH ANY OF THE SCHEDULED BANK (WHETHER SITUATED AT RAIPUR OR OUTSTATION) WITH A CLAUSE TO ENFORCE THE SAME ON THEIR LOCAL BRANCH AT RAIPUR. BONDS ISSUED BY CO- OPERATIVE BANKS ARE NOT ACCEPTED.)

To,
The Director
All India Institute of Medical Sciences (AIIMS),
Tatibandh, GE Road,Raipur-492 099 (CG)

LETTER OF GUARANTEE

WHEREAS All India Institute of Medical Sciences (AIIMS) Raipur (Buyer) have invited Tenders vide Tender No..... Dt..... for purchase of AND WHEREAS the said tender document requires the supplier / firm (seller) whose tender is accepted for the supply of instrument / machinery, etc. in response thereto shall establish an irrevocable Performance Guarantee Bond in favour of "AIIMS Raipur" in the form of Bank Guarantee for Rs..... [10% (ten percent) of the purchase value] which will be valid for entire warranty period from the date of installation & commissioning, the said Performance Guarantee Bond is to be submitted within 30 (Thirty) days from the date of Acceptance of the Purchase Order.

NOW THIS BANK HEREBY GUARANTEES that in the event of the said supplier/firm (seller) failing to abide by any of the conditions referred to in tender document / purchase order/ performance of the instrument / machinery, etc. this Bank shall pay to All India Institute of Medical Sciences (AIIMS) Raipur on demand and without protest or demur Rs (Rupees.....).

This Bank further agrees that the decision of All India Institute of Medical Sciences (AIIMS) Raipur (Buyer) as to whether the said supplier/firm (Seller) has committed a breach of any of the conditions referred in tender document / purchase order shall be final and binding.

We, (name of the Bank & branch) hereby further agree that the Guarantee herein contained shall not be affected by any change in the constitution of the supplier/firm (Seller) and/ or All India Institute of Medical Sciences (AIIMS) Raipur (Buyer).

Notwithstanding anything contained herein:

- a. Our liability under this Bank Guarantee shall not exceed Rs.....(Indian Rupees only).
- b.This Bank Guarantee shall be valid up to(date) and
- c. We are liable to pay the guaranteed amount or any part thereof under this bank guarantee only and only if AIIMS Raipur serve upon us a written claim or demand on or before.....(date)

This Bank further agrees that the claims if any, against this Bank Guarantee shall be enforceable at our branch office at situated at.....(Address of local branch).

Yours truly,

Signature and seal of the Guarantor

Name of the Bank:.....

Complete Postal Address:

ANNEXURE-I**Technical Specifications of Surgical Loups for various department****Schedule-1 (Department of Burn & Plastic Surgery)**

Surgical Loup: high end with illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Purpose	Warranty
<p>Specification of Surgical loupes: high end with illumination accessories</p> <p>The Surgical System should mainly consist of the optical system, adjustable HEAD MOUNT FRAME and standard accessories</p> <p>OPTICAL SYSTEM:</p> <ol style="list-style-type: none"> Kepler optical system featuring compact design, delivering precise image with good colour fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. Loupe must consist of 1 optical system (Kepler), 1 carrier system KF titanium, 1 instructions of use with brief instructions; incl. softcase, contact guards, protective caps and sideshields. Magnification: 4.5X. Working Distance (Eye to operating area) of 400 mm (16 inches) or more Field of view diameter at working distance of 400 mm: 86mm or more. Quick adjustment of interpupillary distance from 55mm to 75 mm by means of left and right coaxial knobs to suit individual surgeons. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection. 	01	For plastic surgery cases under magnification	2 Years

Description of Equipment	Quantity Required	Purpose	Warranty
<p>7. High quality surface of the optical system resistant to standard disinfectants.</p> <p>8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment.</p> <p>9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options.</p> <p>10. Flip up function for unobstructed vision and eye contact with patients with single adjustment.</p> <p>11. Mount to a high quality medical grade HEAD MOUNT frame.</p> <p>12. Serializable contact guard for reliable swinging of the optical system up and down.</p> <p>LED LIGHT ILLUMINATION</p> <p>1. Should include illumination system attached to the loupes with easy detachability option.</p> <p>2. Powerful LED light source with integrated temperature control.</p> <p>3. Should be able to attach easily to the optical system.</p> <p>4. Light intensity up to 50000 Lux.</p> <p>5. Shock proof protection for the light and accessories.</p> <p>6. Light should be resembling day light and should illuminate the whole field with even illumination.</p> <p>7. Extremely flexible with outstanding light transmission.</p> <p>8. Should include 2 Medical grade Lithium ion rechargeable battery with charging device. Battery charger should be processor controlled with country specific power adapter. Charge level indicator should be present.</p>			

Description of Equipment	Quantity Required	Purpose	Warranty
<p>9. Should have a run time of above 3 - 4 hrs. at 100 % intensity.</p> <p>10. Belt clip for the freedom of movement.</p> <p>11. All the necessary accessories for the proper functioning of the unit and suitable high quality soft case for storage.</p> <p>STANDARD ACCESSORIES:</p> <p>Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, accessories for the LED light Illumination, High quality soft case for protection of the surgical loupe and accessories</p> <p>MANUFACTURER AND COMPLIANCE</p> <p>1. The Loupe should be manufactured by a well-known international manufacturer /company.</p> <p>2. Should be European CE or US FDA approved.</p> <p>3. Should carry at least 2 years of warranty.</p>			

Schedule-2 (Department of Burn & Plastic Surgery)

Surgical Loups: without illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Purpose	Warranty
<p>Surgical Loupes: without illumination Accessories</p> <p>The Surgical System should mainly consist of the optical system, adjustable TITANIUM frame and standard accessories</p> <p>OPTICAL SYSTEM:</p> <ol style="list-style-type: none"> 1. Kepler optical system featuring compact design, delivering precise image with good colour fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. Loupe must consist of 1 optical system (Kepler), 1 carrier system KF titanium, 1 instructions of use with brief instructions; incl. softcase, contact guards, protective caps and sideshields. 2. Magnification: 4.3 X. 3. Working Distance (Eye to operating area) of 400 mm or more. 4. Field of view diameter at working distance of 400mm: 65mm or more. 5. Quick adjustment of interpupillary distance from 55mm to 75 mm by means of left and right coaxial knobs to suit individual surgeons. 6. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection. 7. High quality surface of the optical system resistant to standard disinfectants. 8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment. 9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options. 	01	For plastic surgery cases under magnification	2 Years

Description of Equipment	Quantity Required	Purpose	Warranty
<p>10. Flip up function for unobstructed vision and eye contact with patients with single adjustment.</p> <p>11. Mount to a Medical grade Titanium Frame for easy attachment and removal of the optical system.</p> <p>12. Contact guard for reliable swinging of the optical system up and down.</p> <p>13. Should have compatibility with LED Light source.</p> <p>SPECTACLE SYSTEM:</p> <ol style="list-style-type: none"> 1. Sturdy Medical Grade Titanium base high quality for wearing comfort. 2. Should be tied with a light weight belt for support to of the operator’s head. 3. Should provide even weight distribution on operator’s head. 4. Should be able to lock the loupe in the desired viewing angle. <p>STANDARD ACCESSORIES:</p> <p>Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, High quality soft case for protection of the surgical loupe and accessories.</p> <p>MANUFACTURER AND COMPLIANCE:</p> <ol style="list-style-type: none"> 1. The Loupe should be manufactured by a well-known international manufacturer /company. 2. Should be European CE or US FDA approved. 3. Should carry at least 2 years of warranty. 			

Schedule-3 (Department of Otorhinolaryngology)

Surgical Loup: high end with illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Purpose	Warranty
<p>SURGICAL LOUPS: HIGH END WITH ILLUMINATION ACCESSORIES</p> <p>The Surgical System should mainly consist of the optical system, adjustable on a head frame and standard accessories</p> <p>OPTICAL SYSTEM:</p> <ol style="list-style-type: none"> 1. Galilean optical system featuring compact design, delivering precise image with good colour fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. 2. Magnification approx. : 4.5 X 3. Working Distance (Eye to operating area) of 350 mm or more 4. Field of view diameter at working distance of 350 mm: 55mm or more 5. Quick adjustment of interpupillary distance from 55mm to 80 mm by means of left and right coaxial knobs to suit individual surgeons. 6. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection 7. High quality surface of the optical system resistant to standard disinfectants 8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment. 9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options 10. Flip up function for unobstructed vision and eye contact with patients with single adjustment 	01	For better Magnification during surgery	1 Year

Description of Equipment	Quantity Required	Purpose	Warranty
<p>11. Mount for easy attachment and removal of the optical system</p> <p>12. Serializable contact guard for reliable swinging of the optical system up and down.</p> <p>LED LIGHT ILLUMINATION:</p> <p>1. Should include illumination system attached to the loupes</p> <p>2. Powerful LED light source with integrated temperature control</p> <p>3. Should be able to attach easily to the optical system</p> <p>4. Light intensity up to 50000 Lux</p> <p>5. Shock proof protection for the light and accessories</p> <p>6. Light should be resembling day light and should illuminate the whole field with even illumination.</p> <p>7. Low weight</p> <p>8. Should include 2 Medical grade Lithium ion rechargeable battery with charging device</p> <p>9. Should have a run time of above 3 – 4 hrs. at 100 % intensity.</p> <p>10. Belt clip for the freedom of movement.</p> <p>11. All the necessary accessories for the proper functioning of the unit and suitable high quality soft case for storage.</p> <p>HEAD BAND</p> <p>1. Sturdy head band with washable textile pads for wearing comfort and easy cleaning.</p> <p>2. Should be able to adjust to the shape of the operator’s head.</p> <p>3. Should provide even weight distribution on operator’s head.</p> <p>4. Should be able to lock the loupe in the desired viewing angle.</p>			

Description of Equipment	Quantity Required	Purpose	Warranty
<p>STANDARD ACCESSORIES:</p> <p>Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, accessories for the LED light illumination, High quality soft case for protection of the surgical loupe and accessories.</p> <p>MANUFACTURER AND COMPLIANCE:</p> <ol style="list-style-type: none"> 1. The Loupe should be manufactured by a well-known international manufacturer/ company 2. Should be European CE or US FDA approved 3. Should carry at least 2 years of warranty 			

Schedule-4 (Department of Otorhinolaryngology)

Surgical Loups: without illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Purpose	Warranty
<p>SURGICAL LOUPES: WITHOUT ILLUMINATION ACCESSORIES</p> <p>The Surgical System should mainly consist of the optical system, adjustable on a head frame and standard accessories</p> <p>OPTICAL SYSTEM:</p> <ol style="list-style-type: none"> 1. Galilean optical system featuring compact design, delivering precise image with good color fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. 2. Magnification: 4 X 3. Working Distance (Eye to operating area) of 350 mm or more 4. Field of view diameter at working distance of 350 mm : 55mm or more 5. Quick adjustment of interpupillary distance from 55mm to 80 mm by means of left and right coaxial knobs to suit individual surgeons. 6. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection 7. High quality surface of the optical system resistant to standard disinfectants 8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment. 9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options 10. Flip up function for unobstructed vision and eye contact with patients with single adjustment 	01	For better Magnification during surgery	1 Year

Description of Equipment	Quantity Required	Purpose	Warranty
<p>11. Mount for easy attachment and removal of the optical system</p> <p>12. Seriazable contact guard for reliable swinging of the optical system up and down.</p> <p>13. Should have compatibility with LED Light source</p> <p>HEAD BAND:</p> <p>1. Sturdy head band with washable textile pads for wearing comfort and easy cleaning.</p> <p>2. Should be able to adjust to the shape of the operator’s head.</p> <p>3. Should provide even weight distribution on operator’s head.</p> <p>4. Should be able to lock the loupe in the desired viewing angle.</p> <p>STANDARD ACCESSORIES:</p> <p>Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, High quality soft case for protection of the surgical loupe and accessories.</p> <p>MANUFACTURER AND COMPLIANCE:</p> <p>1. The Loupe should be manufactured by a well-known international manufacturer/ company</p> <p>2. Should be European CE or US FDA approved</p> <p>3. Should carry at least 2 years of warranty</p>			

Schedule-5 (Department of Paediatric Surgery)Magnifying Loup (Qty: 01 No.)

Description of Equipment	Quantity Required	Purpose	Warranty
<p>MAGNIFYING LOUPE</p> <p>Specifications:</p> <ol style="list-style-type: none"> 1. Should be Galilean type, higher resolution, and wider field. 2. Should be magnification approx. : 2.5 X 3. Should be customized. 4. Should be configurations type: TTL(Through the lens) 5. Should be field Width & Depth: 10 cm or more 6. Weight shouldn't be more than 50 grams 7. Should be customized working distance between 10" and 18" as per requirement. 8. Should be waterproof 9. Should be hypo allergenic 10. Should be corrosion resistant 11. Should be unique adjustable good quality nose pad 12. Should be lightweight frame made of Titanium in variety of design 13. Should be unparalleled in strength 14. Should be European CE and US FDA approved 15. Should be provided with suitable box 	01	For better Magnification during surgery	3 Years

Schedule-6 (Department of Paediatric Surgery)Magnifying Loup: high end with illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Purpose	Warranty
<p>MAGNIFYING LOUPE WITH LED LIGHT</p> <p>Specifications for Magnifying Loupe</p> <ol style="list-style-type: none"> 1. Should be Galilean type, higher resolution, and wider field. 2. Should be magnification approx. : 2.5 X 3. Should be customized. 4. Should be configurations type: TTL(Through the lens) 5. Should be field Width & Depth: 10 cm or more 6. Wight shouldn't be more than 50 grams 7. Should be customized working distance between 10" and 18" as per requirement. 8. Should be waterproof 9. Should be hypo allergenic 10. Should be corrosion resistant 11. Should be unique adjustable good quality nose pad 12. Should be lightweight frame made of Titanium in variety of design 13. Should be unparalleled in strength 14. Should be European CE and US FDA approved 15. Should be provided with suitable box. <p>Specifications for LED LIGHT</p> <ol style="list-style-type: none"> 16. LED Light should be powerful and should generate up to 7000 foot candles of light for brilliant illumination. 17. Should be portable, eliminates being tethered to your power source. 18. Should allow complete station to station mobility. 19. Should be light weight. 20. Should be comfortable to wear, even during all 	01	For better Magnification during surgery	3 Years

Description of Equipment	Quantity Required	Purpose	Warranty
<p>day use.</p> <p>21.Should deliver superior brightness, sharpness and field depth</p> <p>22.Should be facility to attach through – the-lens loupe configurations.</p> <p>23.Should be coaxial mounting to eliminate shadows from the operating field.</p> <p>24.Should be push button type mechanism for power & light intensity.</p> <p>25.Should be 3 stage brightness control facility, 3000, 5000 & 7000 candle foot.</p> <p>26.Should be lithium Ion type battery, weight not more than 250 gms.</p> <p>27.Should be runtime of the battery from 7 hrs. To 17 hrs. depending on intensity.</p> <p>28.Should be supplied complete with LED light, cables, battery, charger & carrying case.</p> <p>29.Should be European CE / US FDA approved.</p>			

Schedule-7 (Department of Trauma & Emergency)

Surgical Loup: high end with illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Warranty
<p>Specification of Surgical loupes: high end with illumination accessories</p> <p>The Surgical System should mainly consist of the optical system, adjustable Titanium eye glass frame and standard accessories</p> <p>OPTICAL SYSTEM:</p> <ol style="list-style-type: none"> 1. Kepler optical system featuring compact design, delivering precise image with good color fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. Loupe must consist of 1 optical system (Kepler), 1 carrier system KF titanium, 1 instructions of use with brief instructions; incl. soft case, contact guards, protective caps and side shields. 2. Magnification: 2.5 X. 3. Working Distance (Eye to operating area) of 400 mm (16 inches) or more 4. Field of view diameter at working distance of 400 mm: 86mm or more. 5. Quick adjustment of interpupillary distance from 55mm to 75 mm by means of left and right coaxial knobs to suit individual surgeons. 6. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection. 7. High quality surface of the optical system resistant to standard disinfectants. 8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment. 9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options. 	01	3 Years

Description of Equipment	Quantity Required	Warranty
<p>10. Flip up function for unobstructed vision and eye contact with patients with single adjustment.</p> <p>11. Mount to a high quality medical grade titanium eyeglass frame.</p> <p>12. Sterializable contact guard for reliable swinging of the optical system up and down.</p> <p>LED LIGHT ILLUMINATION:</p> <ol style="list-style-type: none"> 1. Should have LED refractive Technology. 2. It Should include professional Headband- Occipital Basket type which Balances and distributes the weight of the headlight and reduces neck strain 3. Cranial support of head band should stabilizes the headlight, adjustable headband without temporal compression 3- point support system for comfortable stability with Extended linkage for better Vertical alignment with surgical loupe . 4. Head band should be Very Lightweight and flexible with less than 3.0mm diameter cable • Bifurcated design allows even weightdistribution and balance • 5. Powerful LED light source with integrated temperature control. 6. Should be able to attach easily to the optical system. 7. Light intensity 2,25000+ lux of pure white homogenous light (at least 6100 degrees Kelvin) for true rendition and best illumination for deep cavity surgery specially Cardiac surgery purpose. Illumination Spot adjustable from 2” to 5”.LED Lamp last up to 50,000hrs. 8. Shock proof protection for the light and accessories. 9. Light should be resembling day light and should illuminate the whole field with even illumination. 10. Extremely flexible with outstanding light transmission. 11. Should include 2 Medical grade Lithium ion rechargeable batteries with charging device. Battery 		

Description of Equipment	Quantity Required	Warranty
<p>charger should be processor controlled with country specific power adapter. Charge level indicator should be present.</p> <p>12. Should have a run time of above 3 - 5 hrs. at 100 % intensity.</p> <p>13. Belt clip for the freedom of movement.</p> <p>14. All the necessary accessories for the proper functioning of the unit and suitable high quality soft case for storage.</p> <p>STANDARD ACCESSORIES: Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, accessories for the LED light Illumination, High quality soft case for protection of the surgical loupe and accessories</p> <p>MANUFACTURER AND COMPLIANCE:</p> <p>1. The Loupe should be manufactured by a well-known international manufacturer/company.</p> <p>2. Should be European CE or US FDA approved.</p> <p>3. Should carry at least 2 year Warranty for Surgical loupe,3 years of warranty for LED and 1 year warranty for battery.</p>		

Schedule-8 (Department of Trauma & Emergency)Surgical Loups: without illumination accessories (Qty: 01 No.)

Description of Equipment	Quantity Required	Warranty
<p>Surgical Loupes: without illumination Accessories</p> <p>The Surgical System should mainly consist of the optical system, adjustable frame and standard accessories</p> <p>Optical System:</p> <ol style="list-style-type: none"> 1. Kepler optical system featuring compact design, delivering precise image with good colour fidelity extending to the peripheral zones, excellent depth of field of view ensuring clear visualization of anatomical structures. Loupe must consist of 1 optical system (Kepler), 1 carrier system KF titanium, 1 instructions of use with brief instructions; incl. soft case, contact guards, protective caps and side shields. 2. Magnification: 4.0 X. 3. Working Distance (Eye to operating area) of 400 mm or more. 4. Field of view diameter at working distance of 400mm: 65mm or more. 5. Quick adjustment of interpupillary distance from 55mm to 75 mm by means of left and right coaxial knobs to suit individual surgeons. 6. High Quality scratch free lens protection device with anti-reflection coating shielding the objective lens against tissue debris for increased protection. 7. High quality surface of the optical system resistant to standard disinfectants. 8. Telescopic rail enabling quick positioning of the eye piece with a single adjustment. 9. Quick adjustment of eyepiece tilt to desired viewing angles even in extreme treatment options. 10. Flip up function for unobstructed vision and eye contact with patients with single adjustment. 	01	2 Years

Description of Equipment	Quantity Required	Warranty
<p>11. Mount to a Medical grade Titanium Frame for easy attachment and removal of the optical system.</p> <p>12. Contact guard for reliable swinging of the optical system up and down.</p> <p>13. Should have compatibility with LED Light source.</p> <p>Spectacle System</p> <p>1. Sturdy Medical Grade Titanium base high quality for wearing comfort.</p> <p>2. Should be tied with a light weight belt for support to of the operator’s head.</p> <p>3. Should provide even weight distribution on operator’s head.</p> <p>4. Should be able to lock the loupe in the desired viewing angle.</p> <p>STANDARD ACCESSORIES:</p> <p>Objective lens protective device (2), Contact Guard (2), cleaning cloth for optical components, Allen Key, High quality soft case for protection of the surgical loupe and accessories.</p> <p>MANUFACTURER AND COMPLIANCE</p> <p>1. The Loupe should be manufactured by a well-known international manufacturer/company.</p> <p>2. Should be European CE or US FDA approved.</p> <p>3. Should carry at least 2 years of warranty.</p>		

 (Signature & Stamp of the bidder)

Note- Please sign each page of document including terms & conditions & tender

ANNEXURE-II (A)

PRICE SCHEDULE FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA

Schedule	Item Description	Qty	Unit Price in INR	Taxes (if applicable) VAT / Sales Tax/etc in INR	Any Other Charges in INR if applicable (Specify)	Unit Price including tax and other charges 7=4+5+6	Total in INR (Unit price x Quantity + Other Charges) 7x3
1	2	3	4	5	6	7	8
* Total							

Total Tender price in Rupees:..... In words

.....

- Note:** - 1. The schedule wise total * will be the deciding factor for L1 provided all the other conditions mentioned in the tender document are fulfilled.
 2. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
 3. L1 will be decided Schedule wise separately.
 4. Price Bid should be submitted in given format only. For additional information/extra items above format may be typed and used.

Name :

Place:.....

Business Address:.....

Date :.....

Signature and Seal of Tenderer

ANNEXURE-II (B)
PRICE SCHEDULE FOR GOODS TO BE IMPORTED FROM ABROAD

1	2	3	4	5					6
				Price per unit / basis at AIIMS, Raipur					
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	FOB price at port/ airport of Lading (a)	Carriage & Insurance (port of loading to port of entry) and other Incidental costs** (b)	Incidental services including installation & commissioning, supervision, Demonstration and Training) at the Consignee's site (c)	Transportation from port of entry to the consignee site, storage and extended Insurance for a period of 3 months beyond date of delivery** (d)	Unit Price on FOR of Destination + Extended Insurance (local transportation and storage) (e)	Total price on FOR of Destination + Insurance (local transportation and storage) 4 x 5
*Total									

Total Tender price in foreign currency: _____

In words: _____

- Note: -**
1. The schedule wise total * will be the deciding factor for L1 provided all the other conditions mentioned in the tender document are fulfilled.
 2. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
 3. The tenderer will be fully responsible for the safe arrival of the goods at the named port of entry in good condition as per terms of CIP as per INCOTERMS, if applicable. Purchaser will make payment after receipt of goods at consignee's site/store in good condition
 4. Price Bid should be submitted in given format only. For additional information/extra items above format may be typed and used.
 5. All the component of total price (i.e. column no. 6 will be paid by tenderer including Custom Duty, Custom Clearing charges as per terms of contract. CDEC will be issued to tenderer on demand.

Indian Agent:

Indian Agency Commission - __% of FOB

Signature of Tenderer _____

Place: _____

Date: _____

Name _____

Business Address _____

Signature of Tenderer _____

Seal of the Tenderer _____

SPECIAL CONDITIONS

1. The bidders should submit the relevant certificates / evidence of previous supply to other hospitals. If required, the technical committee may enquire from the other hospitals where the bidders have supplied the material.
2. The bidders should also submit a certificate from the relevant authority as to the quality of the equipment.
3. The bidder should not have been blacklisted before.
4. The bidders shall also arrange for the demonstration of their equipment to the concerned committee regarding the quality aspect.

The above quote should be made schedule wise and should include all applicable taxes. Schedule wise comparison of the quotes will be made and L1 for each item will be determined accordingly. If the rates of L1 are quoted for various items from different bidders, then AIIMS Raipur reserve the right to either accept the L1 of different firm/agencies or will negotiate, with the firm who has quoted the maximum gross value of L1 items to lower the rate of other item up to the limit of L1 quoted by other firms, provided that such quoted items are not interdependent on each other. In this context, final decision of the committee will be binding to all and no claim in this regard can be entertained. The quantity indicated is tentative and may vary, and any decision in this regard by Director AIIMS Raipur shall be final.

Declaration by the Bidder:

1. This is to certify that I/We before signing this tender have read and fully understood all the terms and conditions contained in Tender document regarding terms & conditions of the contract, rules regarding tender for Supply of Surgical Loups at AIIMS Raipur. I/we agree to abide them.
2. No other charges would be payable by Client and there would be no increase in rates during the Contract period.

Place:.....

(Signature of Bidder with seal)

Date:.....

Name :

Seal :

Address :

MANUFACTURER's / PRINCIPAL's AUTHORIZATION FORM

To
The Store Officer,
All India Institute of Medical Sciences Raipur

Dear Sir,

TENDER: _____.

we, _____, who are established and reputable manufacturers of _____, having factories at _____ and _____, hereby authorize (Authorized Dealer/Distributor/Supplier) Messrs. _____ (name and address of agents) to bid, negotiate and conclude the contract with you against Tender No. _____ for the above goods manufactured by us. No company or firm or individual other than Messrs. _____ are authorized to bid, negotiate and conclude the contract in regard to this business against this specific tender.

We hereby extend our full guarantee and warranty as per the conditions of tender for the goods offered for supply against this tender by the above firm.

The authorization is valid up to _____

Yours faithfully,

(Name)

For and on behalf of Messrs. _____

(Name of manufacturers)/Principal