



अखिल भारतीय आयुर्विज्ञान संस्थान ,रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Admin/Store/Purchase/Instrument/2015 AIIMS-RPR

Date: 19.02.2015

Inviting Quotations for Supply of Instrument for Department of obstetrics & Gynecology, AIIMS, Raipur

QUOTATION NOTICE

Sealed quotations are invited from intending registered Stockist / Distributors having TIN and relevant documents for Supply of Instrument for Department of obstetrics & Gynecology. Should be submitted to office of Store officer on 25.02.2015 Up to 3:00 pm. The quotations will be opened on the same day at 03.15 pm. The list should cover the following Items:-

	Item:	Specification	Quantity Required	Unit Rate in RS.
		Equipments		
1	fetal Doppler	<p>1. FETAL Doppler: Product Specification Safety:- should Comply with: IEC 60601-1:1988+A1:1991 +A2:1995</p> <p>Should be suitable for use after the 12th week of pregnancy.</p> <p>Physical Characteristic</p> <p>LCD display</p> <p>Should be both battery operated and electrically operated</p> <p>FHR Performance</p> <p>FHR Measuring Range: 50~240BPM (BPM: beat per minute) Resolution: 1BPM Accuracy: ±2BPM Power consumption:<0.8W Auto Shut-OFF: After 1 minute no signal, power off automatically. Battery Type Recommended: Two pieces of 1.5 V DC battery (SIZE AA LR6).</p> <p>Probe:</p> <p>Nominal Frequency: 2.0MHz Working Frequency: 2.0MHz±10% P-: <0.5MPa</p>	4 nos.	

		<p>Iob: <10 mW/cm² Ispta: <50mW/cm² Ultrasonic Output Intensity: Isata<5 mW/cm² Ultrasonic Output Power: P<10 mW Working Mode: Continuous wave Doppler Effective Radiating Area of Transducer: 157mm²±15%</p>		
2	Suction apparatus (Electrically operated)	<p>Should be FDA/CE certified. Manufacturer's name should be encrypted in the machine</p> <p>Specification:-</p> <ul style="list-style-type: none"> • Wide application range • Heavy Duty motorized pump- for fast vacuum built-up • Absolutely noiseless • Finely regulable-0715mm+- 10 hg vacuum-easy to read off • Over floe safety Device over – floe check valve and removable filter to protect the pump from moister. • Non collapsible tubing. • Self lubricating system for trouble free services. <p>Technical Specification. Max vacuum 715 _+ 10mm Hg regulable through Vacuum control knob. Max Vacuum 35 lits / MIN. Vacuum gauge bourdon Type 3" 0-760mm Hg calibration. Suction bottle Galss jars or 2 x 2000cc poly carbonate jars (Unbraekable) with over flow check valve (Float Type) Motor ¼ hp motor vacuum. Pump Rotary type. Overall size L330mm x W 405mm x H730mm (Approx). Finish powder coated.</p>	3 nos.	
3	Spot light LED	<p>SPOT LIGHT :Perfect for the OB/GYN office or for specialties that require very focused light in specific areas, the GS Exam Light IV should provide with 3X the typical light output and features a highly flexible light pipe that should be twice as long as conventional pipes. And with a compact, unobtrusive head, it ensures pinpoint accuracy and should provide true tissue color rendition during an exam.</p> <ul style="list-style-type: none"> • Should have LED Lamp • Should be quoted with 5 years warranty for the LED source and power supply • Should be provided with wall mounted as well stand 	2 nos.	

- Taxes, if any (Kindly Mention)-
- Delivery Schedule – within 21days from the date of issue of PO.
- Price should be FOR- Destination basis
- LD @ 0.5% of PO value per week or Part of week for delay of supply of material subject to maximum up to 10% of PO value.

Store Officer
AIIMS, Raipur (C.G.)