



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
 Tatibandh, GE Road,
 Raipur-492099 (CG)
www.aiimsraipur.edu.in

Tender Enquiry No. ADMIN/Tender/Ophthalmology_Instruments/1/2013

Corrigendum for Ophthalmology Instrument

Clause	Existing Parameter	Amendment
Pg 1 Clause 3	Close for Bidding – Submission of Tender 23.10.2013 15:00	Close for Bidding – Submission of Tender 26.10.2013 12:30 PM & Opening at same date 1:00 PM
Pg 5,6,7,8,9	FINANCIAL BID FOR SCHEDULE – A/B/C/D and Price Schedule for A.M.C and C.M.C	Delete from here and it is attached as annexure I, II, III
Page no. 25 Point 1	1. Slit width : 0-14 mm , continuous 2. Slit length: 1 –14mm adjustable continuous 5. Diaphragm sizes : 0.2 – 14mm 7. Light source: halogen lamps 17. Slit lamp camera – 12.2 mega pixel	1. Slit width : 0-8 mm / 0-14mm , continuous 2. Slit length: 1 –8 mm / 1-14mm adjustable continuous 5. Diaphragm sizes : 0.2 – 8 mm/ 0.2 – 14mm 7. Light source: halogen lamps / Tungsten / LED 17. Slit lamp camera – 12.2 mega pixel / Integrated camera with high resolution
Page no. 25 Point 2	IOL Master	IOL Master / Optical biometer
Page no. 26	Specifications: Measurement range: i. Axial length 14-38mm ii. Anterior chamber depth 1.5-6.5mm 8. The equipment should have integrated hardware & software (CPU) within the single unit.	Specifications: Measurement range: i. Axial length 14-32mm ii. Anterior chamber depth 1.5-5.5mm 8. The equipment should have integrated hardware & software (CPU) within the single unit / External unit
Page no. 26 Point 3	Humphery Field Analyser	Visual field analyser

Clause	Existing Parameter	Amendment
Page no. 26	Specifications: 2. Computer & Monitor should be integrated in the perimeter (No external computer). 7. Threshold test strategies Full threshold, Fast Pac, SITA, SITA for SWAP	Specifications: 2. Computer & Monitor should be integrated in the perimeter / External Computer. 7. Fast Screening strategies Full threshold, Fast Pac, SITA, SITA for SWAP / (Blue on Yellow)
Page no 27	15. Internal hard disk drive, backup on USB drive 23. HFA Net connects software.	15. Internal hard disk drive, backup on USB drive or external hard disk drive with DVD writer 23. HFA Net connects software. / Database software DICOM facility
Page no 27	1. BIOMETER (A-B Scan) WITH PACHYMETER:	1. BIOMETER (A-B Scan) WITH PACHYMETER / BIOMETER (A-B Scan) and PACHYMETER (can be separately supplied)
Page no 28	2. Auto Refractometer with Keratometer: Specifications: 1. Objective and subjective mode & measuring corneal astigmatism, low contrast glare acuity testing.	2. Auto Refractometer with Keratometer: Specifications: 1. Objective ± subjective mode & measuring corneal astigmatism, low contrast glare acuity testing.

“Ophthalmology Instruments”
AIIMS, Raipur
FINANCIAL BID for SCHEDULE –A/ B/ C/ D/E/ F

S. No.	Item Description	Unit Price in INR		Custom duty in INR (if applicable)		Taxes (if applicable) VAT / Sales Tax/etc in INR		Service Tax (if applicable) in INR		Qty of units	Any Other Charges in INR if applicable (Specify)		Total in INR (Unit price x Quantity + Other Charges)	
		Figures	Words	Figures	Words	Figures	Words	Figures	Words		Figures	Words	Figures	Words
	Equipment													
	Accessories													
CMC Total for Five Years after warranty period	Name of the instrument	1												
		2												
		3												
Grand Total*														

The grand total * will be the deciding factor for L1 provided all the other conditions mentioned in the tender document (GCC, SCC and the technical conditions) are fulfilled. Total CMC of 5 years of each instrument, which is costing 5 lakhs or more, will also be a part of grand total to decide L1. (AMC value should not be mentioned here)

DATE:

SIGNATURE

NAME

SEAL

PRICE SCHEDULE FOR COMPEREHNSIVE MAINTENANCE CONTRACT
(C.M.C.)
AFTER EXPIRY OF WARRANTY
(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)

S. No.	SME Code No.	Name of the Equipment	For Sixth year with spare parts & labour	For Seventh year with spare parts & labour	For Eighth year with spare parts & labour	For Ninth year with spare parts & labour	For Tenth year with spare parts & labour
(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)

PRICE SCHEDULE FOR ANNUAL MAINTENANCE CONTRACT
(A.M.C.)
AFTER EXPIRY OF WARRANTY
(RATES SHOULD BE QUOTED IN INDIAN RUPEES ONLY)

S. No.	SME Code No.	Name of the Equipment	For Sixth year with spare parts & labour	For Seventh year with spare parts & labour	For Eighth year with spare parts & labour	For Ninth year with spare parts & labour	For Tenth year with spare parts & labour
(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)