



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)  
All India Institute of Medical Sciences, Raipur (Chhattisgarh)  
Tatibandh, GE Road,  
Raipur-492 099 (CG)  
[www.aiimsraipur.edu.in](http://www.aiimsraipur.edu.in)

No. AIIMS/RPR/Port.Dopp./2013-14/RPR


Raipur, the 09<sup>th</sup> Sept., 2013

### **Corrigendum Technical Specifications**

With reference to the tender AIIMS/RPR/Port.Dopp/2013-14 and the corrigendum issued before the following amendments are being made. The last date of submission of the bids is extended to 18.9.2013, 15.00 Hrs.

S. No.	Page / Clause point	Existing parameter	Correction
	<b>Technical Specifications - 3</b>	Monitor should be 10" or more, high- resolution Colour Monitor with VGA Format Display size 640x480, Recording area 640x480, adjustable tilt, Stereo sound via inbuilt and integrated loudspeaker dynamic upto 90 db or more,	Monitor should be of 12" or more, high-resolution Colour Monitor and integrated loudspeaker dynamic upto 160 db or more,  Also mention whether the system has an additional monitor.
		Cine loop memory more than 100 frames	Cine loop memory should be of 30seconds or more.
	<b>Clause 5</b>	<b>Transducers</b>	<b>Transducers:</b> The transducers should be light weight

**Last date of submission of bid is extended to 18-09-2013.**

  
Deputy Director Administration  
AIIMS Raipur (CG)