



Corrigendum

Tender Enquiry No. ADMIN/Tender/Anaesthesia_Workstation/1/2014

Date: 28-11-2014

Page no /Clause no / Point no	Existing sentence	To be read as
Page no. 1, Sr. no. 1	ADMIN/Tender/Accessory Equipment & Furniture/2/2014	ADMIN/Tender/ Anaesthesia workstation/ 1/2014
Page no. 2, Sr. no. 1	All India Institute of Medical Sciences Raipur (CG) invites sealed tenders from eligible and qualified tenderers for supply, installation and commissioning of Low End Modular Operation Theatres and Medical Equipments.	All India Institute of Medical Sciences Raipur (CG) invites sealed tenders from eligible and qualified tenderers for supply, installation and commissioning of anaesthesia workstation.
Page no. 11, Sr. no. 9.1	At any time prior to the deadline for submission of tenders, the purchaser, for any reason deemed fit by it, modify the TE documents by issuing suitable amendment(s) to it.	At any time prior to the deadline for submission of tenders, the purchaser, for any reason deemed fit, may modify the TE documents by issuing suitable amendment(s) to it.
Page no. 23, Sr. no. 28	Minor Infirmity/Irregularity/Non-Conformity	Infirmity/Irregularity/Non-Conformity
Page no. 23, Sr. no. 28.1	If during the preliminary examination, the purchaser find any 'minor' infirmity and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenderers. Wherever necessary, the purchaser will convey its observation on such 'minor' issues to the tenderer by email/registered post/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.	If during the examination, the purchaser find any infirmity and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenderers. Wherever necessary, the purchaser will convey its observation on such issues to the tenderer by email/registered post/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

Page no /Clause no / Point no	Existing sentence	To be read as
Page no. 28	The special instructions to tenderers are same as described in SECTION II general instructions to tenderers.	During technical evaluation, the bidders can be asked to demonstrate the anaesthesia work station and the accessories they have offered.
Page no. 31, Sr. no. 5.3	In the event of any failure / default of the supplier with or without any quantifiable loss to the consignee, the bank Guarantee for CMC security as per Proforma in Section XV, and the amount of the performance security is liable to be forfeited. The Administration Department may do the needful to cover any failure/default of the supplier with or without any quantifiable loss to the consignee.	In the event of any failure / default of the supplier with or without any quantifiable loss to the consignee, the bank Guarantee for CMC security as per Proforma in Section XV, and the amount of the performance security is liable to be forfeited. The Administration shall do the needful to cover any failure/default of the supplier with or without any quantifiable loss to the consignee.
Page no. 32, Sr. no. 7.1	The packing for the goods to be provided by the supplier should be strong and durable enough to withstand, without limitation, the entire journey during transit including trans-shipment (if any), rough handling, open storage etc. without any damage, deterioration etc. As and if necessary, the size, weights and volumes of the packing cases shall also take into consideration, the remoteness of the final destination of the goods and availability or otherwise of transport and handling facilities at all points during transit up to final destination as per the contract.	The packaging for the goods to be provided by the supplier should be strong and durable enough to withstand, without limitation, the entire journey during transit including trans-shipment (if any), rough handling, open storage etc. without any damage, deterioration etc. As and if necessary, the size, weights and volumes of the packing cases shall also take into consideration, the remoteness of the final destination of the goods and availability or otherwise of transport and handling facilities at all points during transit up to final destination as per the contract.
Page no. 34, Sr. no. 8.8	Principal/ Foreign supplier shall also have the equipment inspected by recognized / reputed agency like SGS, Lloyd or equivalent (acceptable to the purchaser) prior to dispatch at the supplier's cost and furnish necessary certificate from the said agency in support of their claim.	Deleted
Page no. 36, Sr. no. 13.1 (III)	Incidental services Providing required jigs and tools for assembly, minor civil works required for the completion of the installation;	Incidental services Providing required jigs and tools for assembly, civil works required for the completion of the installation;

Page no /Clause no / Point no	Existing sentence	To be read as
Page no. 37, Sr. No. 14 B(VII)	Inspection Certificate for the dispatched equipments issued by recognized/ reputed agency like SGS, Lloyd or equivalent (acceptable to the purchaser) prior to dispatch	Inspection Certificate for the dispatched equipments issued by recognized / reputed agency (acceptable to the purchaser) prior to dispatch
Page no. 38, Sr. No. 15.2 (b)	<ul style="list-style-type: none"> ▪ Any kind of OT or 	<ul style="list-style-type: none"> ▪ Any kind of motor
Page no. 38, Sr. No. 15.5	In the event of any rectification of a defect or replacement of any defective goods during the warranty period, the warranty for the rectified/replaced goods shall be extended to a further period of twenty-four (24) months from the date such rectified / replaced goods starts functioning to the satisfaction of the purchaser.	Deleted
Page no. 39, Sr. No. 15.8	The Purchaser/Consignee reserve the rights to enter into Annual Comprehensive Maintenance Contract for the period as mentioned in Section VII, Technical Specifications after the completion of warranty period.	The Purchaser/Consignee reserve the rights to enter into Annual Comprehensive Maintenance Contract for the period as mentioned in Section VII.
Page no. 42, Sr. No. 21.1 B (IX)	Inspection Certificate for the dispatched equipments issued by recognized/reputed agency like SGS, Lloyd or equivalent (acceptable to the purchaser)prior to dispatch.	Deleted
Page no. 52 to 55	<u>TECHNICAL SPECIFICATIONS FOR ANESTHESIA WORK STATION</u>	Deleted
Page no. 56	Specifications / Particulars Competitive bids are invited for supply and installation and commissioning of anesthesia machine, the bidder should quote their mode which is state-of-the-art, top-of-the line system. The specifications & features given below should be essential provided. However, in case the company can offer any other technical features which are better than these features asked in the tender, they should be quoted separately as these will also be considered. All commercially available software for clinical use should be supplied the time of installation. All future comprehensive software upgrade	Deleted

Page no /Clause no / Point no	Existing sentence	To be read as
	guarantee must be provided free of cost.	
Page no. 56 2 nd para	Anesthesia machine, ventilator, patient monitor and charting solution should be from the same manufacturer.	Anesthesia machine, ventilator, vaporizer , patient monitor and charting solution should be from the same manufacturer.
Page no. 57 Point ii.	The anesthesia machine should provide for desflurane compensation.	Deleted.
Page no. 57 Point no. X	Should be provided with two or more drawers.	Should be provided with one or more drawers.
Page no. 57 Point no. I	Should have 12" color touch screen monitor and displayed parameters should be color coded.	Should have 12" or more color screen monitor and displayed parameters should be color coded.
Page no. 57 Add point no. xiv & xv	-	xiv. Should provide Multi – Gas analysis with auto detection of all anesthetic agents. xv. Should be able to automatically detect and calculate MAC of all anaesthetic gases.
Page no. 58 Point no. VII	Should have SIMV and pressure support mode of ventilation	Should have SIMV and/with pressure support mode of ventilation
Page no. 58 Point no. iii	Pressure VS volume	Pressure VS volume (Loop)
Page no. 58 Point no. 7 & 8	7. Should provide Multi – Gas analysis with auto detection of all anesthetic agents. 8. Should be able to automatically detect and calculate MAC of all anaesthetic gases.	Deleted from here and shifted to page no.57 as new point no. xiv & xv.
Page no. 59 Point no. 22	Should be ready to import clinical reports such as radiology films report from PACS or Hospital information system.	Should be ready to import clinical reports such as radiology images from web based PACS or Hospital information system.
Page no. 59 After Point no. 23	The Anesthesia work station should be provide with a keyboard and a mouse for manually entering user defined data. The system should be capable	The Anesthesia work station should be provide with a keyboard and a mouse for manually entering user defined data such as fluid intake/output/drugs administered/staffing etc. The system should be capable
Page no. 68 Section XI Coloumn 5(d)	Extended Insurance (local transportation and storage) from port of entry to the consignee site for a period including 3 months beynd date of delivery ** (d)	Transportation from port of entry to the consignee site, storage and extended Insurance for a period of 3 months beyond date of delivery** (d)

Administrative Officer
AIIMS, Raipur