



अखिल भारतीय आयुर्विज्ञान संस्थान

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

3

No.Admin./Port.Dopp./2013-AIIMS.RPR

Raipur, Dated 05th Sept. 2013


Corrigendum

CORRIGENDUM IN COMMERCIAL SECTION

S.No.	Page / Clause point	Existing parameter	Correction
1.	Page 3	"Last date for submission of original EMD and other relevant documents is 26.08.2013"	Last date for submission of original EMD and other relevant documents is 10.09.2013

CORRIGENDUM IN TECHNICAL SPECIFICATIONS

S. No.	Page / Clause point	Existing parameter	Correction
	Clause 3 in technical specifications	Monitor should be 9" or more, high resolution Colour monitor with VGA Format Display size...	The System must have integrated high – resolution TFT/LCD/Single monitor of 10 Inches or more.
	Physical and Ergonomics	Portable unit cannot have swivel facility	Mention whether the unit is locked or with swivel facility
	Scanning capabilities	Split Screen imaging (B+B,B+M,M+M,B+PWI/CDI preferable)	Mention whether triplex imaging is possible. Give details
	Accessories	B/W Thermal Printer	System should be supplied with Thermal printer and should be compatible to support coloured laser printer.


Deputy Director Administration
AIIMS Raipur (CG)