



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Tender no. ADMIN/Tender/ Cold_Light_Source_with_Biliblanket/2/2014 dated 19-12-2014
Corrigendum for Cold Light Source with Biliblanket

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Page no. 20 Point no. 2	Should have option for overhead spotlight as well as Fibreoptic panel	Should have Fibreoptic pad of large and small size for phototherapy of the baby to allow flexibility of area coverage and irradiance level. One pad of each size should be provided.

Store officer
AIIMS, Raipur