



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Tender enquiry no. **ADMIN/Tender/OPG/1/2013**

Corrigendum for OPG

A: Commercial Section

S. No	Clause No	Old	Updated
1.	Page no. 01 paragraph 02	“Tender For Dentistry Equipments” and should reach at the office of “The Administrative Officer, AIIMS, Tatibandh Raipur (CG) - 492001, by or before on 03.00 PM on 04-10-2013.	“Tender For OPG ” and should reach at the office of “The Administrative Officer, AIIMS, Tatibandh Raipur (CG) - 492001, by or before on 03.00 PM on 15-10-2013 .
2.	Page no. 01 Key Dates Table	Close for Bidding – Submission of Tender - 04-10-2013 15:00	Close of Bidding – Submission of Tender - 15-10-2013 15:00
3.	Page no. 01 Key Dates Table	Open EMD & Technical / PQ bid - 04-10-2013 15:00	Open EMD & Technical / PQ bid- 15-10-2013 15:00
4.	Page no. 01 paragraph 03	The amount of bid security (EMD) for Tender For Dentistry Equipments of Rs. 1,00,000/- (Rupees One Lakh) of tender documents should be paid by FDR/DD in favour of “AIIMS, Raipur” payable at Raipur and will be placed in cover-1 with technical bid.	The amount of bid security (EMD) for Tender For OPG of Rs. 1,00,000/- (Rupees One Lakh) of tender documents should be paid by FDR/DD/ BG in favour of “AIIMS, Raipur” payable at Raipur and will be placed in cover-1 with technical bid.

S. No	<u>Clause No</u>	<u>Old</u>	<u>Updated</u>
5.	Page no 02 point 13 of the Technical Bid	Details of the FDR/DD of bid security (EMD) FDR/DD No: Date: Payable at	Details of the FDR/DD/ BG of bid security (EMD) FDR/DD/ BG No: Date: Payable at
6.	Page no 08, point (A) 1.	not later the date of 04-10- 2013,	not later the date of 15-10-2013,
7.	Page no 08, point (A) 2.	third sealed cover super scribed "Tender for Purchase of "Dentistry Equipments" should reach AIIMS, Raipur by or before 03.00 PM on 04-10-2013.	third sealed cover super scribed "Tender for Purchase of OPG " should reach AIIMS, Raipur by or before 03.00 PM on 15-10-2013 .
8.	Page no 10, point (B) 2.	All the rates should be mention in Indian national currency (INR) only. The rates quoted in foreign currency will not be entertained in this tender enquiry & such tenders will be cancelled straightway.	<ul style="list-style-type: none"> • The payment shall be made in the currency / currencies authorized in the contract. • Irrevocable & non – transferable LC shall be opened by the respective consignees.

B: Technical Specification Section

<u>S. No</u>	<u>Clause No</u>	<u>Old</u>	<u>Updated</u>
1.	Page no 16, Point 1 - c.	Weight: It should be lightweight system with a weight 140 kg or less (inclusive of Cephalostat).	Weight: It should be lightweight system with a weight 200 kg or less (inclusive of Cephalostat).
2.	Page no 16, Point 2 - a. Comment Column	60KHz -150KHz or above	40KHz -150KHz or above.
3.	Page no 16, Point 3 - d.	Adjustable focal trough (Min 9-way adjustable.)	Adjustable focal trough. (Min 9-way adjustable). Mention whether focal trough adjustment is manual or automatic
4.	Page no 16, Point 3 - g.	Should have an automatic exposure control	Does the system has automatic exposure /dose control?
5.	Page no 16, Point 3 - l.	Magnification of panoramic exposure: 1.2-1.29	Should have a constant magnification for panoramic images with magnifications 1.3 or less.
6.	Page no 18, Point 5 - d. Console and Control	Should have a DICOM-3 compatibility for easy image transfer to the PACS/HIS through the networking for image transfer to view at various work station and to print from remote dry chemistry printers.	Should have a DICOM-3 compatibility for easy image transfer to the PACS/HIS through the networking for image transfer to view at various work station and to print from remote dry chemistry printers. Image output by USB/DVD/CD writer should be possible.
7.	Page no 18, Point 6 - a.	The image should be of excellent quality better than 1530 x 2730 pixels	The image should be of excellent quality better than 1530 x 2730 pixels. The image quality in pixels to be mentioned
8.	Page no 18, Point 6 - b.	The Cepahlometric Images should give true measurements with a Scale shown on each image. Magnification should be 1 and should have means for adjusting the magnifications.	The Cepahlometric Images should give true measurements with a Scale shown on each image. Magnification should not be greater than 1.15 and should have means for adjusting the magnifications.

S. No	<u>Clause No</u>	<u>Old</u>	<u>Updated</u>
9.	Page no 18, Point 8.		b. Mention whether the system has capability to give output on to USB device (storage Device)
10.	Page no 18, Point 9. a, Comment Column		With a viewing window.
11.	Page no 19, Point 9 - c.	UPS suitable for the operability back up of the complete equipment for 30 minutes.	<ul style="list-style-type: none"> • 3KVA Online UPS with 30 minute backup of the complete equipment, • 3KV servo stabilizer with Spike suppressor & isolation transformer.
12.	Page no 19, Point 9 - d.	Lead Apron – Ultra light weight (zero lead) with protection of 0.5 mm lead equivalent.	Lead Apron – Ultra light weight (zero lead) with protection of 0.5 mm lead equivalent. Quantity - 01
13.	Page no 19. Special Condition no 3	It will be the responsibility of the supplier to ensure compliance with AERB guidelines.	It will be the responsibility of the supplier to ensure compliance of the equipment and the installation site with AERB guidelines. The bidder would undertake necessary civil alteration if any. Also should quote the rate of lead lining for the doors and walls per square feet area in a separately in the financial bid.