



Tender enquiry no. ADMIN/Tender/ Ophthalmoscope\_Otoscope\_&\_Laryngoscope\_Equipments/1/2014

## Corrigendum

### Tender for “Ophthalmoscope, Otoscope & Laryngoscope Equipments”

#### DIRECT OPHTHALMOSCOPE (05 Nos.) SCHEDULE-A

No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 21 of Annexure-I	Illumination: 3.5V, 2.8W Mini Halogen Bulb, BT-3000	<b>Illumination: 3.5V, 2.8W Mini Halogen / Mini LED Bulb</b>
2	Page no. 21 Annexure-I	Viewing Lenses: 0, $\pm 1$ , $\pm 2$ , $\pm 3$ , $\pm 4$ , $\pm 5$ , $\pm 6$ , $\pm 8$ , $\pm 10$ , $\pm 12$ , $\pm 16$ , $\pm 20$ , $-25$ , $-35$	<b>Viewing Lenses: +42 to -35</b>

#### FIBREOPTIC OTOSCOPE (05 Nos.) SCHEDULE-B

No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 22 of Annexure-I	Otoscope with fibreoptic illumination. 3.5, volts Halogen bulb.	Otoscope with fibreoptic illumination. 3.5, volts <b>Halogen / Mini LED Bulb.</b>

#### LARYNGOSCOPES WITH ALL BLADES (21 Nos.) SCHEDULE-C

No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 22 of Annexure-I	Should have Xenon light illumination of more than 50000 Lux	Should have <b>Xenon / LED Light</b> illumination.
2	Page no. 22 of No. 1	<ul style="list-style-type: none"><li>Batteries should be rechargeable preferably induction charging</li><li>Charger to attach two laryngoscope handles for charging at one time</li><li>Life of the battery should be more than 120 minutes when fully charged</li><li>Charger should have no corrosion and contact problem</li><li>Batteries should have provision to be charged without taking out from the laryngoscope handle</li><li>Charger should have indicator to show the charging and charged mode.</li></ul> The product should be imported	<b>Deleted</b>