



Tender enquiry no. ADMIN/Tender/ Paediatric_Equipments/1/2015 **Dated 14/02/2015**

Corrigendum
Tender for “Paediatric Equipments”

SCHEDULE- A BUBBLE CPAP (2 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 22 Point no. A Patient circuits :	Suction – suction device that can function even without power (e.g. using <i>Venturi</i>)	Deleted
2	Page no. 22 Point no. A Patient circuits :	<ul style="list-style-type: none">Internal light – for illuminationVentilator – basic ventilator with at least CPAP and IMV modes with controls for CPAP/PEEP, PIP, rate, Ti and FiO₂.	Deleted
3	Addendum as new bullet in Point no. A Patient circuits :		Air oxygen Blender should be there and it should be USFDA and European CE approved
4	Page no. 22 Point no. A Power system:	Rechargeable battery with charge lasting for at least 4-6 hours	Rechargeable battery with UPS system is required which can provide backup of atleast 1 hour.
5	Page no. 22 Point no. A Power system:	<ul style="list-style-type: none">The battery should be capable of recharging from mains as well as the ambulance power sourceIt should be able to run the following equipments when disconnected from the power source: heater, suction machine, and ventilator	Deleted

SCHEDULE- B TRANSPORT INCUBATOR + VENTILATOR (1 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 23 Point no. 8	Two 10L integrated oxygen cylinders, regulator and flow meter	Two 10L or four 5L integrated oxygen cylinders, regulator and flow meter
2	Page no. 23 Point no. 18	2 x empty 10 L oxygen cylinders.	2 x empty 10 L / 4 x empty 5L oxygen cylinders.

Store Officer
AIIMS, Raipur