



Tender enquiry no. ADMIN/Tender/ Paediatric_Equipments/1/2014 **Dated 27/11/2014**

Corrigendum

Tender for "Paediatric Equipments"

SCHEDULE-B TRANSPORT INCUBATOR + VENTILATOR (1 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 24	2) Front and head access door, slide-out mattress tray	2) Front and head access door, slide-out mattress tray with access port holes and tubing access ports
2	Page no. 24	7) Digital displays outside shows air temperature	7) Digital displays outside shows air temperature & skin temperature
3	Page no. 24	12) Battery and AC supported	12) Battery and AC supported with indicator for battery level and mains mode
4	Page no. 24	19) It should be US FDA or European CE approved product.	19) It should be US FDA and European CE approved product.
5	Addendum Page no. 24	-	<ul style="list-style-type: none">Should have examination light.Should have inbuilt / external compressor for air supply during transport.

SCHEDULE-C WEIGHING MACHINE (2 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 25	<ul style="list-style-type: none">Remote display (at least 2 meter)	Deleted
2	Addendum Page no. 25	-	<ul style="list-style-type: none">Instrument should have transport castors base should be slip proof.

SCHEDULE-F INFANTOMETER (03 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 27	7. Should be ISO 9001 certified	7. Should be US FDA / CE (European) certified

SCHEDULE-G STADIOMETER (02 Nos.)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	Amendment
1	Page no. 27	<ul style="list-style-type: none">Weight 3.6 KG	<ul style="list-style-type: none">Weight < 3.5 KG
2	Addendum Page no. 27	-	<ul style="list-style-type: none">Foldable height measuring deviceMust have a measurement reading window