



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road,
Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Tender enquiry no. ADMIN/Tender/ Pediatrics Equipment/1/2013

Corrigendum for Pediatrics Equipment

Clause	Existing Parameter	Amendment
Page no. 01 Para 02	"The Administrative Officer, AIIMS, Tatibandh Raipur (CG) - 492001, by or before on 03.00 PM on 04-10-2013.	"The Administrative Officer, AIIMS, Tatibandh Raipur (CG) - 492099, by or before on 03.00 PM on 11-10-2013.
Page no. 06 Point 01	date of 04-10-2013, along with their bid in the Cover-I containing "Technical Bid".	date of 11-10-2013 , along with their bid in the Cover-I containing "Technical Bid".
Page no. 06 Point 02	"Dentistry Equipments" should reach AIIMS, Raipur by or before 03.00 PM on 04-10-2013.	"Pediatrics Equipments" should reach AIIMS, Raipur by or before 03.00 PM on 11-10-2013 .
Page no. 11 Point no. 5a	Mattress Width should be 55-60 cm	Mattress Width should be 45-60 cm
Page no.11 Point no.5c	Mattress Thickness should be 3-5 cm	Mattress Thickness should be 2.5-5 cm
Page no. 17 Point no. 20	Battery should be integrated and should provide back up to both ventilator & Air Compressor	Battery backup through UPS/integrated and should provide back up to both ventilator & Air Compressor
Page no. 17 Point no. 27	Should have permanent Electronic O2 Sensor.	Should have electrochemical O2 sensor.
Page no. 17 Point no. 29	Ventilator should have Up gradation facility with EtCO2	Deleted

Administrative Officer
AIIMS, Raipur