



Tender enquiry no. ADMIN/Tender/ICU_&_Neonatal_Ventilator/2/2014 dated 19-12-2014
Corrigendum for ICU & Neonatal Ventilator

ICU Ventilator (Schedule-A)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Pg. no. 21 Point (A), 2	Screen size should be 10" inch or more.	Screen size should be 12" inch or more.
2	Pg. no. 21 Point (A), 4, e	Advanced mode like Pressure Regulated volume control mode	Advanced mode like Pressure Regulated volume control mode/ Auto Flow
3	Pg. no. 23 (A). Point No.16	Should have Ultrasonic / Paramagnetic / Galvanic O2 sensor. All sensor should be covered in warranty-guarantee and CMC.	Should have Paramagnetic cell for O2 or if other type of sensor, then it should be provide 10 nos. extra sensor along with unit. All sensors should be covered in warranty-guarantee and CMC.

Neonatal Ventilator (Schedule-B)

Sr. No.	Page No/ Clause no/ point no	Existing sentence	To be read as
1	Pg. no. 24 S.No. (A) Point no 4	Should have an integrated high resolution screen with color display of at least 12 inches screen size	Should have an integrated high resolution screen with color display of at least 12 inches touch screen color
2	Pg. no. 24 S.No.(B) Point No 2	Non-invasive ventilation: CPAP and NIMV	Nasal CPAP
3	Pg. no. 24 S.No.(D) Point No. 3, g	Tidal Volume: 2 - 200 ml	Tidal Volume Guarantee: 2 - 100 ml
4	Pg. no. 25 S. No. (E) Point No. 9	Trends of set and measured parameters with a trend memory of at least 48 hours	Trends of set and measured parameters with a trend memory of at least 24 hours

Store officer
AIIMS, Raipur